

Section 5 of Traditional 510(K) Submission:**JUL 29 2014****510 (K) Summary****This 510(K) Summary of safety and effectiveness information is being submitted in accordance with requirement of 21 CFR807.92**

1. Date of Submission: Dec. 11, 2013
2. Submitter / 510(K) Holder

Suzhou Kangli Orthopaedics Instrument Co., Ltd.,
Sha Zhou East Road,
Zhangjiagang City,
Jiangsu Province,
China 215625

Contact Person: Miss Jin Hua Huang
Tel: (86) 0512-58560562
Fax: (86) 0512-58358951
E-mail: hjh@kang-li.cn

3. Proposed Device Name

Trade name: KangLi®
Common name: Locking Plate System

Classification Name: Plate, Fixation, Bone
Device Class: Class II
Classification Panel: Orthopedic Panel
Product Code: HRS
Regulation Number: 21 CFR 888.3030

Classification Name: Screw, Fixation, Bone
Device Class: Class II
Classification Panel: Orthopedic Panel
Product Code: HWC
Regulation Number: 21 CFR 888.3040

4. Predicate Devices

510 (k) Number: K130340
Product Name: Locking Bone Plates and Screws
Submitter: Weigao Orthopaedic Device Co., Ltd.

5. Device Description

KangLi[®] locking plate system contains locking plates with various specifications, metal bone and locking screws with various specifications, and various specific instruments. The bone plates are used for fixation of bones. The screws are used for fix the plates on the bones and the instruments are used for completing the surgery.

The bone plates are manufactured from unalloyed titanium that conforms to ASTM F67. The metal bone and locking screws are made of Ti6Al4V ELI that meets to ASTM F136. The materials of titanium and Ti6Al4V ELI are widely used in the industry with well-known biocompatibility. No new materials are used in the development of this implant.

6. Indication for Use/Intended Use

KangLi[®] locking plate system is intended for adult patients as indicated for fixation of fractures of tibia.

7. Non-Clinical Testing

Bench tests were conducted to verify that proposed device meet all design specifications as was Substantially Equivalent (SE) to the predicate device. The test results demonstrated that proposed device complies with the following standards:

ASTM F 382-99 (Reapproved 2008), Standard Specification and Test Method for Metallic Bone Plates, including the following items:

- * Static four point bending
- * Dynamic four point bending

ASTM F 543-07, Standard Specification and Test Methods for Metallic Medical Bone Screws including the following item:

- * Torsional properties
- * Driving torque
- * Pull out test

8. Substantially Equivalent Conclusion

The KangLi[®] locking plate system has less intended use than the predicate device and similar technological characteristics as the predicate device. The proposed device, the KangLi[®] locking plate system, is determined to be Substantially Equivalent (SE) to the predicate device, K130340 locking bone plates and screws, in respect of safety and effectiveness.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

July 29, 2014

Suzhou Kangli Orthopaedics Instrument Co., Ltd.
% Ms. Alice Gong
Shanghai Yarui Consultant Co., Ltd.
503 Room, 8 Building, 600 Liu Zhou Road
Shanghai, Shanghai 200233
China

Re: K133840

Trade/Device Name: KangLi® Locking Plate System

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/multiple component metallic bone fixation appliances and accessories

Regulatory Class: Class II

Product Code: HRS, HWC

Dated: June 12, 2014

Received: June 26, 2014

Dear Ms. Gong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing

Page 2 – Ms. Alice Gong

(21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Section 4 of Traditional 510(K) Submission:**Indications for Use****510(k) Number:** K133840**Device Name:** KangLi[®] Locking Plate System**Indications For Use:**

KangLi[®] locking plate system is intended for adult patients as indicated for fixation of fractures of tibia.

Prescription Use **AND/OR** **Over-The-Counter Use**

(Part 21 CFR 801 Subpart D)

(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH)

Elizabeth L. Frank -S

Division of Orthopedic Devices