

510(k) SUMMARY: ALTERA™ Spacer

JUL 01 2014

Company: Globus Medical Inc.
2560 General Armistead Avenue.
Audubon, PA 19403
(610) 930-1800

Contact: Christina Kichula
Group Manager, Regulatory Affairs

Date Prepared: February 17, 2014

Device Name: ALTERA™ Spacer

Classification: Per 21 CFR as follows:
§888.3080 Intervertebral Body Fusion Device
Product Code: MAX
Regulatory Class: II, Panel Code: 87

Predicate(s): CALIBER® Spacer (K102293)
PATRIOT® Spacer (K072970 & K122097)
CoAlign Innovation AccuLIF® TL Cage (K113465)

Purpose:

The purpose of this submission is to request clearance for the ALTERA™ Spacer.

Device Description:

The ALTERA™ Spacer is an expandable lumbar interbody fusion device used to provide structural stability in skeletally mature individuals following discectomy. The ALTERA™ Spacer accommodates various surgical approaches to the lumbar spine (posterior or transforaminal [posterolateral]) and allows articulation upon insertion. The device is available in various height ranges, allowing continuous expansion within the range, to fit the anatomical needs of a wide variety of patients. This device is to be filled with autogenous bone graft material. Protrusions on the superior and inferior surfaces of each device grip the endplates of the adjacent vertebrae to resist expulsion.

The ALTERA™ Spacer is made from titanium alloy, as specified in ASTM F136, F1295, and F1472. Internal components are made from radiolucent PEEK polymer and cobalt chromium molybdenum alloy, as specified in ASTM F2026 and F1537.

Indications for Use:

The ALTERA™ Spacer is an interbody fusion devices intended for use in patients with degenerative disc disease (DDD) at one or two contiguous levels of

the lumbosacral spine (L2-S1). DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These patients should be skeletally mature and have had at least six (6) months of non-operative treatment. In addition, these patients may have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s).

The ALTERA™ Spacer is to be filled with autogenous bone graft material. The device is intended to be used with supplemental fixation.

Performance Data:

Mechanical testing (static and dynamic compression, static and dynamic compression shear, and subsidence) was conducted in accordance with the "Guidance for Industry and FDA Staff, Class II Special Controls Guidance Document: Intervertebral Fusion Device," June 12, 2007, ASTM F2077, and ASTM F2267 to demonstrate substantial equivalence to the predicate devices. Results demonstrate that the ALTERA™ Spacer performs equivalently to or better than the predicate CALIBER® Spacer.

Basis for Substantial Equivalence:

The ALTERA™ Spacer is similar to the predicate systems with respect to technical characteristics, performance, design, materials, and intended use. The information provided within this premarket notification supports substantial equivalence to the predicate devices. The ALTERA™ Spacer's safety and effectiveness profile is similar to the cited predicates.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

July 1, 2014

Globus Medical, Incorporated
Ms. Christina Kichula
Group Manager, Regulatory Affairs
2560 General Armistead Avenue
Audubon, Pennsylvania 19403

Re: K140411
Trade/Device Name: ALTERA™ Spacer
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral body fusion device
Regulatory Class: Class II
Product Code: MAX
Dated: May 30, 2014
Received: June 2, 2014

Dear Ms. Kichula:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Ronald P. Jean -S for

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K140411

Device Name
ALTERA™ Spacer

Indications for Use (Describe)

The ALTERA™ Spacer is an interbody fusion device intended for use in patients with degenerative disc disease (DDD) at one or two contiguous levels of the lumbosacral spine (L2-S1). DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These patients should be skeletally mature and have had at least six (6) months of non-operative treatment. In addition, these patients may have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s).

The ALTERA™ Spacer is to be filled with autogenous bone graft material. The device is intended to be used with supplemental fixation.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Anton E. Dmitriev, PhD

Division of Orthopedic Devices