

510(k) Summary

K140799

Date Prepared: May 9th, 2014



medCOMP

1499 Delp Drive

Harleysville, PA 19438

Tel: 215-256-4201

Fax: 215-256-1787

www.medcompnet.com

A. Submitter Information

Submitter: MEDCOMP®
1499 Delp Drive
Harleysville, PA 19438
Tel: (215) 256-4201, x2271
Fax: (215) 256-9191
Contact: Jessica Leo, Regulatory Associate

JUN 13 2014

B. Trade Name Celerity™ System
Common Name: PICC placement accessory
Classification Name: Percutaneous, implanted, long-term intravascular Catheter
Product Code: LJS,
CFR: 21CFR880.5970
Class: II

C. Predicate Devices K093775;
Sapiens™ Tip Location System
Romedex International Srl
Common Name: PICC placement accessory
Classification Name: Percutaneous, implanted, long-term intravascular Catheter
Product Code: LJS,
CFR: 21CFR880.5970
Class: II

D. Device Description

The Celerity System includes the Celerity Monitor/Software, ECG Patient Cable, Remote Control Cable, Battery, Power Supply Cord and ECG Clip Cable (alligator clip). Procedural accessories including ECG Snap Leads, ECG Electrodes, Cable Cover and Prep Pads are provided as a convenience for the clinician.

E. Intended Use

The Celerity System is intended to provide real time tip location information of a central venous catheter by utilization of ECG to observe P-wave changes as the tip approaches the right atrium of the heart via the superior vena cava.

F. Indications for Use

The Celerity System is indicated for use as a supplemental aid in positioning of Peripherally Inserted Central Catheters (PICC) in adult patients. It provides real-time catheter tip location information by using the patient's cardiac electrical activity. Confirmation of tip placement should be verified according to clinical judgment and established hospital protocol (e.g., chest x-ray, fluoroscopy).

Note: Limiting, but not contraindicated, situations for this technique, are patients where cardiac rhythms may change presentation of the P wave.

- Atrial fibrillation
- Atrial flutter
- Severe tachycardia
- Pacemaker driven rhythm
- Chronic obstructive pulmonary disease (COPD)

Such patients are easily identified prior to PICC insertion. Use of additional confirmation method is necessary to confirm catheter tip location.

G. Technological Characteristics

Technological Characteristics of the Celerity System are equivalent with respect to the basic system design and function to that of the predicate device. The differences between the predicate and proposed devices do not raise new questions of safety or effectiveness.

H. Safety and Performance

Design Verification and Validation activities were performed in accordance with Design Control requirements per 21 CFR 820.30 and demonstrate that the subject Celerity System meets predetermined performance specifications.

The performance evaluation plan included testing per the following recognized standards to assess conformance to IEC 60601 (3rd Edition).

IEC 60601-1	Medical Electrical Equipment – Part 1: General Requirements for Safety
IEC 60601-1-2	Medical Electrical Equipment – Part 2: General Requirements for Basic Safety and Essential Performance – Collateral Standard Electro Magnetic Compatibility – Requirements and Test

I. Substantial Equivalence Conclusion

The proposed device is substantially equivalent to the predicate device based on:

- The Intended Use
- Operating principles/technology
- Results of safety and performance testing
- Responses to questions raised in the FDA 510(k) decision tree



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

June 13, 2014

Medcomp®
C/O Ms. Lorraine M. Hanley
Medical Components, Inc., d.b.a Medcomp
25 Forest Drive
Marlborough, MA 01752

Re: K140799
Trade/Device Name: Celerity™ System
Regulation Number: 21 CFR 880.5970
Regulation Name: Percutaneous, Implanted, Long-term Intravascular Catheter
Regulatory Class: II
Product Code: LJS
Dated: May 27, 2014
Received: May 27, 2014

Dear Ms. Hanley:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mary  -S

Erin I. Keith, M.S.
Division Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K 140799

Device Name

Celery PICC Tip Confirmation System

Indications for Use (Describe)

The Celery System is indicated for use as a supplemental aid in positioning of Peripherally Inserted Central Catheters (PICC) in adult patients. It provides real-time catheter tip location information by using the patient's cardiac electrical activity. Confirmation of tip placement should be verified according to clinical judgement and established hospital protocol (e.g., chest x-ray, fluoroscopy).

Note: Limiting, but not contraindicated, situations for this technique, are patients where cardiac rhythms may change presentation of the P wave.

- Atrial fibrillation
- Atrial flutter
- Severe tachycardia
- Pacemaker driven rhythm
- Chronic obstructive pulmonary disease (COPD)

Such patients are easily identified prior to PICC insertion. Use of an additional confirmation method is necessary to confirm catheter tip location.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)



Digitally signed by
Richard C. Chapman -S
Date: 2014.06.13
14:50:17 -04'00'

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."