

**5. 510(k) Summary**

**Date Prepared [21 CFR 807.92(a)(1)]**

July 18, 2014

**JUL 22 2014**

**Submitter's Information [21 CFR 807.92(a)(1)]**

Jan Triani  
EDGE Orthopaedics  
6 Mars Court, Unit 6-3  
Boonton, NJ 07005  
Phone: (201)543-9388  
Fax: (973)794-6813

Establishment Registration Number: 3010726797

**Trade Name, Common Name, Classification [21 CFR 807.92(a)(2)]**

Trade Name: EDGE Orthopaedics VIEW and REDUCE Plating Systems  
Common Name: Bone Fixation Plate and Bone Screw  
Classification Name: Class II, 21 CFR 888.3030 Plate, Fixation Bone  
Class II, 21 CFR 888.3040 Screw, Fixation, Bone  
Product Code: HRS/HWC  
Panel Code: Orthopedics

**Predicate Device [21 CFR 807.92(a)(3)]**

DARCO Locking Plating System – K061808 (Darco International)  
ORTHOLOC® 2.0/2.4 Plate System – K090692 (Wright Medical Technology)

**Description of the Device [21 CFR 807.92(a)(4)]**

The EDGE Orthopaedics' VIEW™ Plating System has been designed to support multiple indications within the mid-foot. The system includes titanium alloy, sterile packaged Lapidus, First Metatarsophalangeal (MTP), Calcaneal-Cuboid (CC), Evans and Talonavicular (TN) Plates along with sterile titanium alloy bone screws. The design creates a "window" offering a continuous view into the healing process.

The EDGE Orthopaedics' REDUCE® Fracture Plating System has been designed to support multiple indications within the forefoot and mid-foot. The system includes titanium alloy, sterile packaged T, L, Y, and Straight Plates along with sterile titanium alloy bone screws.

VIEW and REDUCE Plating Systems are offered in a variety of sizes for use with the non-locking and locking bone screws. The screws are available in variety of diameters and lengths to help support the fixation, correction or stabilization of bones.

The corresponding instrumentation (depth gauges, screwdrivers, reamers, and plate benders) to facilitate insertion is found in EDGE's RIVAL Instrument Tray.

**Intended Use [21 CFR 807.92(a)(5)]**

EDGE Orthopaedics' VIEW and REDUCE Plating Systems are intended for use in stabilization and fixation of fresh fractures, revision procedures, joint fusion, and reconstruction of small bones of the hand, feet, wrist, ankles, fingers and toes. The plates are available for use with EDGE Orthopaedics' locking and non-locking bone screws.

Plates and screws are intended for single use only. Screws are not intended for use in the spine.

**Technological Characteristics [21 CFR 807.92(a)(6)]**

The subject device is similar in design, material and indications to the predicate devices.

**Performance Data [21 CFR 807.92(b)(1)]**

The safety and effectiveness of the EDGE Orthopaedics' VIEW and REDUCE Plating Systems are adequately supported by the substantial equivalence information, materials information and the following non-clinical testing provided within this Premarket Notification: torsional strength per ASTM F543, driving and removal torque per ASTM F543, axial pull out strength per ASTM F543, and Single Cycle Bend Testing per ASTM F382.

**Clinical Data [21 CFR 807.92(b)(2)]**

Clinical data was not used to determine substantial equivalence.

**Conclusion [21 CFR 807.92(b)(3)]**

The analysis of the EDGE Orthopaedics' VIEW and REDUCE Plating Systems within this Premarket Notification supports the conclusion that the subject device is substantially equivalent to the predicate devices.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

July 22, 2014

Icon Orthopedic Concepts, LLC DBA EDGE Orthopaedics  
Ms. Jan Triani  
Director of Quality Assurance and Regulatory Affairs  
6 Mars Court, Unit 6-3  
Boonton, New Jersey 07005

Re: K140876

Trade/Device Name: EDGE Orthopaedics VIEW™ and REDUCE™ Plating Systems

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/multiple component metallic bone fixation appliances and accessories

Regulatory Class: Class II

Product Code: HRS, HWC

Dated: April 24, 2014

Received: April 25, 2014

Dear Ms. Triani:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing

Page 2 – Ms. Jan Triani

(21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Lori A. Wiggins**

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

