

510(k) Summary**JUL 11 2014**

Date Prepared: July 1, 2014

Company: Surgical Specialties Corporation
100 Dennis Dr.
Reading, PA 19606

Contact: Kirsten Stowell Franco
Sr. Regulatory Affairs Manager
Phone: 610-404-3367
Fax: 610-404-3924
Email: kstowell@surgicalspecialties.com

Device trade name: Quill™ Monoderm™ Knotless Tissue-Closure Device (PGA-PCL)

Device Common Name: Absorbable Poly(glycolide/l-lactide) Surgical Suture

Device classification: Suture, Absorbable, Synthetic, Polyglycolic Acid
Product code, GAM
21 CFR 878.4493
Class II

Legally marketed device to which the device is substantially equivalent: K072028: Quill™ Monoderm™ (PGA-PCL)
K052437: Monoderm™ (PGA-PCL) Surgical Suture

Description of the device: The Quill™ Monoderm™ Knotless Tissue-Closure Device is a sterile, synthetic absorbable tissue approximation device that is comprised of a copolymer of glycolide and ε-caprolactone, undyed, or dyed with D&C Violet No. 2. The device is designed with small bi-directional barbs along the long axis of the suture monofilament. It is available in diameter Sizes 0 to 5-0, in various lengths affixed to various needle types.

Indications for Use: Quill™ MONODERM™ device is indicated for use in soft tissue approximation where the use of absorbable sutures is appropriate.

**Substantial
Equivalence:**

The proposed additional diameter sizes, Size 4-0 and 5-0, of the Quill™ Monoderm™ Knotless Tissue-Closure Device product line have the same material, design, intended use and technological characteristics as the predicate device. The only difference between the proposed and predicate device is the suture diameter.

Performance tests:

Non-clinical laboratory performance testing was conducted to confirm that the Quill™ Monoderm™ Knotless Tissue-Closure Device conforms to the USP monograph for absorbable sutures (as applicable). This testing was performed in accordance with FDA's Class II Special Controls Guidance Document: Surgical Sutures, Issued June 3, 2003. Additional performance testing was conducted in order to demonstrate substantial equivalence to the predicate device including *in vitro* post-hydrolysis tensile testing and *in vivo* resorption testing.

The results of this testing demonstrates that the Quill™ Monoderm™ Knotless Tissue-Closure Device is substantially equivalent to the predicate device.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

July 11, 2014

Surgical Specialties Corporation
% Angiotech Puerto, Inc.
Kirsten Stowell Franco
Sr. Regulatory Affairs Manager
100 Dennis Drive
Reading, Pennsylvania 19606

Re: K141558

Trade/Device Name: Quill™ Monoderm™ Knotless Tissue-Closure Device
Regulation Number: 21 CFR 878.4493
Regulatory Name: Absorbable poly(glycolide/L- lactide) surgical suture
Regulatory Class: Class II
Product Code: GAM
Dated: June 10, 2014
Received: June 12, 2014

Dear Ms. Franco:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

David Krause -S

for **Binita Ashar, M.D., M.B.A., F.A.C.S.**
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

cc: DMC
510(k) Staff
Division
D.O.

Indications for Use

510(k) Number (if known)

K141558

Device Name

Quill™ MONODERM™ Knotless Tissue-Closure Device (PGA-PCL)

Indications for Use (Describe)

Quill™ MONODERM™ Knotless Tissue-Closure Device is indicated for use in soft tissue approximation where the use of absorbable sutures is appropriate.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Peter L. Hudson -S

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."