



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

September 3, 2014

Zimmer, Incorporated  
Ms. Diana Taylor  
Regulatory Affairs Senior Specialist  
P.O. Box 708  
Warsaw, Indiana 46580

Re: K141840

Trade/Device Name: *NexGen*® Complete Knee Solutions Trabecular Metal Cone Broach  
Regulation Number: 21 CFR 888.3565  
Regulation Name: Knee joint patellofemorotibial metal/polymer porous-coated uncemented  
prosthesis  
Regulatory Class: Class II  
Product Code: MBH, JWH  
Dated: July 07, 2014  
Received: July 08, 2014

Dear Ms. Taylor:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing

(21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Lori A. Wiggins -S**

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K141840

Device Name

NexGen® Complete Knee Solutions Trabecular Metal Cone Broach

Indications for Use (Describe)

Trabecular Metal Tibial Cone Augments are intended for use where severe degeneration, trauma, or other pathology of the knee joint indicates total knee arthroplasty. When used with the NexGen Complete Knee Solution - Rotating Hinge Knee (RHK) System, the Trabecular Metal Tibial Cone Augments are for cemented use only. When used with the NexGen Complete Knee Solution - Legacy Constrained Condylar Knee System, the Trabecular Metal Tibial Cone Augments are for cementless or cemented use.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

### FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*



### 510(k) Summary

**Sponsor:** Zimmer, Inc.  
P.O. Box 708  
Warsaw, IN 46581-0708

**Contact Person:** Diana Taylor  
Regulatory Affairs  
Telephone: (574) 267-6131  
Fax: (574) 372-4605

**Date:** July 7, 2014

**Trade Name:** NexGen® Complete Knee Solutions Trabecular Metal  
Cone Broach

**Common Name:** Broach

**Classification:** Class II Accessory  
21 CFR 21 888.3560 and 21 CFR 888. 3565

**Product Code:** JWH and MBH

**Classification Panel:** Orthopedics

**Predicate Device(s):** Trabecular Metal Tibial Cone Augments manufactured by  
Trabecular Metal Technologies, K102896, 12/13/2010

NexGen Trabecular Metal Coupled Tibial Cones  
manufactured by Trabecular Metal Technologies, K120990,  
06/29/2012.

**Description:** The NexGen TM Cone Broach, a 2-piece instrument (Broach  
and Handle), are manufactured from 17-4 stainless steel and  
designed to provide a multi-functional instrument to  
streamline the surgical process used when implanting the  
Trabecular Metal Tibial Cone Implants.

The broaches are used to prepare the tibial metaphysis; the  
broach have teeth that are necessary to cut through bone and  
form the tibial canal to the correct shape and depth for a  
Trabecular Metal Tibial Cone Implant. The broaches are  
manually impacted, using a sequentially larger broach until  
reaching the final size for the Implant.

<b>Indications for Use:</b>	Trabecular Metal Tibial Cone Augments are intended for use where severe degeneration, trauma, or other pathology of the knee joint indicates total knee arthroplasty. When used with the NexGen Complete Knee Solution - Rotating Hinge Knee (RHK) System, the Trabecular Metal Tibial Cone Augments are for cemented use only. When used with the NexGen Complete Knee Solution - Legacy Constrained Condylar Knee System, the Trabecular Metal Tibial Cone Augments are for cementless or cemented use.
<b>Comparison to Predicate:</b>	Both the NexGen TM Cone Broach and the predicate broach are non-sterile, reusable manual orthopedic instruments that perform the same function as the predicate, are manufactured from identical materials, design mimics the currently marketed cone implants in order to prepare the tibial canal for Trabecular Metal Cone Implants using the same principal of operation (manual, sequential impaction with a mallet). The addition of design specific Coupled broaches and Large broach sizes provides design specific instrumentation for the full line of Cone Implants. The Large broach replaces a challenging manual burring step. The additional functionality of the instrumentation allows the surgical procedure to be streamlined. The TM Cone Broach is as safe and effective as the predicate instrumentation used to implant the Coupled, Medium and Large Trabecular Metal Cone Implants.
<b>Substantial Equivalence Conclusions</b>	Zimmer considers this 2-piece broach equivalent to the predicate one-piece broach based on design, size range and performance.