



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

November 14, 2014

ConforMIS Incorporated
Ms. Amita S. Shah
Senior Vice President, Regulatory and Quality Affairs
28 Crosby Drive
Bedford, Massachusetts 01730

Re: K142161

Trade/Device Name: ConforMIS iTotal Cruciate Retaining Knee Replacement System
(iTotal CR KRS)
ConforMIS iTotal Posterior Stabilized Knee Replacement System
(iTotal PS KRS)

Regulation Number: 21 CFR 888.3560

Regulation Name: Knee joint patellofemorotibial polymer/metal/polymer semi-constrained
cemented prosthesis.

Regulatory Class: Class II

Product Code: JWH, OOG, OIY

Dated: October 14, 2014

Received: October 15, 2014

Dear Ms. Shah:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act

or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins -S

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K142161

Device Name:

ConforMIS iTotal Cruciate Retaining Knee Replacement System (iTotal CR KRS)

Indications for Use:

The iTotal CR Knee Replacement System (KRS) is intended for use as a total knee replacement in patients with knee joint pain and disability whose conditions cannot be solely addressed by the use of a prosthetic device that treats only one or two of the three knee compartments, such as a unicondylar, patellofemoral or bicompartamental prosthesis.

The Indications for Use include:

- Painful joint disease due to osteoarthritis, traumatic arthritis, rheumatoid arthritis or osteonecrosis of the knee.
- Post traumatic loss of joint function.
- Moderate varus, valgus or flexion deformity in which the ligamentous structures can be returned to adequate function and stability.
- Failed osteotomies, hemiarthroplasties, and unicondylar, patellofemoral or bicompartamental implants.
- Revision procedures provided that anatomic landmarks necessary for alignment and positioning of the implant are identifiable on patient imaging scans.

This implant is intended for cemented use only.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Indications for Use

510(k) Number (if known): K142161

Device Name:

ConforMIS iTotal Posterior Stabilized Knee Replacement System (iTotal PS KRS)

Indications for Use:

The iTotal PS Knee Replacement System (KRS) is intended for use as a total knee replacement in patients with knee joint pain and disability whose conditions cannot be solely addressed by the use of a prosthetic device that treats only one or two of the three knee compartments, such as a unicondylar, patellofemoral or bicompartamental prosthesis.

The Indications for Use include:

- Painful joint disease due to osteoarthritis, traumatic arthritis, rheumatoid arthritis, polyarthritis or osteonecrosis of the knee.
- Post traumatic loss of joint function.
- Moderate varus, valgus or flexion deformity.
- Failed osteotomies, hemiarthroplasties, and unicondylar, patellofemoral or bicompartamental implants.
- Revision procedures provided that anatomic landmarks necessary for alignment and positioning of the implant are identifiable on patient imaging scans

This implant is intended for cemented use only.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 2 of 2

510(K) SUMMARY (PAGE 1 OF 5)

Submitter's Name and Address: ConforMIS Inc.
28 Crosby Drive
Bedford, MA 01730

Establishment Registration Number: 3009844603 and 3004153240

Date of Summary: July 31, 2014

Contact Person: Amita S. Shah, Sr. Vice President, Regulatory and Quality Affairs
Telephone Number: (781) 345-9164
Fax Number: (781) 345-0147

Name of the Device(s): ConforMIS iTotal[®] Cruciate Retaining Knee Replacement System (iTotal CR KRS)
ConforMIS iTotal[®] Posterior Stabilized Replacement System (iTotal PS KRS)

Common Name(s): Cruciate Retaining Knee Replacement System
Posterior Stabilized Knee Replacement System

Regulatory Status and Regulation Number: Class II
21 CFR 888.3560

Classification Name: Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis

Device Classification: Product Code:
JWH: Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis.

OOG: Knee Arthroplasty Implantation System.
Intended to be used to assist in the implantation of a specific knee arthroplasty device or a set of specific knee arthroplasty devices. Indicated to include guiding alignment, making or establishing cuts, selecting, sizing, attaching, positioning or orienting implant components.

OIY: Prosthesis, knee, patellofemorotibial, semi-constrained, cemented polymer + additive/metal/polymer + additive. This generic type of device includes prostheses that have a femoral component made of alloys, such as cobalt-chromium-molybdenum, and a tibial component(s) and/or a retropatellar resurfacing component made of ultra-high molecular weight polyethylene plus an additive, such as a-tocopherol.

510(K) SUMMARY (PAGE 2 OF 5)**Indications for Use:** iTotal® CR KRS

The iTotal® CR Knee Replacement System (KRS) is intended for use as a total knee replacement in patients with knee joint pain and disability whose conditions cannot be solely addressed by the use of a prosthetic device that treats only one or two of the three knee compartments, such as a unicondylar, patellofemoral or bicompartamental prosthesis.

The Indications for Use include:

- Painful joint disease due to osteoarthritis, traumatic arthritis, rheumatoid arthritis or osteonecrosis of the knee.
- Post traumatic loss of joint function.
- Moderate varus, valgus or flexion deformity in which the ligamentous structures can be returned to adequate function and stability.
- Failed osteotomies, hemiarthroplasties, and unicondylar, patellofemoral or bicompartamental implants.
- Revision procedures provided that anatomic landmarks necessary for alignment and positioning of the implant are identifiable on patient imaging scans.

This implant is intended for cemented use only.

iTotal® PS KRS

The iTotal PS Knee Replacement System (KRS) is intended for use as a total knee replacement in patients with knee joint pain and disability whose conditions cannot be solely addressed by the use of a prosthetic device that treats only one or two of the three knee compartments, such as a unicondylar, patellofemoral or bicompartamental prosthesis.

The Indications for Use include:

- Painful joint disease due to osteoarthritis, traumatic arthritis, rheumatoid arthritis, polyarthritis or osteonecrosis of the knee.
- Post traumatic loss of joint function.
- Moderate varus, valgus or flexion deformity.
- Failed osteotomies, hemiarthroplasties, and unicondylar, patellofemoral or bicompartamental implants.
- Revision procedures provided that anatomic landmarks necessary for alignment and positioning of the implant are identifiable on patient imaging scans.

This implant is intended for cemented use only.

510(K) SUMMARY (PAGE 3 OF 5)**Identification of the Legally Marketed Device(s) (Predicate Device(s)):**

ConforMIS iTotal CR Knee Replacement System (ITOTAL CR KRS)
 ConforMIS iTotal PS Knee Replacement System (ITOTAL PS KRS)

Device Class: II
 Product Code: JWH, OOG, OIY
 Regulation Number: 21 CFR 888.3560
 510(k) Number: K140423 and K140833
 Device Class: II

Device Description:

The iTotal[®] Knee Replacement Systems (hereafter referred to as the "iTotal[®] KRS") are patient specific tricompartmental faceted knee replacement systems. The iTotal[®] KRS are semi-constrained, cemented knee implants which consist of femoral, tibial, and patellar components.

Using patient imaging and a combination of proprietary and off the shelf software a patient specific implant is designed, that best meets the geometric and anatomic requirements of the specific patient. The femoral component is manufactured from cobalt chromium molybdenum ("CoCrMo") alloy. The tibial component includes a metal tray manufactured from CoCrMo alloy and either one or two (CR only) polyethylene inserts. The polyethylene inserts may be manufactured from either UHMWPE, or for iTotal CR KRS, a highly cross-linked Vitamin E infused polyethylene (iPoly XE™). The patellar component is also manufactured from either UHMWPE or from a highly cross-linked Vitamin E infused polyethylene (iPoly XE™).

For user convenience, and similar to the predicate iTotal KRS, accessory orthopedic manual surgical instruments designed for use with the modified iTotal[®] KRS are provided to assist with implantation. The ancillary instruments are provided sterile and for single-use only. These patient specific instruments are provided to assist in the positioning of total knee replacement components intra-operatively and in guiding the cutting of bone.

The function and general design features of the patient specific implants and ancillary instruments remain similar to those described in the predicate 510(k)'s **K140423** and **K140833**.

510(K) SUMMARY (PAGE 4 OF 5)**Substantial
Equivalence:**

The products subject of this premarket notification are substantially equivalent to the iTotals CR KRS (**K140423**, cleared May 27, 2014) and the iTotals PS KRS (**K140833**, cleared June 30, 2014). The following testing was performed to establish substantial equivalence:

- Software verification and validation testing of proprietary software

**Description and
Assessment
Nonclinical Testing:**

The modified iTotals KRS and predicate iTotals KRS have the same technological characteristics. There have been no changes to the modified iTotals KRS devices with respect to design, materials, and methods of manufacture, packaging, or sterilization. The designs of the devices remain similar to those cleared via **K140423** and **K140833**.

Modifications which represent improving manufacturing efficiencies are primarily focused on increasing automation of the CAD manufacturing process via the proprietary software modules as shown in **Table 1** below.

Table 1: Comparison between the Modified and Predicate Devices

Predicate iTotals Knee Replacement Systems (K140423 and K140833)	Modified iTotals Knee Replacement Systems (This submission)
SegSurf(T) version 2.6 or Manual Process	SegSurf(T) version 3.0 or Manual Process
iTotals FemJigs version 2.0 (CR only) or Manual Process	iTotals FemJigs version 3.0 or Manual Process
iTotals Tib version 2.1 (CR only) or Manual Process	iTotals Tib version 3.0 (CR only) or Manual Process
iTotals TibJigs version 1.1 (CR only) or Manual Process	iTotals TibJigs version 2.0 (CR only) or Manual Process
iTotals iView version 2.0 (CR only) or Manual Process	iTotals iView version 3.0 (CR only) or Manual Process

510(K) SUMMARY (PAGE 5 OF 5)

Nonclinical Testing: The determination of substantial equivalence for this device was based on a detailed device description. The following non-clinical laboratory testing was performed demonstrating that the device can be considered substantially equivalent to the predicate device for the intended use:

- Detailed software description and software verification and validation testing of proprietary software Segsurf(T)
- Detailed software description and software verification and validation testing of proprietary software iTotal FemJigs
- Detailed software description and software verification and validation testing of proprietary software iTotalTib
- Detailed software description and software verification and validation testing of proprietary software iTotal TibJigs

Detailed software description and software verification and validation testing of proprietary software iTotal iView

Conclusion:

The determination of substantial equivalence for these devices was based on detailed device descriptions. Non-clinical laboratory testing was performed demonstrating that the devices can be considered substantially equivalent to the predicate devices for the intended uses. Clinical data is not necessary to demonstrate substantial equivalence.

Based on the testing conducted, it is concluded that the modified iTotal KRS are substantially equivalent to the predicate iTotal® KRS (**K140423** cleared May 27, 2014 and **K140833** cleared June 30, 2014).