

January 16, 2015

Food and Drug Administration 10903 New Hampshire Avenue Document Control Center - WO66-G609 Silver Spring, MD 20993-0002

Yung Sheng Optical Co., Ltd. Mr. Wen-Han Chen Regulatory Affairs Supervisor 3F-1, No. 6 Jhongke Road, Daya District Taichung City 42881 Taiwan

Re: K143190

Trade/Device Name: Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic)

Contact Lens

Regulation Number: 21 CFR 886.5925

Regulation Name: Soft (Hydrophilic) Contact Lens

Regulatory Class: Class II Product Code: LPL, MVN Dated: November 17, 2014 Received: November 19, 2014

Dear Mr. Chen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

# Kesia Y. Alexander -S

for Malvina Eydelman, M.D.
Director
Division of Ophthalmic and Ear,
Nose, and Throat Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

# DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

# Indications for Use

Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017 See *PRA Statement below.* 

510(k) Number *(if known)* K143190

**Device Name** 

Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens

Indications for Use (Describe)

The Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens is indicated for daily wear single use only for the correction of refractive ametropia (myopia) in aphakic and not-aphakic persons with non-diseased eyes. The lenses may be worn by person who exhibit astigmatism of 2.00 diopters or less that does not interfere with visual acuity. The eye care professional should prescribe the lenses for daily wear single use only. The lenses are to be discarded upon removal: Therefore no cleaning or disinfecting is required. The Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lenses help protect against transmission of harmful UV radiation to the cornea and into the eye.

Type of Use (Select one or both, as applicable)

➤ Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

#### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

#### \*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\*

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff PRAStaff@fda. hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Yung Sheng Optical Co., Ltd. 510(k) Notification

Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens

# 510(k) Summary

#### K143190

**5.1 Type of Submission:** Traditional

**5.2 Submitter:** Yung Sheng Optical Co., Ltd.

Address: 3F-1, No.6, Jhongke Road, Daya District, Taichung

City 42881 Taiwan

Manufacturer No.8, Keya 2<sup>nd</sup> Road, Daya District, Taichung City

Address 42881, Taiwan

**Phone:** (04) 25658384 #156 / 157

**Fax:** (04) 25658387

Contact: Wen-Han Chen / Tsung-Jen Yeh

**Establishment Registration Number:** N/A **Date Prepared** December 05, 2014

## 5.3 Identification of the Device:

**Proprietary/Trade name:** Eye Secret 38 UV Aspheric (polymacon)

1-Day Soft (hydrophilic) Contact Lens

Common Name: Contact Lens

Classification Name: Lenses, Soft Contact, Daily Wear

**Device Classification:** 

**Regulation Number:** 21 CFR 886.5925 (b) (1)

Panel: Ophthalmic

Product Code: LPL for Lenses, Soft Contact, Daily Wear

MVN for Lenses, Soft Contact (Disposable)

#### **5.4 Identification of the Predicate Device:**

**Predicate Device Name:** Eye Secret 38 UV Aspheric (polymacon)

Soft (hydrophilic) Contact Lens for Daily

Wear

Manufacturer: Yung Sheng Optical Co., Ltd.

**510(k) Number or Clearance Information:** K132854

## 5.5 Intended Use and Indications for Use of the subject device.

The Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens is indicated for daily wear single use only for the correction of refractive ametropia (myopia) in aphakic and not-aphakic persons with non-diseased eyes. The lenses may be worn by person who exhibit astigmatism of 2.00 diopters or less that does not interfere with visual acuity. The eye care professional should prescribe the lenses for daily wear single use only. The lenses are to be discarded upon removal: Therefore no cleaning or disinfecting is required. The Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lenses help protect against transmission of harmful UV radiation to the cornea and into the eye.

#### **5.6 Device Description**

The Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens is manufactured by using cast molding method. The lens material, polymacon, is a random copolymer of 2-hydroxyethyl methacrylate (HEMA) crosslinked with ethylene glycol dimethacrylate (EGDMA). A UV absorbing monomer is used to block UV radiation. The average transmittance characteristics are less than 5 % in the UVB range of 280 to 315 nm and less than 50 % in the UVA range of 316 to 380 nm. The lenses are tinted blue for visibility purposes with the color additives, C.I. Reactive Blue No. 4.

The Lenses are available as aspheric lenses. Each finished lens is supplied in a plastic blister container with a sterile isotonic phosphate buffered saline solution containing 0.01 % sodium hyaluronate and 0.1 % trehalose wetting agents.

The purpose of this 510(k) Notification is to notify the FDA of the new device, intended for commercial distribution by Yung Sheng Optical Co., Ltd. under the trade name of Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens.

# **5.7 Summary of Clinical Study**

Polymacon lenses have been used widely. Its safety and effectiveness have been well documented and cleared by FDA. Eye Secret 38 UV Aspheric (polymacon) Soft (hydrophilic) Contact Lens for Daily Wear (K132854) submitted by Yung Sheng Optical Co., Ltd. is an example.

Clinical study for Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens is not required for the premarket notification as the USAN name and process are the same as the above mentioned predicate devices.

#### 5.8 Non-clinical Testing

A series of preclinical testing were performed to demonstrate the safety and effectiveness of the Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens. The results of all testing demonstrated that the safety and effectiveness of the Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens are equivalent to the Eye Secret 38 UV Aspheric (polymacon) Soft (hydrophilic) Contact Lens for Daily Wear (K132854). The following tests were conducted as recommended by the FDA Premarket Notification 510(k) Guidance Document for Daily Wear Contact Lenses, February 27, 1997:

- Toxicity
  - 1. Acute Systemic Injection Study: The lens material meets the requirements of the systemic injection test and is considered non-toxic.
  - 2. White Rabbit Ocular Irritation Test: Ocular irritation test was performed and produced no ocular irritation.
  - 3. Cytotoxicity Test: The test article meets the requirements of ISO 10993-5.
- Extractables (Leachability)
- Finished Lens Parameters
- Light Transmittance
- Refractive Index
- Water Content

- Shelf-life
- Mechanical Properties Comparative Testing
- Specific Gravity Comparative Testing
- Physical Compatibility Test with Contact Lens Care Solution and Packaging Solution

The results of the non-clinical testing demonstrated that the Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens is substantially equivalent to the predicate device.

## **5.9 Substantial Equivalence Determination**

The Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens submitted in this 510(k) file is substantially equivalent in intended use, design, technology/principles of operation, materials and performance to the cleared Eye Secret 38 UV Aspheric (polymacon) Soft (hydrophilic) Contact Lens for Daily Wear which is the subject of K132854. Differences between the devices cited in this section do not raise any new issues of safety or effectiveness.

Comparison Table		
Item	Eye Secret 38 UV 1-Day	Predicated Device (K132854)
Product Name	Eye Secret 38 UV Aspheric	Eye Secret 38 UV Aspheric
	(polymacon) 1-Day Soft (hydrophilic)	(polymacon) Soft (hydrophilic)
	Contact Lens	Contact Lens for Daily Wear
Regulatory	886.5925	886.5925
Number		
Classification	II	II
Intended Use	The Eye Secret 38 UV Aspheric	The Eye Secret 38 UV Aspheric
	(polymacon) 1-Day Soft (hydrophilic)	(polymacon) Soft (hydrophilic)
	Contact Lens is indicated for daily	Contact Lens for Daily Wear is
	wear single use only for the	indicated for daily wear for the
	correction of refractive ametropia	correction of refractive ametropia
	(myopia) in aphakic and not-aphakic	(myopia) in aphakic and not-aphakic
	persons with non-diseased eyes.	persons with non-diseased eyes.
	The lenses may be worn by person	The lenses may be worn by person
	who exhibit astigmatism of 2.00	who exhibit astigmatism of 2.00
	diopters or less that does not	diopters or less that does not
	interfere with visual acuity.	interfere with visual acuity.

Comparison Table		
Item	Eye Secret 38 UV 1-Day	Predicated Device (K132854)
Prescription Use	Yes	Yes
Material	Polymacon	Polymacon
Manufacturing Method	Cast molding	Cast molding
Water Content	38 %	38 %
Powers	-0.50 D ~ -20.00 D	-0.50 D ~ -12.00 D
Light Transmittance	95 ± 5 %	95 ± 5 %
UV-A	< 50 %	< 50 %
UV-B	< 5 %	< 5 %
Refractive Index	1.440 ± 0.005 n <sub>d</sub>	1.440 ± 0.005 n <sub>d</sub>
Base Curve	7.85 ~ 10.00 ± 0.20 mm	8.60 ± 0.20 mm
Diameter	12.00 ~ 15.00 ± 0.20 mm	14.00 ± 0.20 mm
Center Thickness	0.060 ± 0.020 mm	0.044 ~ 0.131 mm (varies with power)
Tint	C.I. Reactive Blue No. 4	C.I. Reactive Blue No. 4

# 5.10 Conclusion

After analyzing bench tests, safety testing data, it can be concluded that the Eye Secret 38 UV Aspheric (polymacon) 1-Day Soft (hydrophilic) Contact Lens is substantially equivalent to the predicate device.