



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

January 9, 2015

K2M, Incorporated
Ms. Nancy Giezen
Manager, Regulatory Affairs
751 Miller Drive Southeast
Leesburg, Virginia 20175

Re: K143334

Trade/Device Name: Range/Denali/Mesa Spinal System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: NKB, OSH, MNH, MNI, KWP, KWQ
Dated: November 19, 2014
Received: November 20, 2014

Dear Ms. Giezen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins -S

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K143334

Device Name

RANGE /DENALI/MESA Spinal System

Indications for Use (Describe)

RANGE /DENALI/MESA and SMALL STATURE and ARI are cleared for the following indications: Non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e. scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

Except for hooks, when used as an anterolateral thoracic/lumbar system the Range Spinal System may also be used for the same indications as an adjunct to fusion.

Except for the ARI staples, the Range Spinal System is indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis when used for posterior noncervical fixation in pediatric patients. The Range Spinal System for pediatric use is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY
Range/Mesa/Denali Spinal System
K2M, Inc.

Submitter

K2M, Inc.
751 Miller Drive SE
Leesburg, VA 20175

Contact Person: Nancy Giezen
Telephone: 571 919-2168
Date Prepared: 1/7/2015

Classification

Trade Name: Range/Denali/Mesa Spinal System
Common Name: Spinal Fixation System
Regulatory Class: Class II, Class III (NKB)

Classification Name(s):

Pedicle Screw Spinal System (21 CFR 888.3070, Product Codes: NKB, OSH, MNH, MNI)
Spinal Intervertebral body fixation Orthosis (21 CFR 888.3060, Product Code: KWQ)
Spinal Interlaminar fixation Orthosis (21 CFR 888.3050, Product Code: KWP)

Predicate Device(s)

Primary Predicate: K2M Range/Mesa/Denali Spinal System (K141873)

Additional Predicates:

- K2M Everest Spinal System (K140529)

Device Description

The Range/Mesa/Denali Spinal System is a top-loading, multiple component, posterior (thoracic-lumbar) spinal fixation system which consists of pedicle screws, rods, hooks and rod connectors. The subject 510(k) adds Mesa 2 implants and connectors to the system.

Function: The system functions as a spinal fixation device to provide support and stabilization of the posterior thoracic and lumbar spine.

Indications For Use

RANGE /DENALI/MESA and SMALL STATURE and ARI are cleared for the following indications:

Non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e. scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

Except for hooks, when used as an anterolateral thoracic/lumbar system the Range Spinal System may also be used for the same indications as an adjunct to fusion.

Except for the ARI staples, the Range Spinal System is indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis when used for posterior noncervical fixation in pediatric patients. The Range Spinal System for pediatric use is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

Technological Comparison to Predicate(s)

The Range/Mesa/Denali Spinal System was compared to predicate systems and the design features, materials and sizes were found to be substantially the same as these systems.

Non-clinical Performance Evaluation

Performance evaluations were conducted on constructs representing the worst case components (including static torsion, static compression and dynamic compression bending in accordance with ASTM F1717) and the proposed implants were found to be substantially the same as predicate devices.

Conclusion

There are no significant differences between the Range/Mesa/Denali System and other systems currently being marketed which would adversely affect the use of the product. It is substantially equivalent to these other devices in design, function, material and intended use.