



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

March 25, 2015

Cayenne Medical, Incorporated
Ms. Shima Hashemian
Senior Manager of Regulatory Affairs and Quality Assurance
16597 North 92nd Street
Scottsdale, Arizona 85260

Re: K143473

Trade/Device Name: Short SureLockTM All-Suture Anchor
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: MBI
Dated: February 16, 2015
Received: February 19, 2015

Dear Ms. Hashemian:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins -S

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

Device Name: Short SureLock™ All- Suture Anchor

Indications for Use:

The Cayenne Medical, Inc. Short SureLock™ All- Suture Anchor is intended to reattach soft tissue to bone in orthopedic surgical procedures. The system may be used in either arthroscopic or open surgical procedures. After the anchor is deployed in the bone, the floating sutures can be used to reattach soft tissue, such as ligaments, tendons, or joint capsules to the bone. The suture anchor system thereby stabilizes the damaged soft tissue, in conjunction with appropriate postoperative immobilization, throughout the healing period.

The Cayenne Medical, Inc. Short SureLock™ All-Suture Anchors are intended to be used for the reattachment of soft tissue to bone for the following indications:

Foot and Ankle

- Hallux valgus reconstruction
- Midfoot reconstruction

Hand and Wrist

- Ulnar or lateral collateral ligament reconstruction
- Repair/reconstruction of collateral ligaments
- Flexor and extensor tendon at the PIP, DIP, and MCP joints for all digits
- Scapholunate ligament reconstruction

Prescription Use X AND/OR
(Part 21 CFR 801 Subpart D)

Over-The-Counter Use _____ (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



510(k) Summary**Cayenne Medical, Inc.**
SureLock™ All- Suture Anchor

ADMINISTRATIVE INFORMATION

Date of summary: 12/04/2014

Manufacturer Name: Cayenne Medical, Inc.
16597 N. 92nd St., Suite 101
Scottsdale, AZ 85260
Telephone (480) 502-3661
FAX (480) 502-3670

Official Contact: Shima Hashemian
16597 N. 92nd St., Suite 101
Scottsdale, AZ 85260
shashemian@cayennemedical.com
Telephone (480) 502-3661
FAX (480) 502-3670

DEVICE NAME

Classification Name: Smooth or threaded metallic bone fixation fastener

Trade/Proprietary Name: Short SureLock™ All- Suture Anchor

Common Name: Suture Anchor

DEVICE CLASSIFICATION

FDA has classified bone screws as Class II devices (21 CFR 888.3040). The product code for Fastener, Fixation, Nondegradable, and Soft Tissue is MBI. These devices are reviewed by the Orthopedic Joint Devices Branch.

INTENDED USE

The Cayenne Medical, Inc. Short SureLock™ All- Suture Anchor is intended to reattach soft tissue to bone in orthopedic surgical procedures. The system may be used in either arthroscopic or open surgical procedures. After the anchor is deployed in the bone, the floating suture with needles can be used to reattach soft tissue, such as ligaments,

tendons, or joint capsules to the bone. The suture anchor system thereby stabilizes the damaged soft tissue, in conjunction with appropriate postoperative immobilization, throughout the healing period.

The Cayenne Medical, Inc. Short SureLock™ All-Suture Anchors are intended to be used for the reattachment of soft tissue to bone for the following indications:

Foot and Ankle

- Hallux valgus reconstruction
- Midfoot reconstruction

Hand and Wrist

- Ulnar or lateral collateral ligament reconstruction
- Repair/reconstruction of collateral ligaments
- Flexor and extensor tendon at the PIP, DIP, and MCP joints for all digits
- Scapholunate ligament reconstruction

DEVICE DESCRIPTION

The Short SureLock™ All-Suture Anchor is a sterile (using ethylene oxide sterilization method), manually operated, single procedure all suture anchor device for reattachment of soft tissue to bone. The all-suture anchor is preloaded with floating suture with needles and loaded on a disposable inserter. Short SureLock™ All-Suture Anchor incorporates design features that facilitate suture anchor placement under arthroscopic, open, or limited access conditions in soft tissue to bone reattachment procedures.

The Short SureLock™ All-Suture Anchor is offered in two configurations, 1.4mm anchor with one size 0 Ultra High Molecular Weight Polyethylene (UHMWPE) suture strand with attached needles and 1.4mm anchor with one size 2-0 UHMWPE suture strand with attached needles. The anchors and floating sutures are made out of non-absorbable Ultra High Molecular Weight Polyethylene (UHMWPE) surgical sutures. The needles attached to the floating suture are made out of medical grade stainless steel.

The anchor is formed by passing one end of a length of suture perpendicularly back through itself in alternating directions a number of times. This results in a construct resembling a ladder. The four suture tails are cut and trimmed. A floating suture is passed through the loops in the anchor to form the anchor construct.



The following table summarizes the two configurations of the All-suture Anchor device.

Catalog number	All-Suture anchor size	Anchor-suture size and color	Floating suture size	Number of floating sutures	Floating suture type and color
CM-961400	1.4 mm	USP size 2 - White	USP size 0	1	Co-braid green/white
CM-961420	1.4 mm	USP size 2- White	USP size 2-0	1	Co-braid blue/white

The disposable inserter has a working shaft length of 4.14 cm with an outer shaft diameter of 4.0 mm. The inserter shaft is made out of surgical grade stainless steel and the handle and knob are made out of ABS plastic. The inserter pushes the suture anchor construct into a hole drilled in the bone. The knob on the inserter handle is rotated to apply tension to the floating suture limbs to expand and deploy the anchor in the bone tunnel. The floating suture limbs with needles are then released from the inserter and the inserter is removed.

NON-CLINICAL TESTING

Non-clinical testing data submitted, referenced, or relied upon to demonstrate substantial equivalence is included. Mechanical testing (pull-out strength) was performed on the Short SureLock™ Anchor and the predicate device. Testing showed that the Short SureLock™ Anchor ultimate pull-out strength was comparable to that of the predicate device.

EQUIVALENCE TO MARKETED PRODUCT

Cayenne Medical, Inc. demonstrated that, for the purposes of FDA's regulation of medical devices, the Short SureLock™ Anchor is substantially equivalent in indication and design principles to predicate devices, each of which has been determined by FDA to be substantially equivalent to preamendment devices: Depuy Mitek MINILOK® QuickAnchor® Plus device (K071257) and Riverpoint Medical HS Fiber Polyblend non-absorbable surgical suture (K100006). The substantial equivalence of Short SureLock™ Anchor is based on similarities in indications for use, intended use, design features, technology, and materials to the predicate device.



The subject Short SureLock™ device has the same intended use as the predicate device, the Depuy Mitek MINILOK® QuickAnchor® Plus. The subject device has a broader indication for use. The subject device is indicated for repair/reconstruction of collateral ligaments and Flexor and extensor tendon at the PIP, DIP, and MCP joints for all digits in addition to the predicate device indications for use. Cayenne Medical tested both predicate and subject devices for the range of the subject device indications using two bone block densities. The two added indications for use do not raise different questions of safety and effectiveness.

This subject device differs from the predicate device, Depuy Mitek MINILOK® QuickAnchor® Plus, in terms of the material of the anchor, the inserter role during deployment, and the offered sizes.

