



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

CarboFix Orthopedics Ltd.
Ms. Yael Rubin
Director of Regulatory Affairs
11 Ha'hoshlim Street
Herzeliya 46724
Israel

May 29, 2015

Re: K143496
Trade/Device Name: Piccolo Composite[®] Plate System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/multiple component metallic bone fixation appliances and accessories
Regulatory Class: Class II
Product Code: HRS, KTT
Dated: April 27, 2015
Received: April 28, 2015

Dear Ms. Rubin:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing

(21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation

(21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(K) Number (if known): K143496 (page 1/1)

Device Name: Piccolo Composite® Plate System

Indications for Use:

The Piccolo Composite Distal Femur Plate System is indicated for buttressing multifragmentary distal femur fractures including supra-condylar, intra-articular and extra-articular condylar fractures, periprosthetic fractures, fractures in normal or osteopenic bone, nonunions and malunions.

Prescription Use √ AND/OR Over-The-Counter Use
(Part 21 CFR 801 Subpart D) (Part 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

CarboFix Orthopedics, Ltd.
Piccolo Composite® Plate System – Distal Femur

510(K) Summary
CarboFix Orthopedics Ltd.
Piccolo Composite® Plate System

Applicant Name

CarboFix Orthopedics, Ltd.
11 Ha'hoshlim St., Herzeliya 46724, Israel

Contact Person

Yael Rubin
CarboFix Orthopedics, Ltd.
11 Ha'hoshlim St., Herzeliya 46724, Israel
Tel: +972 9 9511511, Fax: +972 9 9548939

Date Prepared

December 2014

Trade/Proprietary Name

Piccolo Composite® Plate System

Common Name

Bone Plating System

Classification Name

Single/multiple component metallic bone fixation appliances and accessories; (21 CFR §888.3030; Class II; Product Code HRS, KTT).

CarboFix Orthopedics, Ltd.**Piccolo Composite® Plate System – Distal Femur**

Predicate Devices

- Piccolo Composite® Plate System (CarboFix Orthopedics Ltd.; K102597, K120409, K130061)
- Polyax™ Distal Femoral Locked Plating System (DePuy; K060969)
- Synthes 4.5mm VA-LCP Curved Condylar Plate System (Synthes; K083025)
- Synthes LISS DF System (Synthes; K062564)

Indications for Use

The Piccolo Composite Distal Femur Plate System is indicated for buttressing multifragmentary distal femur fractures including supra-condylar, intra-articular and extra-articular condylar fractures, periprosthetic fractures, fractures in normal or osteopenic bone, nonunions and malunions.

System Description

The Piccolo Composite Distal Femur Plate System comprises implants (plates and screws), and a set of instruments.

The plates are made of carbon fiber reinforced polyetheretherketone (CFR-PEEK), and are marked with a tantalum thread, to provide for their visualization under fluoroscopy.

The screws are made of titanium alloy. Both non-locking screws and locking screws are available, in various dimensions.

Substantial Equivalence

The Piccolo Composite Distal Femur Plate System intended use, design, materials, technological characteristics, and principles of operation are substantially equivalent to those of the predicate devices, as applicable.

Performance characteristics, evaluated per ASTM F 382, are also comparable to those of predicate devices, thus demonstrating that the device is safe and effective for its intended use.
