



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

June 11, 2015

BTL Industries Incorporated  
% Dr. Michail Pankratov  
MMP Medical Associates, LLC  
16 Appleton Street  
Waltham, Massachusetts 02453

Re: K143559

Trade/Device Name: XP1000 RF  
Regulation Number: 21 CFR 878.4400  
Regulation Name: Electrosurgical cutting and coagulation device and accessories  
Regulatory Class: Class II  
Product Code: GEI  
Dated: May 6, 2015  
Received: May 7, 2015

Dear Dr. Pankratov:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

  
Jennifer R. Stevenson -S

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.  
Director  
Division of Surgical Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)

K143559

Device Name

XP1000 RF

Indications for Use (Describe)

The XP1000 RF device is indicated for temporary reduction in circumference of the abdomen.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRAStaff@fda.hhs.gov](mailto:PRAStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



K143559

## Section 5 – 510(k) Summary

### General Information

Sponsor: BTL Industries, Inc.  
47 Loring Drive  
Framingham, MA 01702  
Tel: [+1-866-285-1656](tel:+1-866-285-1656)  
Fax: +1-888-499-2502

Applicant: BTL Industries, Inc.  
47 Loring Drive  
Framingham, MA 01702  
Tel: [+1-866-285-1656](tel:+1-866-285-1656)  
Fax: +1-888-499-2502

Contact Person: Michail Pankratov  
MMP Medical Associates, LLC  
16 Appleton Street  
Waltham, MA 02453  
Tel: [+1-617-480-4543](tel:+1-617-480-4543)  
Fax: +1 781 207 4676

Summary Preparation  
Date: June 3, 2015

### Device Names

Trade/Proprietary Name: XP1000 RF

Primary Classification Name: Electrosurgical Cutting and Coagulation Device & Accessories

Classification Regulation: 878.4400

Product Code: GEI

### Legally Marketed Predicate Devices

The XP1000 RF system is a state-of-the-art high-frequency energy device with accessory, and is substantially equivalent to the current product that is already cleared for USA distribution under the following 510(k) Premarket Notification number:

- Transcend System (K120510)



## Product Description

The XP1000 RF system is designed for temporary reduction in circumference of the abdomen by means of high-frequency electromagnetic field.

The multi-jointed arm firmly supports the applicator during treatment. Large knobs lock and unlock the joints to make positioning the applicators to the patient quick, easy and secure.

The control unit consists of the control system and the electronic system. The control system contains the main microcomputer and software for control of the entire equipment; the electronic system contains the complete electronics for electromagnetic field generation. Easy-to-use color touch screen allows for maximum operator comfort. A large control knob is provided to increase and decrease output power.

The XP1000 RF system is placed in a specially designed cart, the shape of which provides maximum operator comfort and easy movement of the device in the office.

The XP1000 RF consists of the following main components:

- microprocessor-driven control unit
- high-frequency electromagnetic energy generator
- user interface with 8.4" color touch screen
- applicator on a multi-jointed arm

## Indications for Use

The XP1000 RF is indicated for temporary reduction in circumference of the abdomen.

## Non-clinical Testing

The XP1000 RF device has been thoroughly evaluated for electrical safety and performance. The XP1000 RF has been found to conform with applicable medical device safety standards. The system complies with the following standards:

- ISO 14971 – Medical devices – Application of risk management to medical devices
- IEC 62304 – Medical Device Software – Software Life Cycle Processes

### Medical Electrical Equipment

- IEC 60601-1 General requirements for safety
- IEC 60601-1-2 Electromagnetic compatibility–Requirements and Tests
- IEC 60601-1-6 Usability
- IEC 60601-2-2 Particular requirements for basic safety and performance of high frequency surgical equipment
- ISO 10993-1 Evaluation and testing within a risk management process
- ISO 10993-5 Biological Evaluation of Medical Devices–Tests for In Vitro toxicity



ISO 10993-10 Biological Evaluation of Medical Devices–Test for Irritation and Skin Sensitization

**Clinical testing**

The device was clinically tested for reduction of abdominal circumference by disruption of adipocyte cells.

Sixty subjects, ranging from 19 to 53 years of age, were enrolled in the double blinded study and divided into 2 groups: 40 to the XP1000 RF Treatment Group and 20 to the Sham (Placebo) Group. All subjects but one finished the study.

The Treatment Group subjects have met the primary efficacy outcome measure by achieving  $\geq 3$  cm waist circumferential reduction in more than 80% of subjects. The mean circumferential reduction was  $\geq 1$  cm than the mean circumferential reduction of the Placebo Group at the 30-day Follow Up evaluation point.

Waist circumference changes (Baseline data vs the 30-day and 90-day Follow Up data) between the XP1000 RF Treatment subjects and the Placebo Group were statistically highly significant.

The secondary objective was absence of adverse events (AE) associated with the treatment procedure. The secondary objective was met with no adverse events recorded.

The clinical study supports the effectiveness of the device for temporary reduction of abdominal circumference.

The efficacy of XP1000 RF was not established at power level settings below maximum power of 200 points. The device is indicated for patients 22 years or older because its safety and effectiveness have not been established in pediatric patients or patients younger than 22 years.

**Comparison with the Predicate Device**

<b>510(k) number</b>	<b>K143559</b>	<b>K120510</b>
<b>Device name</b>	XP1000 RF	Transcend System
<b>Company name</b>	BTL Industries, Inc.	Syneron Medical Ltd.
<b>Product Code Regulation</b>	<u>General &amp; Plastic Surgery</u> 21 CFR 878.4400 -GEI, Electrosurgical Cutting and Coagulation Device &	<u>General &amp; Plastic Surgery</u> 21 CFR 878.4400 -GEI, Electrosurgical Cutting and Coagulation Device &



<b>510(k) number</b>	<b>K143559</b>	<b>K120510</b>
<b>Device name</b>	XP1000 RF	Transcend System
<b>Company name</b>	BTL Industries, Inc.	Syneron Medical Ltd.
	Accessories.	Accessories.
<b>Indications for Use</b>	The XP1000 RF is indicated for temporary reduction in circumference of the abdomen.	The Transcend is indicated for temporary reduction in circumference of the abdomen.
<b>Deep tissue Heating Electromagnetic Energy</b>	Radiofrequency	Radiofrequency + InfraRed
<b>Vacuum</b>	No	Yes
<b>Electrical Protection</b>	Class II, BF	Class I, BF
<b>Unit Construction</b>	Constructed of materials that conform with safety standards and requirements.	Constructed of materials that conform with safety standards and requirements.
<b>Interface</b>	Touch-screen user applied interface to program and set the controls for the patient application; there is an applicator utilized to deliver the treatment.	Touch-screen user applied interface to program and set the controls for the patient application; there is a hand-piece utilized to deliver the treatment.
<b>Color Touch Screen</b>	8.4" (21.5cm)/640x480 pixel	10.1" (25.7 cm)
<b>Modes Of Operation</b>	Bipolar	Bipolar



<b>510(k) number</b>	<b>K143559</b>	<b>K120510</b>
<b>Device name</b>	XP1000 RF	Transcend System
<b>Company name</b>	BTL Industries, Inc.	Syneron Medical Ltd.
<b>Nominal Operating Power</b>	Max: 200W RF	Max: 151.7W (150W RF, 1.7 or 3.3W IR)
<b>RF Carrier Frequency</b>	27.12 MHz (±400kHz)	1 MHz
<b>Operating Temperature</b>	10°C to 30°C	15°C to 30°C
<b>Operating Humidity</b>	30% - 75%	Up to 80%
<b>Skin Temperature Monitoring</b>	Based on patient's feedback. External IR thermometer.	Based on patient's feedback. Built-in IR thermometer.
<b>Treatment Temperature Range</b>	40°C – 45°C	40°C – 45°C
<b>Cooling</b>	Not Available	Not Available
<b>Cooling Mechanism</b>	Not Available	Not Available
<b>Applicator Holder Availability</b>	YES	YES
<b>Waveform</b>	Sinusoid	Sinusoid
<b>Applicator Effective Area</b>	20,106 mm <sup>2</sup>	1,250 mm <sup>2</sup> 750 mm <sup>2</sup> 130 mm <sup>2</sup>
<b>Treatment Time</b>	Up to 45 min. per treated area (waist)	Up to 45 min. per treated area (waist)
<b>Total Energy Density (fluence)</b>	3.6 - 27 J/mm <sup>2</sup>	3.6 J/mm <sup>2</sup> 6.0 J/mm <sup>2</sup> 34.6 J/mm <sup>2</sup>
<b>Material of the Generator Case</b>	Plastic/metal	Plastic/metal
<b>Dual Dispersive Patch Electrode Grounding</b>	NO	NO
<b>Patch Electrode Contact Quality Monitoring</b>	YES	N/A





<b>510(k) number</b>	<b>K143559</b>	<b>K120510</b>
<b>Device name</b>	XP1000 RF	Transcend System
<b>Company name</b>	BTL Industries, Inc.	Syneron Medical Ltd.
<b>RF Energy Emission Indicator</b>	YES; Information displayed on the screen of the unit	YES
<b>Applicator Weight</b>	3.9 lbs (1760 g)	2.2 lbs (1 kg)
<b>Energy Source</b>	100 – 240 VAC, max 5A, 50 – 60Hz	110 – 230 VAC, max 5A, 50Hz
<b>External Exchangeable Fuse</b>	2 × T4A / 250V, tube fuse 5 × 20 mm	Available
<b>System Dimensions</b>	22" × 39" × 22" (56 cm x 98 cm x 56 cm)	15" × 19" × 51.8" (132 cm × 49 cm × 38 cm)
<b>System Weight</b>	83 lbs (38 kg)	44.1 lbs (20 kg)

### Substantial Equivalence

Based upon the intended use and technical information provided in this pre-market notification, the XP1000 RF device has been shown to be substantially equivalent to currently marketed predicate device.

### Conclusion

Based on the aforementioned information, the XP1000 RF is safe and effective and substantially equivalent to the identified predicate device.