



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

APEX BIO TECHNOLOGY CORP
HSUE-MEI LEE, MANAGER OF QUALITY ASSURANCE DEPARTMENT
NO. 7, LI-HSIN ROAD V
HSINCHU SCIENCE PARK
HSINCHU 30078, TAIWAN

May 7, 2015

Re: K143750

Trade/Device Name: MEG-2B Blood Glucose Monitoring System,
MEG-2B Pro Blood Glucose Monitoring System

Regulation Number: 21 CFR 862.1345

Regulation Name: Glucose test system

Regulatory Class: II

Product Code: NBW, CGA

Dated: April 2, 2015

Received: April 8, 2015

Dear Hsue-Mei Lee:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulations (21 CFR Parts 801 and 809), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Katherine Serrano -S

For : Courtney H. Lias, Ph.D.
Director
Division of Chemistry and Toxicology Devices
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K143750

Device Name
MEG-2B Blood Glucose Monitoring System

Indications for Use (Describe)

The MEG-2B Blood Glucose Monitoring System, includes MEG-2B Blood Glucose Meter and MEG-2B Blood Glucose Test Strip, is intended for the quantitative measurement of glucose in fresh capillary whole blood samples drawn from the fingertips, forearm, or palm. Alternate site testing should be done only during steady-state times (when glucose is not changing rapidly). It is indicated for use by lay people with diabetes at home as an aid to monitor the effectiveness of diabetes control. This system is intended for self-testing outside the body (in vitro diagnostic use) and should not be shared. It is not indicated for the diagnosis or screening of diabetes or for neonatal use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Indications for Use

510(k) Number (if known)
K143750

Device Name
MEG-2B Pro Blood Glucose Monitoring System

Indications for Use (Describe)

The MEG-2B Pro Blood Glucose Monitoring System includes MEG-2B Pro Blood Glucose Meter and MEG-2B Pro Blood Glucose Test Strip, is intended for the quantitative measurement of glucose in fresh capillary whole blood samples drawn from the fingertips, forearm, or palm. Alternate site testing should be done only during steady -state times (when glucose is not changing rapidly). This system is indicated for multiple-patient use in professional healthcare settings as an aid to monitor the effectiveness of diabetes control programs. It is only used with single-use, auto-disabling lancing device. It is intended for testing outside the body (in vitro diagnostic use). It is not indicated for the diagnosis or screening of diabetes or for neonatal use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Submitter	<p>Hsue-mei Lee Manager of Quality Assurance Department Apex BioTechnology Corp. No. 7, Li-Hsin Road V, Hsinchu Science Park Hsinchu, 30078 CHINA (TAIWAN)</p> <p>email: hsue-mei@apexbio.com Phone: 011-886-3-5641952 FAX: 011-886-3-5678302</p>
Contact Person	<p>Hsue-mei Lee Manager of Quality Assurance Department Apex BioTechnology Corp. No. 7, Li-Hsin Road V, Hsinchu Science Park Hsinchu, 30078 CHINA (TAIWAN)</p> <p>email: hsue-mei@apexbio.com Phone: 011-886-3-5641952 FAX: 011-886-3-5678302</p>
Date Prepared	March 31, 2015
Trade Names	MEG-2B Blood Glucose Monitoring System MEG-2B Pro Blood Glucose Monitoring System
Classification	Glucose test system, 21 CFR 862.1345, Class II
Product Codes	CGA, NBW
Predicate Devices	MEG-2B Blood Glucose Monitoring System (k120448) MEG-2B Pro Blood Glucose Monitoring System (k120448)
Device Description	The MEG-2B Blood Glucose Monitoring System consists of the MEG-2B meter and MEG-2B Test Strips. It is used for testing of blood glucose by self-testers at home. The MEG-2B Pro Blood Glucose Monitoring System consists of the MEG-2B Pro meter and MEG-2B Pro Test Strips. It is used for testing of blood glucose by professional testers in healthcare facilities. The MEG-2B and MEG-2B Pro systems are identical other than trade names and details of product labeling.

510(k) Summary (Continued)

<p>Intended Use</p>	<p>MEG-2B Blood Glucose Monitoring System: The MEG-2B Blood Glucose Monitoring System, includes MEG-2B Blood Glucose Meter and MEG-2B Blood Glucose Test Strip, is intended for the quantitative measurement of glucose in fresh capillary whole blood samples drawn from the fingertips, forearm, or palm. Alternate site testing should be done only during steady -state times (when glucose is not changing rapidly). It is indicated for use by lay people with diabetes at home as an aid to monitor the effectiveness of diabetes control. This system is intended for self-testing outside the body (in vitro diagnostic use) and should not be shared. It is not indicated for the diagnosis or screening of diabetes or for neonatal use.</p> <p>MEG-2B Pro Blood Glucose Monitoring System: The MEG-2B Pro Blood Glucose Monitoring System includes MEG-2B Pro Blood Glucose Meter and MEG-2B Pro Blood Glucose Test Strip, is intended for the quantitative measurement of glucose in fresh capillary whole blood samples drawn from the fingertips, forearm, or palm. Alternate site testing should be done only during steady -state times (when glucose is not changing rapidly). This system is indicated for multiple-patient use in professional healthcare settings as an aid to monitor the effectiveness of diabetes control programs. It is only used with single-use, auto-disabling lancing device. It is intended for testing outside the body (in vitro diagnostic use). It is not indicated for the diagnosis or screening of diabetes or for neonatal use.</p>
<p>Comparison of Technological Characteristics</p>	<p>The modified MEG-2B and MEG 2B Pro Blood Glucose Monitoring Systems are identical to the predicate other than a) addition of silver paint to the meter case, b) change in manufacturing site for the LCD cover, and c) the recommendation of three additional disinfectants in the User Manuals.</p>
<p>Non-Clinical Testing</p>	<p>Disinfection (viral inactivation) and “robustness” testing were done to qualify all recommended disinfection solutions. Results demonstrate substantial equivalence to the predicate.</p>
<p>Clinical Testing</p>	<p>No clinical testing was conducted.</p>
<p>Conclusion</p>	<p>Testing showed that the modified MEG-2B and MEG 2B Pro Blood Glucose Monitoring Systems are substantially equivalent to the predicate.</p>