



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Extremity Medical, LLC
Mr. Brian Smekal
Vice President, Regulatory Affairs and Quality Assurance
300 Interpace Parkway, Suite 410
Parsippany, New Jersey 07054

March 4, 2015

Re: K150309

Trade/Device Name: Extremity Medical Calcaneal Osteotomy Device
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: HWC, HRS
Dated: February 6, 2015
Received: February 9, 2015

Dear Mr. Brian Smekal:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K150309

Device Name

Extremity Medical Calcaneal Osteotomy Device

Indications for Use (Describe)

The Extremity Medical Calcaneal Osteotomy Device is intended for fixation of osteotomies of the calcaneus.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Special 510(k) Summary of Safety and Effectiveness:

Calcaneal Osteotomy Device

Submitter:	Extremity Medical, LLC. 300 Interpace Parkway Suite 410 Parsippany, NJ 07054
Contact Person	Brian Smekal, MS, RAC Director, Regulatory Affairs Phone: (973) 588-8988 Email: bsmekal@extremitymedical.com
Date Prepared	February 6, 2015
Trade Name	Calcaneal Osteotomy Device
Classification Name and Number	Smooth or threaded metallic bone fixation fastener 21 CFR 888.3040
Product Code	HWC (screw, fixation, bone), HRS (plate, fixation, bone)
Predicate Devices	K143402 – Calcaneal Osteotomy Device (Extremity Medical)
Device Description	The Calcaneal Osteotomy Device is a bone fixation device consisting of a bracket and lag screws and a set of instruments used for implant site preparation and delivery. The bracket is offered in height sizes of 6mm to 13mm. The modified device allows for placement of the implant with 3 lag screws. The Calcaneal Osteotomy Device is intended for fixation of osteotomies of the calcaneus.
Indications for use	The Extremity Medical Calcaneal Osteotomy Device is intended for fixation of osteotomies of the calcaneus.
Statement of Technological Comparison	The Calcaneal Osteotomy Device and predicate devices are equivalent in terms of indications for use, design, and material mechanical properties.
Non-clinical Testing	Pull-out, torque and bending strength was assessed compared to the predicate device and worse case constructs have been tested to demonstrate equivalence.
Clinical Testing	No clinical testing was performed.
Conclusion	The Calcaneal Osteotomy Device is substantially equivalent to its predicate device. This conclusion is based upon indications for use, materials, design, test data and principles of operation.