



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

May 22, 2015

MicroAire Surgical Instruments LLC  
Mr. Chris Spofford  
Regulatory Affairs and Compliance Officer  
3590 Grand Forks Boulevard  
Charlottesville, Virginia 22911

Re: K150779  
Trade/Device Name: Microaire LipoFilter System  
Regulation Number: 21 CFR 878.5040  
Regulation Name: Suction lipoplasty system  
Regulatory Class: Class II  
Product Code: MUU  
Dated: March 25, 2015  
Received: March 26, 2015

Dear Mr. Spofford:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**David Krause -S**

for Binita S. Ashar, M.D., M.B.A., F.A.C.S.  
Director  
Division of Surgical Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K150779

Device Name

Microaire LipoFilter System

Indications for Use (Describe)

The MicroAire LipoFilter System is used in the aspiration, harvesting, filtering and transferring of autologous adipose tissue for aesthetic body contouring. If the harvested fat is to be re-implanted, the harvested fat is only to be used without any additional manipulation.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

---

## 510(k) Summary

- 1) Preparation Date: March 24, 2015
- 2) Submitted by: MicroAire Surgical Instruments LLC  
3590 Grand Forks Boulevard  
Charlottesville, Virginia 22911  
Owner/Operator #: 9004658
- 3) Contact Person/Prepared by: Chris Spofford  
Regulatory Affairs and Compliance Officer  
MicroAire Surgical Instruments  
(434) 975-8344  
[Chriss@microaire.com](mailto:Chriss@microaire.com)  
3590 Grand Forks Boulevard  
Charlottesville, Virginia 22911
- 4) Device Identification:  
  
Trade Name: MicroAire LipoFilter System  
Common Name: Lipoplasty Suction System  
Classification: Suction lipoplasty system (21 CFR 878.5040, Product Code MUU)
- 5) Predicate Device:  
Sound Surgical Technologies – Origins LipoHarvesting System (K120328)  
Medical Device Resource Corporation – Lipisystems Aquavage Model AV2000 & 1200 (K092284)
- 6) Device Description: The MicroAire LipoFilter System consists of a single use, closed loop tissue collection system comprised of a medical grade canister, vacuum port, collection port, tissue port and lid intended to be used with a standard liposuction aspiration system pump to collect fatty tissue for aesthetic body contouring. As the tissue is harvested from the patient, it enters the canister via the collection port in the canister lid. The physician removes unwanted waste materials from the collection system via the vacuum port by closing a valve. This process leaves fatty tissue that can be transferred to syringes via the tissue port for autologous fat re-injection.
- 7) Intended Use: The MicroAire LipoFilter System is used in the aspiration, harvesting, filtering and transferring of autologous adipose tissue for aesthetic body contouring. If the harvested fat is to be re-implanted, the harvested fat is only to be used without any additional manipulation.

8) Comparison to Predicate:

Table 1: Substantial Equivalence Comparison Chart

FEATURES	DEVICES		Medical Device Resource Corp Lipisystems Aquavage
	MicroAire LipoFilter System	Origins LipoHarvesting System	
Device Description:	The MicroAire LipoFilter System consists of a single use, closed loop tissue collection system comprised of a medical grade canister, vacuum port, collection port, tissue port and lid intended to be used with a standard liposuction aspiration system pump to collect fatty tissue for aesthetic body contouring. As the tissue is harvested from the patient, it enters the canister via the collection port in the canister lid. The physician removes unwanted waste materials from the collection system via the vacuum port by closing a valve. This process leaves fatty tissue that can be transferred to syringes via the tissue port for autologous fat re-injection.	The Origins LipoHarvesting System consists of a reusable, closed loop tissue collection system comprised of a medical grade canister, a drain port and a lid intended to be used with a standard liposuction aspiration pump to collect fatty tissue for aesthetic body contouring. As the tissue is harvested from the patient, it enters the collection canister via a port in the canister lid. The physician removes unwanted waste materials from the collection system via the drain port at the base. This process leaves fatty tissue that can be transferred to syringes for autologous fat re-injection.	<p>Device function:</p> <ol style="list-style-type: none"> <li>1. Contains port interfaces between: canister to tubing and aspirator to canister.</li> <li>2. Sterile tubing to connect the interfaces.</li> <li>3. Funnel to interface port to tubing.</li> <li>4. Syringe to collect fat.</li> </ol> <p>Device Design:</p> <ol style="list-style-type: none"> <li>1. Contents subjected to sterility.</li> </ol> <p>Material used: Plastic canister, syringe, silicone tubing.</p>
Picture of Product			

Intended Use:	The MicroAire LipoFilter System is used in the aspiration, harvesting, filtering and transferring of autologous adipose tissue for aesthetic body contouring. If the harvested fat is to be re-implanted, the harvested fat is only to be used without any additional manipulation.	The Origins LipoHarvesting System is used in the aspiration, harvesting, filtering and transferring of autologous tissue. If harvested fat is to be re-implanted, the harvested fat is only to be used without any additional manipulation.	For use in aspirating subcutaneous fatty tissue including autologous fat collection.
Technology Comparison:	The MicroAire LipoFilter System employs the same technological characteristics as the predicate device.	The Origins LipoHarvesting System employs the same technological characteristics as the predicate devices.	The Lipisystems Aquavage employs the same technological characteristics as the predicate device.
Suction Source(s)	Aspiration device	Aspiration device	Aspiration device
Volume Range	Up to 2500mL	100 to 2,000 mL	2000cc or 1200cc
Shipped sterile	Yes	No, sterilized by user prior to use	Yes
Initial method	Gamma radiation	Not applicable	Gamma radiation
Sterility assurance level	10 <sup>-6</sup>	10 <sup>-6</sup>	10 <sup>-6</sup>
Disposable or reusable	Single use, disposable	Reusable	Single use, disposable
Resterilization method	Not applicable	Autoclave	Not applicable

## 9) Pre-Clinical Testing

The MicroAire LipoFilter System has been tested in various ways to demonstrate equivalence to the predicate devices including Gamma Radiation validation, shelf life validation, biocompatibility, and various performance testing. Specifics include the following:

Validated for sterility using gamma radiation. Test results and analyses indicate that the MicroAire LipoFilter System sterility complies in accordance with ISO 11137-1:2006 and ISO 11137-2:2006.

Validated for a 2 year shelf life. Test results and analyses indicate that the MicroAire LipoFilter System packaging complies in accordance with ISO 11607-1: 2006 and ISO 11607-2: 2006.

The MicroAire LipoFilter System has patient contact materials and is made from medical grade biocompatible materials. Test results and analyses indicate that the MicroAire LipoFilter System materials' complies in accordance with ISO 10993-1: 2009.

The MicroAire LipoFilter System was tested for performance in accordance with its predetermined specifications on file at MicroAire which includes functionality testing, ISO 10079-1 testing, tubing connection strength testing, vacuum leak testing, adipocyte viability testing, and Aging Studies.

#### 10) Conclusion:

The MicroAire LipoFilter System uses decantation to separate the tissue which is the same method as the two identified predicate devices. The MicroAire LipoFilter System does not have any new intended uses as compared to the two predicate devices. The MicroAire LipoFilter System has a slightly larger volume (2500 mL) whereas the predicate devices have a volume of up to 2000 mL which is insignificant given the intended use of the devices. The MicroAire LipoFilter System is provided sterile whereas the predicate devices are either provided sterile and/or sterilized by the user. The results of these activities demonstrate that the MicroAire LipoFilter System is considered substantially equivalent to the two predicate devices.

The results of the evaluation above demonstrate that the MicroAire LipoFilter System is considered substantially equivalent to the two predicate devices.