



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

November 18, 2015

Andersen Sterilizers, Inc.  
Dr. William Andersen  
President  
3154 Caroline Drive  
Haw River, North Carolina 27258

Re: K151585  
Trade/Device Name: Andersen EOgas 4 Steritest  
Regulation Number: 21 CFR 880.2800  
Regulation Name: Sterilization Process Indicator  
Regulatory Class: Class II  
Product Code: FRC  
Dated: October 20, 2015  
Received: October 22, 2015

Dear Dr. William Andersen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Erin I. Keith -S**

Erin I. Keith, M.S.

Director

Division of Anesthesiology,

General Hospital, Respiratory,

Infection Control, and Dental Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K151585

Device Name

Andersen EOGas 4 SteriTTest

Indications for Use (Describe)

The Andersen EOGas 4 SteriTTest consists of a self-contained biological indicator inoculated with viable *Bacillus atrophaeus* bacterial spores that is placed in a dedicated BI receptacle in the sterilizer. It monitors the efficacy of the 3 hour sterilization cycle at 50°C in the EOGas 4 Ethylene Oxide Gas Sterilizer.

Critical process parameters for the cycle are summarized below in Table 1.

Table 1. Critical parameters for the 3 hour cycle in the EOGas 4 Ethylene Oxide Gas Sterilizer

| EO Amount   | Temperature | Relative Humidity | EO Exposure Time | Total Cycle Time |
|-------------|-------------|-------------------|------------------|------------------|
| 17.6 g ± 5% | 50°C ± 3°C  | 35-90%            | 3 hours          | 3.5 hours        |

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

**FOR FDA USE ONLY**

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

---

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
*PRAStaff@fda.hhs.gov*

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*

**510(k) Summary**

|     |                              |   |
|-----|------------------------------|---|
| 5.1 | Applicant's Name and Address | Andersen Sterilizers, Inc.<br>3154 Caroline Drive<br>Haw River, NC 27258                  |
| 5.2 | Contact Person               | William K. Andersen, BE, MD, FAAOS<br>President<br>Phone: 336-376-8622, Fax: 336-376-5428 |
| 5.3 | Date of Preparation          | November 17, 2015   |
| 5.4 | Device                       | Andersen EOGas 4 <sup>®</sup> SteriTest   |
|     | Proprietary Name             | Biological Sterilization Process Indicator  |
|     | Common Name                  | Class II (21 CFR 880.2800)  |
|     | Classification               | FRC   |
|     | Product Code                 |   |
| 5.5 | Predicate Device             |   |
|     | Device Name                  | EZTest - Gas Biological Indicators  |
|     | 510(k) number                | K930683   |
|     | Manufacturer                 | Mesa Laboratories Inc.  |

This 510(k) submission modifies the indications for use of the predicate device in order to include it as a component in a process challenge device. No modifications were made to the technology or intended use.

**5.6 Device Description**

The Andersen EOGas 4<sup>®</sup> SteriTest consists of a single-use self-contained biological indicator (SCBI) placed in a reusable BI receptacle. It is designed for monitoring the efficacy of the 3 hour sterilization cycle at 50°C in an EOGas 4 Ethylene Oxide Gas Sterilizer.

The SCBI, the EZTest - Gas Biological Indicator, consists of a plastic vial that serves as the culture tube and a cap including a filter material port to allow ethylene oxide to enter the vial. The plastic vial contains *Bacillus atrophaeus* spores inoculated onto a paper carrier, and a glass ampoule containing modified soybean casein digest broth and phenol red acting as a pH indicator. There is a chemical indicator printed on the unit label of the SCBI to indicate EO exposure.

Following manufacturer's instructions, the operator inserts the Andersen EOGas 4 SteriTest into the reusable BI receptacle on the purge probe of the EOGas 4 Ethylene Oxide Gas Sterilizer, and initiates a 3 hour cycle at 50°C. After cycle completion, the SCBI is retrieved and activated by crushing the glass ampoule. The chemical indicator on the SCBI changes from blue to a green/brown color depending on the duration of ethylene oxide exposure. The activated SCBI and an unprocessed control are incubated at 30-35°C for 48 hours, and monitored for any color change and/or turbidity. Evidence of microbial growth by color change from red-orange to

yellow and/or turbidity must be interpreted as a failure to meet the conditions necessary for sterilization (cycle failed); no color change or turbidity indicates conditions for sterilization were achieved (cycle passed).

### 5.7 Indications for Use

The Andersen EOGas 4 SteriTTest consists of a self-contained biological indicator inoculated with viable *Bacillus atrophaeus* bacterial spores that is placed in a dedicated BI receptacle in the sterilizer. It monitors the efficacy of the 3 hour sterilization cycle at 50°C in the EOGas 4 Ethylene Oxide Gas Sterilizer. Critical process parameters for the cycle are summarized in **Table 5-1**.

**Table 5-1.** Critical parameters for the 3 hour cycle in the EOGas 4 Ethylene Oxide Gas Sterilizer

| Ethylene Oxide | Temperature | Relative Humidity | Ethylene Oxide Exposure Time | Total Cycle Time |
|----------------|-------------|-------------------|------------------------------|------------------|
| 17.6 g ± 5%    | 50°C ± 3°C  | 35-90%            | 3 hours                      | 3.5 hours        |

### 5.8 Device Comparison

The technological characteristics of the Andersen EOGas 4 SteriTTest are substantially equivalent to the predicate device - EZTest - Gas Biological Indicators, because both indicators are intended for the same use, use the same technology, are designed in the same way, and perform substantially equivalently.

The Andersen EOGas 4 SteriTTest differs from the predicate device in the configuration used in a load. The difference raises no issues related to safety or effectiveness of the subject device in the sterilization cycle. A comparison between the devices is listed in **Table 5-2**.

**Table 5-2.** Device Comparison

| Element                   | EZTest - Gas BI   | EOGas 4 SteriTTest  |
|---------------------------|---|---|
| Intended Use              | Sterilization method: EO gas<br>Process parameters: EO concentration, time, temperature, and relative humidity  | Identical   |
| Organism                  | <i>Bacillus atrophaeus</i> (ATCC 9372)  | Identical   |
| Viable Spore Population   | $\geq 1.0 \times 10^6$  | Identical   |
| Device Design             | Paper strip containing indicator organism;<br>Glass ampoule containing growth medium;<br>Capped vial serving as a culture tube;<br>A pH indicator in medium for color change;<br>A process indicator indicating EO exposure | Identical for the BI;<br>Additionally, a BI receptacle is added to create a greater challenge to the sterilization process. |
| Materials of Construction | Paper, glass, and polypropylene   | Paper, glass, polypropylene, and aluminum   |
| Configuration in Load     | SCBI placed within the load   | SCBI in a receptacle  |

|                     |               |  |
|---------------------|---------------|--|
| Indications for Use | Not available | The Andersen EOGas 4 SteriTTest consists of a self-contained biological indicator inoculated with viable <i>Bacillus atrophaeus</i> bacterial spores that is placed in a dedicated BI receptacle in the sterilizer. It monitors the efficacy of the 3 hour sterilization cycle at 50°C in the EOGas 4 Ethylene Oxide Gas Sterilizer. |
|---------------------|---------------|--|

## 5.9 Performance Testing

The Andersen EOGas 4 SteriTTest has been validated using applicable tests in FDA 2007 “Guidance for Industry and FDA Staff: Biological Indicator (BI) Premarket Notification [510(k)] Submissions”, and AAMI/ANSI/ISO 11138-1 “Sterilization of health care products – Biological indicators - Part 1: General requirements” (FDA Recognition Number 14-296).

For the EZTest - Gas Biological Indicators, tests include viable spore population assay, resistance characteristics study, carrier and primary packaging materials (growth inhibition) evaluation, holding time assessment, and reduced incubation time validation. The results of all studies met the established acceptance criteria when applicable.

We also demonstrated that under the same exposure conditions, the resistance characteristics of the EZTest - Gas Biological Indicators were the same when measured in a Biological Indicator Evaluator Resistometer in the presence or absence of vacuum.

The Andersen EOGas 4 SteriTTest represents a rigorous challenge to the EOGas 4 sterilization process. Its resistance characteristics are greater than the same EZTest-Gas Biological Indicator placed in the worst-case locations of the maximum fabric, metal, plastic, and endoscope validation loads. The performance of the Andersen EOGas 4 SteriTTest in the 3 hour cycle at 50°C in an EOGas 4 Ethylene Oxide Gas Sterilizer is summarized in **Table 5-3**.

**Table 5-3.** Summary of bench tests performed to demonstrate safety and effectiveness of the Andersen EOGas 4 SteriTTest

| Test          | Description  | Result |
|---------------|--|--------|
| Functionality | 1) Critical parameters include temperature, time, and gas concentration under a relative humidity of 35-90%;<br>2) Device is appropriate for monitoring the efficacy of the sterilization process claimed                  | Pass   |
| Shelf Life    | Maintains performance specifications (resistance characteristics and correctly indicate pass/fail in cycles) throughout the stated shelf life of 2 years;<br>Stability demonstrates reasonable assurance for effectiveness | Pass   |

In conclusion, the Andersen EOGas 4 SteriTTest is substantially equivalent to the legally marketed predicate, the EZTest - Gas Biological Indicator.