



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

NuVasive, Incorporated
Ms. Olga Lewis
Lead Specialist, Regulatory Affairs
7475 Lusk Boulevard
San Diego, California 92121

September 25, 2015

Re: K151651

Trade/Device Name: NuVasive® X-Core® Mini Cervical Expandable VBR System
Regulation Number: 21 CFR 888.3060
Regulation Name: Spinal intervertebral body fixation orthosis
Regulatory Class: Class II
Product Code: PLR
Dated: August 26, 2015
Received: August 27, 2015

Dear Ms. Lewis:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K151651

Device Name

NuVasive® X-Core® Mini Cervical Expandable VBR System

Indications for Use (Describe)

The NuVasive® X-Core® Mini Cervical Expandable VBR System is a vertebral body replacement device indicated for use in the cervical spine (C3-C7 vertebral bodies) in skeletally mature patients to replace a diseased or damaged vertebral body caused by tumor, fracture, or osteomyelitis, or for reconstruction following corpectomy performed to achieve decompression of the spinal cord and neural tissues in cervical degenerative disorders. The NuVasive® X-Core® Mini Cervical Expandable VBR System is intended to be used with supplemental fixation cleared by the FDA for use in the cervical spine.

These implants are intended for use with autograft or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft, as an adjunct to fusion. The NuVasive® X-Core® Mini Cervical Expandable VBR System is also intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the cervical spine in whom life expectancy is of insufficient duration to permit achievement of fusion, with bone graft used at the surgeon's discretion.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) Summary

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular 21 CFR §807.92, the following summary of information is provided:

A. Submitted by:

Olga Lewis
Lead Specialist, Regulatory Affairs
NuVasive, Incorporated
7475 Lusk Blvd.
San Diego, California 92121
Telephone: (858) 909-1800

Date Prepared: August 26, 2015

B. Device Name

Trade or Proprietary Name: *NuVasive® X-Core® Mini Cervical Expandable VBR System*
Common or Usual Name: Spinal Vertebral Body Replacement Device
Classification Name: Spinal Intervertebral Body Fixation Orthosis

Device Class: Class II
Classification: 21 CFR § 888.3060
Product Code: PLR

C. Predicate Devices

The subject *NuVasive® X-Core® Mini Cervical Expandable VBR System* is substantially equivalent to the primary predicate device *NuVasive® X-Core® Expandable VBR System* (K142205) and the additional predicate device, *NuVasive® CoRoent® System* (K081611).

D. Device Description

The *X-Core® Mini Cervical Expandable VBR System* is manufactured from Ti-6Al-4V ELI conforming to ASTM F136 and ISO 5832-3. The implants are available in a variety of different shapes and sizes to suit the individual pathology and anatomical conditions of the patient. This 510(k) is to expand the indications for use of *NuVasive® X-Core® Mini Expandable VBR System* (K142205) to include treatment of tumors, trauma, and degenerative disorders of the cervical spine.

E. Indications for Use

The *NuVasive® X-Core® Mini Cervical Expandable VBR System* is a vertebral body replacement device indicated for use in the cervical spine (C3-C7 vertebral bodies) in skeletally mature patients to replace a diseased or damaged vertebral body caused by tumor, fracture, or osteomyelitis, or for reconstruction following corpectomy performed to achieve decompression of the spinal cord and neural tissues in cervical degenerative disorders. The *NuVasive® X-Core® Mini Cervical Expandable VBR System* is intended to be used with supplemental fixation cleared by the FDA for use in the cervical spine.

These implants are intended for use with autograft or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft, as an adjunct to fusion. The *NuVasive® X-Core® Mini Cervical Expandable VBR System* is also intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the cervical spine in whom life expectancy is of insufficient duration to permit achievement of fusion, with bone graft used at the surgeon's discretion.

F. Technological Characteristics

As was established in this submission, the subject *NuVasive® X-Core® Mini Cervical Expandable VBR System* is substantially equivalent to other predicate devices cleared by the FDA for commercial distribution in the United States. The subject device was shown to be substantially equivalent based on clinical data and have the same technological characteristics to its predicate devices through comparison in areas including design, intended use, performance, material composition, and function.

G. Performance Data

The purpose of this 510(k) is to modify the Indications for Use for the subject *NuVasive® X-Core® Mini Cervical Expandable VBR System*. Clinical performance data on the subject device, as well as clinical literature review and retrospective clinical data analysis was performed to support the use of the subject device for expanded indications for use. Based on the published clinical literature review and retrospective clinical data analysis, it was determined that the *NuVasive® X-Core® Mini Cervical Expandable VBR System* used in the treatment of tumors, trauma, and degenerative disorders of the cervical spine has a safety and effectiveness profile that is similar to the predicate device. No other changes have been made to the system design since its clearance in the *NuVasive® X-Core® Expandable VBR System* 510(k) K142205. Therefore, no new nonclinical testing was performed for the purpose of this submission.

H. Conclusions

Based on the indications for use, technological characteristics, mechanical testing, and comparison to predicate devices, the subject *NuVasive® X-Core® Mini Cervical Expandable VBR System* has been shown to be substantially equivalent to legally marketed predicate devices.
