



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Toshiba Medical Systems Corporation
% Orlando Tadeo, Jr.
Manager, Regulatory Affairs
Toshiba America Medical Systems, Inc.
2441 Michelle Drive
TUSTIN CA 92780

October 9, 2015

Re: K151833
Trade/Device Name: Aquilion Lightning (TSX-035A/2, V7.0)
Regulation Number: 21 CFR 892.1750
Regulation Name: Computed tomography x-ray system
Regulatory Class: II
Product Code: JAK
Dated: September 21, 2015
Received: September 22, 2015

Dear Mr. Tadeo:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink that reads "Robert Ochs". The signature is written in a cursive style with a slight shadow effect behind the letters.

Robert Ochs, Ph.D.
Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K151833

Device Name

Aquilion Lightning, TSX-035A/2, V7.0

Indications for Use (Describe)

This device is indicated to acquire and display cross-sectional volumes of the whole body, to include the head.

The Aquilion Lightning has the capability to provide volume sets. These volume sets can be used to perform specialized studies, using indicated software/hardware, by a trained and qualified physician.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) SUMMARY

1. SUBMITTER'S NAME:

Toshiba Medical Systems Corporation
1385 Shimoishigami
Otawara-Shi, Tochigi-ken, Japan 324-8550

2. OFFICIAL CORRESPONDENT:

Akinori Hatanaka
Senior Manager, Regulatory Affairs and Vigilance

3. ESTABLISHMENT REGISTRATION:

9614698

4. CONTACT PERSON:

Orlando Tadeo, Jr.
Manager, Regulatory Affairs
Toshiba America Medical Systems, Inc
2441 Michelle Drive
Tustin, CA 92780
(714) 669-7459

5. Date Prepared:

July 2, 2015 (Updated September 17, 2015)

6. TRADE NAME(S):

Aquilion Lightning, TSX-035A/2, V7.0

7. COMMON NAME:

System, X-ray, Computed Tomography

8. DEVICE CLASSIFICATION:

Class II (per 21 CFR 892.1750, Computed Tomography X-ray System)

9. PRODUCT CODE / DESCRIPTION:

90JAK / Computed Tomography X-Ray System

10. PERFORMANCE STANDARD:

This device conforms to applicable Performance Standards for Ionizing Radiation Emitting Products [21 CFR, Subchapter J, Part 1020]

11. PREDICATE DEVICE:

Product	Marketed by	Regulation Number	Regulation Name	Product Code	510(k) Number	Clearance Date
Aquilion RXL, TSX-101A	Toshiba America Medical Systems	892.1750	Computed Tomography X-ray System	90JAK	K121553	July 26, 2012

12. REASON FOR SUBMISSION:

New device.

13. DEVICE DESCRIPTION:

The **Aquilion Lightning, TSX-035A/2, v7.0** is a 16-row CT System that is intended to acquire and display cross-sectional volumes of the whole body, including the head. This system is based upon the technology and materials of previously marketed Toshiba CT systems.

14. INDICATIONS FOR USE:

This device is indicated to acquire and display cross-sectional volumes of the whole body, to include the head.

The Aquilion Lightning has the capability to provide volume sets. These volume sets can be used to perform specialized studies, using indicated software/hardware, by a trained and qualified physician.

15. SUBSTANTIAL EQUIVALENCE:

The Aquilion Lightning, TSX-035A/2, is substantially equivalent to the Aquilion RXL, TSX-101A, that received premarket clearance under K121553 and is marketed by Toshiba America Medical Systems. Included in this submission is a comparison table that lists the similarities and differences between the Aquilion Lightning, TSX-035A/2, and the predicate device. See below for a brief comparison of the technological characteristics between the subject and the predicate devices:

Item	Aquilion Lightning, TSX-035A/2	Aquilion RXL, TSX-101A
510(k) Number	This submission	K121553
Gantry Rotation Speed	0.75 sec/rotation (Option: 0.6 sec)	0.5 sec/rotation (Option: 0.4sec)
Detector	800 channels × 16 rows	896 channels × 16 rows
FOV (field of view)	180/240/320/390/500 mm	180/240/320/400/500 mm
Rated output	Max. 36 kW	Max. 60 kW (72kW*)
X-ray tube voltage	80/100/120/135 kV	80/100/120/135 kV
X-ray tube current	10-300 mA	10-500 mA
X-ray fan angle	43.9°	49.2°
Noise reduction processing	QDS, AIDR 3D , AIDR3D Enhanced	QDS, AIDR 3D
Metal Artifact Reduction	SEMAR (Helical Scan)	Not Available

Gantry opening size (diameter)	780 mm	720 mm
Tilt angle	±30°	±30°
Installation area (Short patient couch)	Minimum 17.1m ² (14.3 m ²)	Minimum 27m ² (25 m ²)

16. SAFETY:

The device is designed and manufactured under the Quality System Regulations as outlined in 21 CFR § 820 and ISO 13485 Standards. This device is in conformance with the applicable parts of the following standards IEC60601-1, IEC60601-1-1, IEC60601-1-2, IEC60601-1-3, IEC60601-1-4, IEC60601-1-6, IEC60601-2-28, IEC60601-2-32, IEC60601-2-44, IEC60825-1, IEC62304, IEC62366, NEMA PS 3.1-3.18, NEMA XR-25 and NEMA XR-26. Additionally, this device complies with all applicable requirements of the radiation safety performance standards, as outlined in 21 CFR §1010 and §1020.

17. TESTING

Risk analysis and verification/validation testing conducted through bench testing are included in this submission which demonstrates that the established specifications for the device have been met. CT image quality metrics were performed, utilizing phantoms, which validated that the subject device is substantially equivalent to the predicate device with regard to spatial resolution, CT number and contrast-to-noise ratio and noise properties.

Representative diagnostic images, reviewed by an American Board Certified Radiologist, including head, chest, abdomen, pelvis and peripheral exams were also obtained using the subject device which demonstrates that the device produces images of diagnostic quality and; therefore, performs as intended.

Software Documentation for a Moderate Level of Concern, per the FDA guidance document, "Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices Document" issued on May 11, 2005, is also included as part of this submission.

Additionally, testing of the subject device was conducted in accordance with the applicable standards published by the International Electrotechnical Commission (IEC) for Medical Devices and CT Systems.

18. CONCLUSION

The Aquilion Lightning, TSX-035A/2, performs in a manner similar to and is intended for the same use as the predicate device, as indicated in product labeling. Based upon bench testing, representative clinical images, successful completion of software validation and application of risk management and design controls, it is concluded that the subject device is substantially equivalent in safety and effectiveness to the predicate device.