



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Senecka Spine
% Rich Jansen, Pharm.D.
Silver Pine Consulting, LLC
11821 Bramble Cove Drive
Fort Myers, Florida 33905

September 30, 2015

Re: K151849

Trade/Device Name: Seneka I Polyscrew Pedicle Fixation System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: NKB, MNI, MNH
Dated: July 6, 2015
Received: July 7, 2015

Dear Dr. Jansen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K151849

Device Name

Seneka I Polyscrew Pedicle Fixation System

Indications for Use (Describe)

The Seneka I Polyscrew Pedicle Fixation System is intended for posterior, noncervical pedicle fixation as an adjunct to fusion in skeletally mature patients using autograft and/or allograft for the following indications: degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and or lordosis); tumor; pseudoarthrosis; and failed previous fusion.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Date Prepared: September 23, 2015

Contact: Khalid Sethi
Senecka Spine
46 Harrison Street
Johnson City, NY 13790
607-237-4724
Fax: 607-584-0387

Regulatory Contact: Rich Jansen, Pharm. D.
Silver Pine Consulting, LLC
richj@s-pineconsulting.com

Trade Names: Seneka I Polyscrew Pedicle Fixation System
Product Class: Class III
Classification: 21 CFR §888.3070 Pedicle Screw Spinal System
Common Name: Pedicle Screw System
Product Codes: NKB, MNI, MNH
Panel Code: 87

Indications for Use:

The Seneka I Polyscrew Pedicle Fixation System is intended for posterior, noncervical pedicle fixation as an adjunct to fusion in skeletally mature patients using autograft and/or allograft for the following indications: degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion.

Device Descriptions:

The Seneka I Polyscrew Pedicle Fixation System is comprised of: straight and pre-curved rods, pedicle screw assemblies with both cannulated and non-cannulated screws, reduction screws, domino connectors, offset connectors and a set screw. Various forms and sizes of these implants are available so that adaptations can be utilized to take into account the unique pathology of individual patients. The Seneka I system can be implanted via the open surgical approach.

Components are made of Ti6Al4V ELI, a titanium based alloy, which complies with ASTM F136, cobalt chrome per ASTM F1537 or CP Titanium per ASTM E2371-13.

Predicate Device(s):

The Seneka I Polyscrew Pedicle Fixation System is substantially the primary predicate device which is the Moss Miami System (DePuy Spine) (K022623). Additional predicate devices include Optima Spinal System (U&I) (K031585) and Xia Spinal System (Stryker) (K001319).

Performance Standards:

The pre-clinical testing performed includes static and dynamic compression bending, and static torsion per ASTM F1717-10.

Technological Characteristics:

Senecka Spine has compared the Seneka I Polyscrew Pedicle Fixation System to the predicate devices in regards to indications for use, materials, function, sizes and mechanical test results. These comparisons demonstrate substantial equivalence to the predicate devices.

Conclusion:

Senecka Spine concludes that the Seneka I Polyscrew Pedicle Fixation System is substantially equivalent to the predicate devices and raises no new questions of safety or effectiveness.