



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

December 21, 2015

OsteoMed
Ms. Piedad Peña
Manager, Regulatory Affairs
3885 Arapaho Road
Addison, Texas 75001

Re: K152145

Trade/Device Name: OsteoMed ExtremiLOCK Wrist Plating System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/multiple component metallic bone fixation appliances and accessories
Regulatory Class: Class II
Product Code: HRS, KWC
Dated: November 24, 2015
Received: November 25, 2015

Dear Ms. Peña:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K152145

Device Name: OsteoMed ExtremiLOCK Wrist Plating System

Indications for Use:

OsteoMed ExtremiLOCK Wrist Plating System is indicated for fracture fixation, fusion and osteotomies of wrist and other bones appropriate for the size of the device. It is intended for use in trauma, general surgery, and reconstructive procedures.

OsteoMed ExtremiLOCK Wrist Plating System implants are intended for single use only.

Prescription Use X

(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use

(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



510(k) Summary

Submitter Information OsteoMed
 3885 Arapaho Road
 Addison, Texas 75001
 Phone: (972) 677-4600
 Fax: (972) 677-4601

Contact Person Mrs. Piedad Peña

Date Prepared July 31, 2015

Proprietary Name: ***OSTEOMED EXTREMILOCK WRIST PLATING SYSTEM***
Common Name: Wrist Fixation System
Classification Name: 21 CFR 888.3030, Single/multiple component metallic bone fixation appliances and accessories
Product Code: HRS, HWC

Primary Predicate:

OsteoMed Wrist Plating System, K120015

Classification Name: 21 CFR 888.3030, Single/multiple component metallic bone fixation appliances and accessories; Product Code: HRS, HWC; Class: II

Secondary Predicate:

Synthes 2.4mm VA-LCP Two Column Volar Distal Radius Plate, K083694

Classification Name: 21CFR 888.3030 Single/multiple component metallic bone fixation appliances and accessories; Product Code: Product Code HRS, Class: II

3rd Reference Predicate:

OsteoMed Hand Plate and Screw Fixation System, K090522

Classification Name: 21 CFR 888.3030, Single/multiple component metallic bone fixation appliances and accessories; Product Code: HRS, HWC; Class: II

4th Reference Predicate:

OsteoMed Foot Plate and Screw Rigid Fixation System, K091614

Classification Name: 21 CFR 888.3030, Single/multiple component metallic bone fixation appliances and accessories; Product Code: HRS, HWC; Class: II

5th Reference Predicate:

OsteoMed Headless Cannulated Screw, K063298

Classification Name: 21 CFR 888.40, Smooth or threaded metallic bone fixation fastener; Product Code: HWC; Class: II

Device Description	The OsteoMed ExtremiLOCK Wrist Plating System is a rigid fixation and fusion system consisting of plates and screws in various configurations along with the appropriate instrumentation to facilitate modification and implantation. Plates are anatomically pre-contoured in various shapes and sizes. Screws are provided with variable angle locking and non-locking heads and are either fully threaded or partially threaded in various lengths. These screws are either solid core or cannulated and can be used with or without plates. The system also contains smooth variable angle locking pegs and K-wire implants. The implants of the OsteoMed ExtremiLOCK Wrist Plating System are made from Titanium per ASTM F-67 (Plates), Titanium Alloy per ASTM F-136 (plates, screws, pegs and washer), and Stainless Steel per ASTM F-138(K-wires). The modifications to the plates include the new material of CP Titanium per ASTM F-67 and the screw modifications include the new dual lead technology. The rest of the system has already been cleared through OsteoMed wrist predicate 510(k) K120015. The modification also introduced new sterile packaging configurations for the implants and disposable instruments. The system instruments included facilitate modification and insertion of the implants.
Indications for use/ Intended Use	The OsteoMed ExtremiLOCK Wrist Plating System is indicated for fracture fixation, fusion and osteotomies of wrist and other bones appropriate for the size of the device. It is intended for use in trauma, general surgery and reconstructive procedures. OsteoMed ExtremiLOCK Wrist Plating System implants are intended for single use only.
Performance Characteristics & Testing & Clinical	Performance: Verification testing was conducted to ensure plates and screws performed equal or better compared to the predicate devices. The plates were tested against the Synthes Secondary Predicate Device to ensure the design features and new material met the required mechanical strength criteria for their intended use. The screws with the dual lead technology underwent verification evaluation to ensure the new design features met the mechanical strength criteria for the intended use. The screws were compared to their respective predicates. Performance equivalence was shown through the verification comparison to the predicate devices.

Clinical Testing is not required to support substantial equivalence.

Substantial Equivalence The basis of substantial equivalence for this device, **OsteoMed ExtremiLOCK Wrist Plating System**, is based on similarities in indications for use, intended use, material, function, performance, design, technology, sterilization, and operational principles to the OsteoMed predicates and Synthes predicate. Performance comparisons were performed which verified that the new plates and screws met required mechanical strength criteria for their intended use compared to the legally marketed predicate devices listed in this summary. OsteoMed has shown that the non-clinical tests demonstrate that the devices are as safe and effective as the predicate devices.