



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

September 29, 2015

Grandway Technology (shenzhen) Limited
Mr. Patrick Chow
General Manager
Block 7, Zhu Keng Industrial Zone, Ping Shan District,
Shenzhen, 518118 CN

Re: K152483
Trade/Device Name: Digital Automatic Blood Pressure Monitor BPM30 and BPM31
Series
Regulation Number: 21 CFR 870.1130
Regulation Name: Non-Invasive Blood Pressure Measurement System
Regulatory Class: Class II
Product Code: DXN
Dated: August 31, 2015
Received: August 31, 2015

Dear Mr. Patrick Chow,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply

with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Bram D. Zuckerman", is written over a faint, large, light-gray background watermark of the FDA logo.

for Bram D. Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K152483

Device Name

Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series

Indications for Use (Describe)

This device is for use by medical professional or home users. It is intended to measure the systolic and diastolic blood pressure on an adult individual by using a non-invasive technique, in which an inflatable cuff is wrapped around the upper arm.

Type of Use (Select one or both, as applicable)

☐ Prescription Use (Part 21 CFR 801 Subpart D)

☒ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K152483

Page 1 of 4

1. Submitter identification

510(k) Submitter	GRANDWAY TECHNOLOGY (SHENZHEN) LIMITED
Address	Block 7, Zhu Keng Industrial Zone, Ping Shan, Long Gang District, 518118, Shenzhen, People's Republic of China
Phone Number	(00852)-2851-6789
Fax Number	(00852)-2851-6278
Contact Person	Mr. Patrick Chow
Date of Submission	28-Aug-2015

2. Device identification

Trade Name	Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series
Common Name	Non-invasive Blood Pressure Measurement System
Classification Name	Non-invasive Blood Pressure Measurement System (CFR 870.1130, Class II, Product Code DXN)

3. Predicate device

Predicate Device	Digital Automatic Blood Pressure Monitor BPM18 Series
Manufacturer	GRANDWAY TECHNOLOGY (SHENZHEN) LIMITED
510(k) Number	K133619

4. Device Description

Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series is a non-invasive blood pressure measurement system for use by medical professional or at home. It is designed to measure the systolic and diastolic blood pressure, and pulse rate of an individual in each measurement and then display the readings on a digital panel.

The device utilizes the oscillometric methodology, in which an inflatable cuff is wrapped around the upper arm of an individual, for blood pressure measurement. This means the monitor detects your blood's movement through your brachial artery and converts the movement into a digital reading.

The table below illustrate the feature presence in Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series.

Model	BP Measurement	Pulse Rate Measurement	WHO Classification	IHB	LCD Type	Backlight	DC Jack	V.A Size
MD3000	✓	✓	✓	✓	Positive reflective	✗	✗	60 × 70 mm
MD3010	✓	✓	✓	✓	Positive reflective	✗	✓	60 × 70 mm

Model	BP Measurement	Pulse Rate Measurement	WHO Classification	IHB	LCD Type	Backlight	DC Jack	V.A Size
MD3020	✓	✓	✓	✓	Negative transmissive	✓	X	60 × 70 mm
MD3030	✓	✓	✓	✓	Negative transmissive	✓	✓	60 × 70 mm
MD3040	✓	✓	✓	✓	Positive transmissive	✓	X	60 × 70 mm
MD3050	✓	✓	✓	✓	Positive transmissive	✓	✓	60 × 70 mm
MD3001	✓	✓	✓	✓	Positive reflective	X	X	50 × 60 mm
MD3011	✓	✓	✓	✓	Positive reflective	X	✓	50 × 60 mm
MD3021	✓	✓	✓	✓	Negative transmissive	✓	X	50 × 60 mm
MD3031	✓	✓	✓	✓	Negative transmissive	✓	✓	50 × 60 mm
MD3041	✓	✓	✓	✓	Positive transmissive	✓	X	50 × 60 mm
MD3051	✓	✓	✓	✓	Positive transmissive	✓	✓	50 × 60 mm
MD3100	✓	✓	✓	✓	Positive reflective	X	X	60 × 70 mm
MD3110	✓	✓	✓	✓	Positive reflective	X	✓	60 × 70 mm
MD3120	✓	✓	✓	✓	Negative transmissive	✓	X	60 × 70 mm
MD3130	✓	✓	✓	✓	Negative transmissive	✓	✓	60 × 70 mm
MD3140	✓	✓	✓	✓	Positive transmissive	✓	X	60 × 70 mm
MD3150	✓	✓	✓	✓	Positive transmissive	✓	✓	60 × 70 mm
MD3101	✓	✓	✓	✓	Positive reflective	X	X	50 × 60 mm
MD3111	✓	✓	✓	✓	Positive reflective	X	✓	50 × 60 mm
MD3121	✓	✓	✓	✓	Negative transmissive	✓	X	50 × 60 mm
MD3131	✓	✓	✓	✓	Negative transmissive	✓	✓	50 × 60 mm
MD3141	✓	✓	✓	✓	Positive transmissive	✓	X	50 × 60 mm
MD3151	✓	✓	✓	✓	Positive transmissive	✓	✓	50 × 60 mm

5. Indication for use

This device is for use by medical professional or home users. It is intended to measure the systolic and diastolic blood pressure on an adult individual by using a non-invasive technique, in which an inflatable cuff is wrapped around the upper arm.

6. Comparison of technological characteristics between new Device and predicate Devices

Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series is compared to the predicate device, [Digital Automatic Blood Pressure Monitor BPM18 Series \(K133619\)](#) in the device comparison table below.

Item	Predicate Device	New device	Comment
Indication for Use	Digital Automatic Blood Pressure Monitor BPM18 Series is for use by medical professional or home user. The BPM18 Series is intended to measure the systolic and diastolic blood pressure, and pulse rate of an adult individual by using a non-invasive technique, in which an inflatable cuff is wrapped around the upper arm.	This device is for use by medical professional or home users. It is intended to measure the systolic and diastolic blood pressure on an adult individual by using a non-invasive technique, in which an inflatable cuff is wrapped around the upper arm.	Equivalent
Measurement Method	Non-invasive, Oscillometric	Non-invasive, Oscillometric	Identical
IHB Detection	Yes	Yes	Identical

Item	Predicate Device	New device	Comment
Patient Population	Adult	Adult	Identical
BP Measurement Range	Cuff Pressure: 0 - 300 mmHg Systolic Pressure: 50 - 250 mmHg Diastolic Pressure: 30 - 200 mmHg	Cuff Pressure: 0 - 300 mmHg Systolic Pressure: 50 - 250 mmHg Diastolic Pressure: 30 - 200 mmHg	Identical
Number of User	2 independent users	4 independent users	More user
Memory Space	2 users × 120 memory space	4 users × 120 or 240 memory space	More user memory
Resolution of Measurement	Blood Pressure: 1 mmHg or 0.1kPa Pulse Rate: 1 beat/ min	Blood Pressure: 1 mmHg or 0.1kPa Pulse Rate: 1 beat/ min	Identical
Blood Pressure Measurement Accuracy	± 3 mmHg or 2% of reading	± 3 mmHg	Equivalent
Pulse Rate Measurement Range	40 - 180 beats/min	40 - 180 beats/min	Identical
Pulse Rate Measurement Accuracy	± 5 % of the reading	± 5 % of the reading	Identical
Display Type	LCD	LCD	Identical
Power Source	4 × 1.5 V AAA-batteries and/or AC Adaptor	4 × 1.5 V AAA-batteries and/or AC Adaptor	Identical
Pressurization Mode	Automatic Inflation	Automatic Inflation	Identical
Deflation Mode	Automatic Exhaust/ Deflation	Automatic Exhaust/ Deflation	Identical
Operating Condition	Temperature: +5 to +40 °C Humidity: 15 to 93 % R.H. max Atmospheric Pressure: 700-1060 hPa	Temperature: +5 to +40 °C Humidity: 15 to 93 % R.H. max Atmospheric Pressure: 700-1060 hPa	Identical
Storage and Transportation Condition	Temperature: -25 to +70 °C Humidity: up to 93% R.H. max Atmospheric Pressure: 700-1060 hPa	Temperature: -25 to +70 °C Humidity: up to 93% R.H. max Atmospheric Pressure: 700-1060 hPa	Identical
Material	Resistances, capacitance, transistors, amplifiers, pressure sensor, CPU, PCB, cuff ABS button, ABS cabinet, batteries and packaging	Resistances, capacitance, transistors, amplifiers, pressure sensor, CPU, PCB, cuff ABS button, ABS cabinet, batteries and packaging	Identical
Compatibility with Environment and Other Devices	No influence with environment and other device	No influence with environment and other device	Identical
Applicable Standard	- EN 1060-1:1995+A2:2009 - EN 1060-3:1997+A2:2009 - IEC 60601-1:2012 - EN 60601-1-2:2007 - FCC Part 15 Subpart B - IEC 81060-2:2009	- EN 1060-1:1995+A2:2009 - EN 1060-3:1997+A2:2009 - IEC 60601-1:2012 - IEC 60601-1-2:2007 - FCC Part 15 Subpart B - IEC 81060-2:2013	Equivalent

Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series is a non-invasive measuring device and utilizes the oscillometric methodology to measure the blood pressure reading. The key components of device are: a pressure sensor, an electric valve and an electronic control module together with an electric pump.

K152483

Page 4 of 4

The electric pump inflate (and deflate) the inflatable cuff automatically according to our designed architecture. The predicate device adopts exactly same methodology and key components for measuring blood pressure.

7. Clinical and Non-clinical Tests

Clinical Test Summary

Testing to insure clinical accuracy of the device in accordance with ISO 81060-2 as documented in Clinical Test report.

One hundred patients (49 males and 51 females) were invited for the study. Standard auscultation method was used as the reference blood pressure monitor measuring in the left arm. Blood pressure measurements were repeated alternatively with the device and auscultation in the same arm according to the sequence in ISO 81060-2.

Non-Clinical Test Summary

Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series has performed several non-clinical tests to show that all requirement specifications and standard requirements are met. The tests includes the follows:

- ✧ EN 1060-1:1995+A2:2009
- ✧ EN 1060-3:1997+A2:2009
- ✧ IEC 60601-1:2012
- ✧ IEC 60601-1-2:2007
- ✧ FCC Part 15 Subpart B

All of the clinical and non-clinical testing performed on Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series are same as the predicate device. Therefore, no bench test is conducted to show the performance of Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series is equivalent to the predicate device.

8. Conclusion

Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series has the same intended use and same technological characteristics as the predicate device, [Digital Automatic Blood Pressure Monitor BPM18 Series \(K133619\)](#). Moreover both clinical and non-clinical testing has demonstrated that no differences in the technological characteristics and questioning on safety or effectiveness to be raised. Thus, Digital Automatic Blood Pressure Monitor BPM30 and BPM31 Series is substantially equivalent to the predicate device.