



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

March 17, 2016

Minghao Medical Products Co., Ltd.  
% Chu Xiaoan  
Official Correspondent  
Room 1606 Bldg. 1 Jianxiang Yuan  
No 209 Bei Si Huan Zhong Rd Haidian District  
Beijing, 100083 CHINA

Re: K152707

Trade/Device Name: Minghao Nitrile Powder Free Patient Examination Gloves, Blue Color  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Patient Examination  
Regulatory Class: I  
Product Code: LZA  
Dated: January 28, 2016  
Received: February 1, 2016

Dear Chu Xiaoan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

*Tejashri Purohit-Sheth, M.D.*

Tejashri Purohit-Sheth, M.D.  
Clinical Deputy Director  
DAGRID/ODE/CDRH FOR

Erin I. Keith, M.S.  
Division Director  
Division of Anesthesiology,  
General Hospital, Respiratory,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K152707

Device Name  
Minghao Nitrile Powder Free Patient Examination Gloves, Blue Color

Indications for Use (Describe)

Nitrile Powder Free Patient Examination Gloves, Blue Color is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## Section 6 510(k) Summary

### 510(K) Summary

"This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92."

"The assigned 510(k) number is: K152707 "

#### Premarket Notification [510(k)] Summary

##### 1.0 Submitter:

Submitter's name : Minghao Medical Products Co.,Ltd.  
Submitter's address : West Industrial Park, Luannan County,  
Tangshan City, 063500 Hebei Province,China  
Phone number : (86)315-8230716  
Fax number : (86)315-8010500  
Name of contact person: Pang Cuiru  
Date of preparation : 2016-03-01

##### 2.0 Name of the Device

Device Name: Nitrile Powder Free Patient Examination  
Gloves, Blue Color  
Proprietary/Trade name: Minghao Nitrile Powder Free Patient  
Examination Gloves, Blue Color  
Common Name: Exam gloves  
Classification Name: Patient examination glove  
Device Classification: I  
Regulation Number: 21 CFR 880.6250  
Panel: General Hospital (80)  
Product Code: LZA

##### 3.0 Predicate device

Device Name: Nitrile Powder Free Patient Examination Gloves,  
Blue Color  
Company name: Tangshan Zhonghong Pulin Plastic Co.,Ltd.  
510(K) Number: K120970

##### 4.0 Device Description:

###### 4.1 How the device functions:

Nitrile films form a barrier to body fluids and bloodborne Pathogens

###### 4.2 Scientific concepts that form the basis for the device

The nitrile rubber is water tight under normal conditions of use. It's tensile properties cause it to conform to the hand, allowing movements necessary for a

medical procedure.

**4.3 Physical and performance characteristics such as design, materials and physical properties:**

Nitrile glove is known to create a barrier to bloodborne pathogens and body fluids. ASTM conforming tensile properties create a glove that is strong and flexible. The glove is manufactured in accordance with the requirements of ASTM D6319 and ASTM D5151 requirements.

**5.0 Device Intended Use (Indication for use):**

Nitrile Powder Free Patient Examination Gloves, Blue Color is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

**6.0 Summary of the Technological Characteristics of the Device:**

The Nitrile Powder Free Patient Examination Gloves, Blue Color, non sterile are summarized with the following technological characteristics compared to ASTM or equivalent standard.

<b>Characteristics</b>	<b>Standard</b>	<b>Device performance</b>
Dimension	ASTM standard D 6319-10.	Meets
Physical Properties	ASTM standard D 6319-10.	Meets
Freedom from pinholes	21 CFR 800.20	Meets
Powder Residual	ASTM standard D 6319-10 and D6124-06(Reapproved 2011).	Meets <2mg/glove
Biocompatibility	Primary Skin Irritation in rabbits ISO 10993-10: Third Edition 2010-08-01.	Under the conditions of the study, the subject device is non-irritating
	Dermal sensitization in the guinea pig ISO 10993-10: Third Edition 2010-08-01.	Under the conditions of the study, the subject device is not a sensitizer

**7.0 Substantial Equivalent Based on Assessment of Non-Clinical Performance Data:**

Nitrile Powder Free Patient Examination Gloves, Blue Color, meet requirements per ASTM D6319-10, per ASTM D6124-06(Reapproved 2011), per 21 CFR 800.20 and ISO 10993-10: Third Edition 2010-08-01.

The performance test data of the non clinical tests that support a determination of substantial equivalent is the same as mentioned immediately above.

**8.0 Substantial Equivalent Based on Assessment of Clinical Performance Data:**

Clinical data was not needed for this submission.

**9.0 Substantial Equivalence Comparison:**

Features & Description	Predicate Device	Subject Device	Result of Comparison
Company	Tangshan Zhonghong Pulin Plastic Co.,Ltd.	Minghao Medical Products Co.,Ltd.	--
510(K) Number	K120970	K152707	
Device name	Nitrile Powder Free Patient Examination Gloves, Blue Color	Nitrile Powder Free Patient Examination Gloves, Blue Color	same
Product Code	LZA	LZA	same
Size	Small/ Medium/ Large/X large	Small/ Medium/ Large/X large	same
Intend for use	Nitrile Powder Free Patient Examination Gloves, Blue Color is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	Nitrile Powder Free Patient Examination Gloves, Blue Color is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	Substantially equivalent
Device Description and Specifications	Meets ASTM D 6319-10	Meets ASTM D 6319-10	Substantially equivalent
Dimensions --Length	Meets ASTM D6319-10 ≥230mm min	Meets ASTM D6319-10 ≥230mm min for all sizes	Substantially equivalent
Dimensions -- Width	Meets ASTM D6319-10  Small 70-90 mm Medium 85-105mm Large 100-120mm Xlarge 110-130 mm	Meets ASTM D6319-10  Small 70-90 mm Medium 85-105 mm Large 100-120 mm X large 110-130 mm	Substantially equivalent
Dimensions --Thickness	Meets ASTM D6319-10  Finger 0.05mm min. Palm 0.05mm min.	Meets ASTM D6319-10  Thickness (mm) min. Finger 0.05 mm min. Palm 0.05 mm min.	Substantially equivalent
Physical Properties	Meets ASTM D 6319-10  Before aging/after aging  Elongation ≥500% Tensile Strength ≥ 14MPa	Meets ASTM D6319-10  Before aging/after aging  Elongation ≥500% Tensile Strength ≥ 14MPa	Substantially equivalent
Freedom from Pinholes	Meets • 21 CFR 800.20 • ASTM D6319-10 • ASTM D 5151-06 (Reapproved 2011)	Meets • 21 CFR 800.20 • ASTM D6319-10 • ASTM D5151-06 (Reapproved 2011)  Holes at	Substantially equivalent

		Inspection Level I AQL2.5	
Residual Powder	Meets ASTM D 6124-06 (Reapproved 2011)  below 2mg of residual powder	Meets ASTM D 6124-06 (Reapproved 2011)  Results generated values below 2mg of residual powder	Substantially equivalent
Materials used to fabricate the devices	Nitrile	Nitrile	Substantially equivalent
Dusting or Donning Powder: name	PU	Polyurethane	Substantially equivalent
Compare performance data supporting substantial equivalence	Meets ASTM D5151-06 (Reapproved 2011) ASTM D6319-10 ASTM D6124-06 (Reapproved 2011)	Meets ASTM D5151-06 (Reapproved 2011) ASTM D6319-10 ASTM D6124-06 (Reapproved 2011)	Substantially equivalent
Single Patient Use	Single Patient Use	Single Patient Use	Substantially equivalent
Biocompatibility	SKIN IRRITATION DERMAL and SENSITIZATION STUDIES Meets ISO 10993-10:2002/Amd.1:2006	SKIN IRRITATION DERMAL and SENSITIZATION STUDIES Meets ISO 10993-10: Third Edition 2010-08-01.  Under the conditions of the study, not an irritant and under conditions of the study, not a sensitizer.	Substantially equivalent
Labeling for the legally marketed device to which substantial equivalence is claimed.	-Powder Free -Patient Examination Glove -Single Use Only - Manufactured For: - Lot -Blue color - Non sterile	-Powder Free -Patient Examination Glove -Single Use Only - Manufactured For: - Lot -Blue color - Non sterile	Substantially equivalent

### 10.0 Substantial Equivalence Comparison:

It can be concluded that the Nitrile Powder Free Patient Examination Gloves, Blue Color meet the ASTM standard or equivalent standard and FDA requirements for waterleak test on pinhole AQL., meet labeling claims.

It can be concluded that the Nitrile Powder Free Patient Examination Gloves, Blue Color is as safe, as effective, and performs as well as the predicate device, Nitrile Powder Free Patient Examination Gloves, Blue Color, Tangshan Zhonghong Pulin Plastic Co., Ltd.. K120970.