



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

February 8, 2016

Biomet Manufacturing Corp.
Mr. Adam Cargill
Regulatory Affairs Specialist
56 East Bell Drive
PO Box 587
Warsaw, Indiana 46581

Re: K152868

Trade/Device Name: ExpressBraid Graft Manipulation
Regulation Number: 21 CFR 878.5000
Regulation Name: Nonabsorbable Poly(Ethylene Terephthalate) Surgical Suture
Regulatory Class: Class II
Product Code: GAT
Dated: January 7, 2016
Received: January 11, 2016

Dear Mr. Cargill:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

David Krause -S

for Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K152868

Device Name

ExpressBraid Graft Manipulation

Indications for Use (Describe)

The ExpressBraid Graft Manipulation is intended for use in soft tissue approximation and/or ligation. The suture may be provided individually or be incorporated as a component, into surgeries where constructs including those with allograft or autograft tissue are used for repair.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

In accordance with 21 CFR §807.92 and the Safe Medical Devices Act of 1990, the following information is provided for the ExpressBraid 510(k) premarket notification. The submission was prepared in accordance with the FDA guidance document, 'Format for Traditional and Abbreviated 510(k)s', issued on August 12, 2005.

Sponsor: Biomet Manufacturing Corp.
56 East Bell Drive
PO Box 587
Warsaw, IN 46581
Office: 574-267-6639
Main Fax: 574-267-8137
Establishment Registration Number: 1825034

Contact: Adam Cargill
Regulatory Affairs Specialist

Date: January 7, 2016

Subject Device: Trade Name: ExpressBraid Graft Manipulation

Common Name: Non-absorbable Surgical Suture

Classification Name:
• GAT– Nonabsorbable poly(ethylene terephthalate) Surgical Suture
(21 CFR 878.5000)

Legally marketed devices to which substantial equivalence is claimed:

Predicate Device:
• Arthrex Suture - K122374

Reference Devices:
• Teleflex Force Fiber - K033654/K040472

Device Description

The ExpressBraid Graft Manipulation device consists of a stainless steel, straight needle attached to a strand of MaxBraid suture. The two ends of the strand of suture are crimped into the needle. The suture may be all white or blue/white MaxBraid suture. The MaxBraid sutures are braided, non-absorbable sutures available in a white configuration consisting of 100% Ultra High Molecular Weight Polyethylene (UHMWPE) fibers and a blue and white co-braid configuration consisting of Deklene II, Polypropylene suture braided into 100% UHMWPE fibers. The suture is provided uncoated. The colorant used in the blue of the blue and white co-braid consists of [phthalocyaninato (-2)] cooper <0.5% by weight per 21 CFR Sec. 74.3045.

Intended Use and Indications for Use

The ExpressBraid Graft Manipulation is intended for use in soft tissue approximation and/or ligation. The suture may be provided individually or be incorporated as a component, into surgeries where constructs including those with allograft or autograft tissue are used for repair.

Summary of Technological Characteristics

The ExpressBraid Graft Manipulation device has similar technological characteristics (design features, sizes, etc.) as the predicate.

- ExpressBraid Graft Manipulation has the same indications for use as the predicate Arthrex Suture (K122374).
- Both ExpressBraid Graft Manipulation and Arthrex Suture are manufactured from UHMWPE.
- Both ExpressBraid Graft Manipulation and Arthrex Suture meet USP Reference Standards.
- ExpressBraid Graft Manipulation is provided with needles attached to the suture. Arthrex Suture is available with or without needles attached to the suture.
- ExpressBraid Graft Manipulation colorant used in the blue of the blue and white co-braid configuration consists of [phthalocyaninato (-2)] cooper <0.5% by weight per 21 CFR Sec. 74.3045. The dyes used in the Arthrex Suture may include D&C Blue No. 6, D&C Green No. 6, and Logwood Black.
- MaxBraid Suture is identical to the Force Fiber Suture cleared in Teleflex 510(k)s K033654 and K040472.

Summary of Performance Data

ExpressBraid Graft Manipulation was testing in accordance with USP – non-absorbable surgical sutures for tensile strength, diameter, and needle attachment and meets the requirements of the Class II Special Controls Guidance: Surgical Sutures; Guidance for Industry and FDA; June 3, 2003.

Substantial equivalence testing was performed comparing ExpressBraid Graft Manipulation to the Arthrex #2 Suture. Testing concluded that the ExpressBraid Graft Manipulation will perform equivalently to the Arthrex #2 Suture.

All materials used in ExpressBraid Graft Manipulation were evaluated per ISO 10993-1 and meet the standard requirements for biocompatibility.

Substantial Equivalence Conclusion

The proposed ExpressBraid Graft Manipulation has similar intended use, design characteristics, and mechanical performance as the predicate Arthrex Suture. The performance data identified no new risks and substantial equivalence to the legally marketed predicate device.