



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

ConforMIS, Incorporated
Ms. Amita Shah
Senior Vice President, Regulatory and Quality Affairs
28 Crosby Drive
Bedford, Massachusetts 01730

March 25, 2016

Re: K153721

Trade/Device Name: iTotal Cruciate Retaining (CR) Knee Replacement System
Regulation Number: 21 CFR 888.3560
Regulation Name: Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis
Regulatory Class: Class II
Product Code: JWH, OIY, OOG
Dated: December 23, 2015
Received: December 28, 2015

Dear Ms. Shah:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K153721

Device Name

ConforMIS iTotal Cruciate Retaining Knee Replacement System

Indications for Use (Describe)

The iTotal CR Knee Replacement System (KRS) is intended for use as a total knee replacement in patients with knee joint pain and disability whose conditions cannot be solely addressed by the use of a prosthetic device that treats only one or two of the three knee compartments, such as a unicondylar, patellofemoral or bicompartamental prosthesis:

The Indications for Use include:

- Painful joint disease due to osteoarthritis, traumatic arthritis, rheumatoid arthritis or osteonecrosis of the knee
- Post traumatic loss of joint function.
- Moderate varus, valgus or flexion deformity in which the ligamentous structures can be returned to adequate function and stability.
- Failed osteotomies, hemiarthroplasties, and unicondylar, patellofemoral or bicompartamental implants.
- Revision procedures provided that anatomic landmarks necessary for alignment and positioning of the implant are identifiable on patient imaging scans.

This implant is intended for cemented use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

6.0 510(K) SUMMARY (PAGE 1 OF 5)

Submitter's Name and Address: ConforMIS, Inc.
28 Crosby Drive
Bedford, MA 01730

Establishment Registration Number: 3009844603 and 3004153240

Date of Summary: December 21, 2015

Contact Person: Amita S. Shah, Sr. Vice President, Regulatory and Quality Affairs
Telephone Number: (781) 345-9164
Fax Number: (781) 345-0147

Name of the Device: ConforMIS iTotal® CR Knee Replacement System (iTotal CR KRS)

Common Name: Total Cruciate Retaining Knee Replacement System

Regulatory Status and Regulation Number: Class II
21 CFR 888.3560

Classification Name: Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis

Device Classification: Product Codes:
JWH: Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis.

OOG: Knee Arthroplasty Implantation System.
Intended to be used to assist in the implantation of a specific knee arthroplasty device or a set of specific knee arthroplasty devices. Indicated to include guiding alignment, making or establishing cuts, selecting, sizing, attaching, positioning or orienting implant components.

OIY: Prosthesis, knee, patellofemorotibial, semi-constrained, cemented polymer + additive/metal/polymer + additive. This generic type of device includes prosthesis that have a femoral component made of alloys, such as cobalt-chromium-molybdenum, and a tibial component(s) and/or a retropatellar resurfacing component made of ultra-high molecular weight polyethylene plus an additive, such as a-tocopherol.

510(K) SUMMARY (PAGE 2 OF 5)

Indications for Use: The iTotal® Cruciate Retaining (CR) Knee Replacement System (KRS) is intended for use as a total knee replacement in patients with knee joint pain and disability whose conditions cannot be solely addressed by the use of a prosthetic device that treats only one or two of the three knee compartments, such as a unicondylar, patellofemoral or bicompartamental prosthesis.

The Indications for Use include:

- Painful joint disease due to osteoarthritis, traumatic arthritis, rheumatoid arthritis or osteonecrosis of the knee.
- Post traumatic loss of joint function.
- Moderate varus, valgus or flexion deformity in which the ligamentous structures can be returned to adequate function and stability.
- Failed osteotomies, hemiarthroplasties, and unicondylar, patellofemoral or bicompartamental implants.
- Revision procedures provided that anatomic landmarks necessary for alignment and positioning of the implant are identifiable on patient imaging scans.

This implant is intended for cemented use only.

Identification of the Legally Marketed Device (Predicate Device):

ConforMIS iTotal CR Knee Replacement System (ITOTAL CR KRS)

Device Class: II
 Product Code: JWH, OOG, OIY
 Regulation Number: 21 CFR 888.3560
 510(k) Number: K142161, K152704

510(K) SUMMARY (PAGE 3 OF 5)

Device Description: The iTotal® CR Knee Replacement System (hereafter referred to as the "iTotal CR KRS") is a patient specific tricompartmental faceted posterior cruciate ligament (PCL) retaining knee replacement system. The iTotal® CR KRS is a semi-constrained, cemented knee implant which consists of a femoral, tibial, and patellar component.

Using patient imaging and a combination of proprietary and off the shelf software a patient specific implant is designed, that best meets the geometric and anatomic requirements of the specific patient. The femoral component is manufactured from cobalt chromium molybdenum ("CoCrMo") alloy. The tibial component includes a metal tray manufactured from CoCrMo alloy and either one or two polyethylene inserts. The polyethylene inserts may be manufactured from either UHMWPE or a highly cross-linked Vitamin E infused polyethylene (iPoly XE™). The patellar component is also manufactured from either UHMWPE or from a highly cross-linked Vitamin E infused polyethylene (iPoly XE™).

For user convenience, and similar to the predicate iTotal CR KRS, patient specific accessory orthopedic manual surgical instruments designed for use with the modified iTotal CR KRS are provided to assist in the positioning of the total knee replacement components intraoperatively and in guiding the cutting of bone. In addition, reusable orthopedic manual surgical instruments are provided separately.

The function and general design features of the patient specific implants and ancillary instruments remain similar to those described in the predicate 510(k)s **K142161** and **K152704**.

Substantial Equivalence:

The subject of this premarket notification is substantially equivalent to the iTotal CR KRS (**K142161** cleared November 14, 2014 and **K152704** cleared October 21, 2015). The following testing was performed to establish substantial equivalence:

- Software verification and validation testing of proprietary software

510(K) SUMMARY (PAGE 4 OF 5)

Description and Assessment of Nonclinical Testing:

The modified iTTotal CR KRS and predicate iTTotal CR KRS have the same technological characteristics. There have been no changes to the modified iTTotal CR KRS device with respect to design, materials and methods of manufacture, packaging, or sterilization. The design of the device remains similar to that cleared in **K142161** and **K152704**.

The proposed modifications, which represent improving manufacturing efficiencies, are primarily focused on increasing automation of the CAD manufacturing processes via the proprietary software modules as shown in **Table 1** below.

Table 1: Comparison between the Modified and Predicate Device

Characteristic	Predicate iTTotal Cruciate Retaining Knee Replacement System (K142161 and K152704)	iTotal Cruciate Retaining Knee Replacement System (This submission)
CAD Process for Tibial implants	iTotalTib version 3.0 or manual process	iFit iTTotalTib CR version 4.0 or manual process
CAD Process for Tibial iJigs	iTotal TibJigs version 2.0 or manual process	
CAD Review of tibial components	Manual Process	

Nonclinical testing: The determination of substantial equivalence for this device was based on a detailed device description. The following non-clinical laboratory testing was performed demonstrating that the device can be considered substantially equivalent to the predicate device for the intended use:

- Detailed software description and software verification and validation testing of proprietary software iFit iTTotalTib CR

510(K) SUMMARY (PAGE 5 OF 5)

Conclusion

The determination of substantial equivalence for this device was based on a detailed device description. Non-clinical laboratory testing was performed demonstrating that the device can be considered substantially equivalent to the predicate device for the intended use. Clinical data is not necessary to demonstrate substantial equivalence.

Based on the testing conducted, it is concluded that the modified iTOTAL CR KRS is substantially equivalent to the predicate iTOTAL CR KRS (**K142161** cleared November 14, 2014 and **K152704** cleared October 21, 2015).
