



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

September 30, 2016

Medeon Biodesign, Inc.  
Ms. Greta Chang  
Sr. Manager of Regulatory, Quality & Clinical Affairs  
7F, 116, HouGang St.  
Taipei  
Taiwan 11170 TW

Re: K160117

Trade/Device Name: Abclose - Port Site Closure Device  
Regulation Number: 21 CFR 876.1500  
Regulation Name: Endoscope and accessories  
Regulatory Class: Class II  
Product Code: OCW, GCJ, HCF  
Dated: July 27, 2016  
Received: July 28, 2016

Dear Ms. Greta Chang:

This letter corrects our substantially equivalent letter of September 9, 2016.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent for the indications for use stated in the enclosure to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**David Krause -S**

for Binita S. Ashar, M.D., M.B.A., F.A.C.S.  
Director  
Division of Surgical Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K160117

Device Name

AbClose – Port Site Closure Device

Indications for Use (Describe)

The AbClose - Port Site Closure Device has application in laparoscopic procedures for approximation of tissues and percutaneous suturing for closing incision sites.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*

## 510(k) Summary

<b>510(k) Nwmber</b>	K160117
<b>Date Prepared</b>	16 June 2016
<b>1. <u>Submitter</u></b>	
Mailing Address	<b>Medeon Biodesign, Inc</b> 7F, 116, HouGang St, Taipei, Taiwan 11170 Phone: +886 2 2881 6686
Contact Person	<b>Greta Chang</b> Sr. Manager of Regulatory, Quality & Clinical Affair
Phone:	+886 2 2881 6686
Fax:	+886 2 2881 6685
E-mail:	<a href="mailto:greta@medeonbio.com">greta@medeonbio.com</a>
Date Prepared	
<b>2 <u>Device Name</u></b>	
Common or usual name	Endoscopic tissue approximation device
Trade name	AbClose-Port Site Closure Device
Product Code	OCW, GCJ, HCF
Classification Name	Endoscope and Accessories
CFR Classification	CFR Part 876.1500
Device Class	II
Classification Panel	Gastroenterology/Urology
<b>3 <u>Predicate Device Name</u></b>	
510(k) number:	K132362
Trade or proprietary or model name:	WECK Efx Endo Fascial Closure System
Manufacturer:	Teleflex Medical, Incorporated
<b>4 <u>Device Description:</u></b>	The AbClose - Port Site Closure Device is a sterile, single-use device including 2 major components, Suture Guide and Suture Passer. The device is used with a

commercially cleared suture, and is used as a manual instrument to pass needles with suture through soft tissues for suturing. The device is designed with a suture retrieval system for unassisted fascial closures, facilitating standard suture closure techniques.

The Suture Guide is structured by a Handle and a Suture Catcher. The pyramid-shape Suture Catcher is controlled by a Push Button to fold the Suture Catcher into cylinder shape for insertion into the trocar wound. Once the Suture Guide is in position, Suture Catcher is expanded back to a pyramid shape to stabilize the device on the trocar wound site.

The Suture Passer is a handheld suture grasping device designed to pass sutures through soft tissue. It features a Handle and a stainless steel Needle assembled with a Suture Shaft that grasps a suture. The Suture Passer is designed to work with the Suture Guide to penetrate through soft tissues and deploy suture ends into the Suture Catcher.

Withdraw Suture Guide by fold the Suture Catcher into cylinder shape, the two suture ends are securely captured and are ready for user to type knots to close the trocar wound site.

The AbClose - Port Site Closure Device is intended to be used by clinicians through prescription use only.

5. **Indications for Use:**

The AbClose - Port Site Closure Device has application in laparoscopic procedures for approximation of tissues and percutaneous suturing for closing incision sites.

Special Conditions for Use Statement(s):

For prescription use only

6. **Technological Characteristics and Substantial Equivalence Comparison with Predicate:**

A Comparison of the device feature, intended use, and other information demonstrates that the AbClose - Port Site Closure Device is substantially equivalent to the predicate device as summarized in Table 1.

7. **Materials**

All patient contacting materials are in compliance with ISO 10993-1, as well as FDA Draft Guidance – Use of international Standard ISO 10993, “Biological Evaluation of Medical Devices, Part1: Evaluation and Testing”, dated 04-23-2013

**Table 1**

<b>Similarities</b>		
<b>Device name</b>	<b>Subject device: AbClose - Port Site Closure Device</b>	<b>Predicate device: WECK EFX Endo Fascial Closure System</b>
<b>Indications for use</b>	The AbClose - Port Site Closure Device has application in laparoscopic procedures for approximation of tissue and percutaneous suturing for closing incision sites.	The WECK EFX Endo Fascial Closure System has application in laparoscopic procedures for approximation of tissue and percutaneous suturing for closing incision sites.
<b>Target patient Population</b>	Patient under laparoscopic surgery	Same
<b>Target User Population</b>	Clinician who is qualified to participate a laparoscopic surgery.	Same
<b>Anatomical Site</b>	Abdominopelvic cavity	Same
<b>Where Used</b>	Hospital O.R. room	Same
<b>Contraindications</b>	Do not use where laparoscopic techniques are generally contraindicated	Same
<b>Method of Introduction</b>	-Suture Guide is introduced into abdominopelvic cavity via a 10-15mm trocar port site -Suture Passer is introduced into abdominopelvic cavity by insertion	Same
<b>Compatibility Other Devices</b>	Trocar: 10~ 15 mm	Same
<b>Performance</b>	Maintain pneumoperitoneum and facilitate placement and withdrawal of suture loop	Same
<b>Biocompatible for Intended Use</b>	Limited exposure, external communication device of tissue contact	Same
<b>Differences</b>		
<b>Sterilization Method</b>	Ethylene Oxide sterilization, 10 <sup>-6</sup> SAL	Radiation, 10 <sup>-6</sup> SAL

## **7. Performance Testing**

Performance testing has been carried out to demonstrate that this device meets the performance specification for its intended use. The following tests were performed on the device.

### **Sterilization validation**

Sterilization validation was conducted in accordance with ISO 11135:2014 “Sterilization of health-care products—Ethylene oxide –Requirement for the development, validation and

---

routine control of a sterilization process for medical devices". Sterilization has been validated to achieve a Sterility Assurance Level (SAL) of  $10^{-6}$ .

#### Mechanical testing

Abclose - Port Site Closure Device's mechanical function and structure integrity were tested to demonstrate that the design specification from design input are fulfilled.

Mechanical safety test were also conducted to demonstrate the insertion portion of the device remain intact during the surgery and raise no safety concern.

#### Animal test

The intended use has been conducted by animal test to demonstrate the AbClose - Port Site Closure Device can deliver the intended function of application in laparoscopic procedures for approximation of tissues and percutaneous suturing for closing incision sites.

### **8. Conclusion**

Base on the intended use and/ or indications for use, technological characteristics, performance testing and comparison to predicate device, the AbClose - Port Site Closure Device is substantially equivalent to the predicate device.

---