



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Arthrex, Incorporated
Ms. Ivette Galmez
Regulatory Affairs Specialist
1370 Creekside Boulevard
Naples, Florida 34108-1945

May 10, 2016

Re: K160319

Trade/Device Name: FiberTak DR
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: MBI
Dated: April 20, 2016
Received: April 21, 2016

Dear Ms. Galmez:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

2.5 INDICATIONS FOR USE

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration Indications for Use	Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017 See PRA Statement below.
---	---

510(k) Number (if known)

K160319

Device Name

FiberTak DR

Indications for Use (Describe)

FiberTak DR is intended to be used for suture or tissue fixation in the foot/ankle, knee, hand/wrist, elbow, and shoulder. Specific indications are listed below:

- Elbow: Biceps Tendon Reattachment, Ulnar or Radial Collateral Ligament Reconstruction
- Shoulder: Rotator Cuff Repair, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction
- Hand/Wrist: Scapholunate Ligament Reconstruction, Repair/Reconstruction of collateral ligaments, Repair of Flexor and Extensor Tendons at the PIP, DIP and MCP joints for all digits, digital tendon transfers, Carpal Ligament Reconstruction and Carpometacarpal joint arthroplasty (basal thumb joint arthroplasty)
- Foot/Ankle: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Metatarsal Ligament Repair, Hallux Valgus reconstruction, digital tendon transfers, Mid-foot reconstruction
- Knee: Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis

Type of Use (Select one or both, as applicable)

- Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
 Food and Drug Administration
 Office of Chief Information Officer
 Paperwork Reduction Act (PRA) Staff
 PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

2.1 510(K) SUMMARY OF SAFETY AND EFFECTIVENESS

Date Summary Prepared	April 12, 2016
Manufacturer/Distributor/Sponsor	Arthrex, Inc. 1370 Creekside Boulevard Naples, FL 34108-1945 USA
510(k) Contact	Ivette Galmez Regulatory Affairs Specialist Arthrex, Inc. 1370 Creekside Boulevard Naples, FL 34108-1945 USA Telephone: 239/643.5553, ext. 71263 Fax: 239/598.5508 Email: igalmez@arthrex.com
Trade Name	FiberTak DR
Common Name	Soft Tissue Fixation Device
Product Code - Classification Name CFR	MBI Fastener, fixation, nondegradable, soft tissue 21 CFR 888.3040
Predicate Device	K151230: Arthrex FiberTak Anchors
Purpose of Submission	This special 510(k) premarket notification is submitted to obtain clearance for FiberTak DR .
Device Description	<p><i>FiberTak DR</i> is an "all-suture" soft-tissue fixation device with an expandable push-in design. The anchor and connected sutures are impacted into a pilot hole. The sutures are then manually tensioned to set the anchor by "bulging" the suture sleeve within the pilot hole. Once the anchor is set, the suture is passed around or through soft tissue and tied in a knot; or the repair can be completed in a knotless fashion when used in combination with Arthrex SwiveLock Anchors (K101823).</p> <p><i>FiberTak DR</i> is constructed from a hollow braid of polyester with a suture component combination assembled through the hollow braid. The suture component can be made from either UHMWPE or a polyblend of UHMWPE and Polyester. The anchor configuration of <i>FiberTak DR</i> is similar to the predicate device, except that it combines suture and tape within the same anchor. <i>FiberTak DR</i> is available in three models with combinations in the color of its suture components. <i>FiberTak DR</i> comes preloaded on a disposable inserter made from stainless steel (ASTM F899) whereas the predicate inserter is made of nitinol. The handle of the inserters is made of ABS plastic. <i>FiberTak DR</i> may be sold separately or in a kit with implantation instrumentation. The proposed <i>FiberTak DR</i> is provided sterile for single use.</p>

<p>Intended Use</p>	<p>FiberTak DR is intended to be used for suture or tissue fixation in the foot/ankle, knee, hand/wrist, elbow, and shoulder. Specific indications are listed below:</p> <ul style="list-style-type: none"> • <i>Elbow:</i> Biceps Tendon Reattachment, Ulnar or Radial Collateral Ligament Reconstruction • <i>Shoulder:</i> Rotator Cuff Repair, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction • <i>Hand/Wrist:</i> Scapholunate Ligament Reconstruction, Repair/Reconstruction of collateral ligaments, Repair of Flexor and Extensor Tendons at the PIP, DIP and MCP joints for all digits, digital tendon transfers, Carpal Ligament Reconstruction and Carpometacarpal joint arthroplasty (basal thumb joint arthroplasty) • <i>Foot/Ankle:</i> Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Metatarsal Ligament Repair, Hallux Valgus reconstruction, digital tendon transfers, Mid-foot reconstruction • <i>Knee:</i> Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis
<p>Substantial Equivalence Summary</p>	<p>The proposed FiberTak DR is similar to the predicate device in which the basic design features (all-suture anchor); the materials and intended use are the same.</p> <p>The submitted mechanical testing demonstrates that the pull-out strength of the proposed FiberTak DR exceeds the pull out strength of the predicate device. In addition, the pull out displacements recorded meet the criteria established by the predicate devices and therefore of no clinical relevance. Furthermore, the biocompatibility and packaging/shelf life for the subject device were reviewed and determined to be acceptable due to its similarities in material and packaging.</p> <p>Based on the indication for use, technological characteristics, and the summary of data submitted, ArthroC, Inc. has determined that FiberTak DR is substantially equivalent to currently marketed predicate devices.</p>