



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

October 13, 2016

Hartalega Sdn. Bhd.
Mr. H. Carl Jenkins
Regulatory Counsel to Hartalega Sdn. Bhd.
10 E. Scranton Ave, Suite 201
Lake Bluff, IL 60044

Re: K160562

Trade/Device Name: Nitrile Powder Free Examination Glove with Colloidal Oatmeal USP
Tested for Use with Chemotherapy Drugs

Regulation Number: 21 CFR 880.6250

Regulation Name: Patient Examination Glove

Regulatory Class: I

Product Code: LZA, LZC

Dated: September 7, 2016

Received: September 9, 2016

Dear Mr. Jenkins:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR

Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Michael J. Ryan -S

for Tina Kiang, Ph.D.

Acting Director

Division of Anesthesiology,

General Hospital, Respiratory,

Infection Control, and Dental Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K160562

Device Name

Nitrile Powder Free Examination Glove with Colloidal Oatmeal USP Tested for Use with Chemotherapy Drugs

Indications for Use (Describe)

A non-sterile disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner. These gloves were tested for use with Chemotherapy Drugs in accordance with ASTM D6978-05, Standard Practice for Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs.

The list of Chemotherapy Drugs tested (with breakthrough times) is as below:

Carmustine (3.3 mg/ml)	10.1 minutes
Cisplatin (1.0 mg/ml)	>240 minutes
Cyclophosphamide (20 mg/ml)	>240 minutes
Dacarbazine (10.1 mg/ml)	>240 minutes
Doxorubicin Hydrochloride (2.0 mg/ml)	>240 minutes
Etoposide (20.0 mg/ml)	>240 minutes
Fluorouracil (50.0 mg/ml)	>240 minutes
Methotrexate (25 mg/ml)	>240 minutes
Mitomycin C (0.5 mg/ml)	>240 minutes
Paclitaxel (6.0 mg/ml)	>240 minutes
Thiotepa (10.0 mg/ml)	30.4 minutes
Vincristine Sulfate (1.0 mg/ml)	>240 minutes

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."