



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

September 1, 2016

R&D Systems, Inc.  
Rachel Somsen  
Associate Director, Quality Assurance and Regulatory Affairs  
614 McKinley Place N.E.  
Minneapolis, MN 55413

Re: K160606  
Trade/Device Name: BC-5D Hematology Controls  
Regulation Number: 21 CFR 864.8625  
Regulation Name: Hematology Quality Control Mixture  
Regulatory Class: Class II  
Product Code: JPK  
Dated: August 3, 2016  
Received: August 5, 2016

Dear Ms. Somsen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Kelly Oliner -S**

For

Leonthena R. Carrington, MS, MBA, MT(ASCP)

Director

Division of Immunology and Hematology Devices

Office of *In Vitro* Diagnostics and Radiological

Health

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K160606

Device Name

BC-5D Hematology Controls

Indications for Use (Describe)

BC-5D is an assayed whole blood control designed for Mindray BC-5390 (Mindray 5000 series) Auto Hematology Analyzer for the following parameters: WBC, Neu#, Lym#, Mon#, Eos#, Bas#, Neu%, Lym%, Mon%, Eos%, Bas%, RBC, HGB, HCT, MCV, MCH, MCHC, RDW-CV, RDW-SD, PLT, and MPV.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### **Submitter Information**

R&D Systems, Inc.  
614 McKinley Place N.E.  
Minneapolis, MN 55413  
Contact: Rachel Somsen  
Phone: 612-851-4846  
Fax: 612-379-6580  
Date Prepared: August 26, 2016

### **Device Information**

Proprietary Name:	BC-5D Hematology Controls
Common Name:	Hematology Control
Classification	21 CFR 864.8625
Classification Name:	Hematology Quality Control Mixture
Product Code:	JPK
Device Class:	II
Panel:	Hematology (81)

### **Predicate Device**

R&D Systems CBC-XE Hematology Control, (K042094)

### **Instrument**

Mindray BC-5390 Hematology Analyzer (K160429)

### **Description of Device**

The Mindray BC-5390 Hematology Analyzer is a quantitative, automated hematology analyzer and leukocyte differential counter for In Vitro Diagnostic Use in clinical laboratories. It is only to be used by trained medical professionals to identify the normal patient, with all normal system-generated parameters, and to flag or identify patient results that require additional studies.

The BC-5D Hematology Control consists of three levels of controls (Low, Normal, and High). It is recommended to perform quality control check using these controls established by the laboratory procedures and/or local/national regulations.

BC-5D Hematology Control is an *in vitro* diagnostic reagent composed of human erythrocytes, simulated leukocytes, and mammalian platelets in a plasma-like fluid with preservatives. This control is contains stabilized materials that provide a means of monitoring the performance of hematology blood cell counters. It is sampled in the same manner as a patient specimen.

**Intended Use**

BC-5D is an assayed whole blood control designed for Mindray BC-5390 (Mindray 5000 series) Auto Hematology Analyzer for the following parameters: WBC, Neu#, Lym#, Mon#, Eos#, Bas#, Neu%, Lym%, Mon%, Eos%, Bas%, RBC, HGB, HCT, MCV, MCH, MCHC, RDW-CV, RDW-SD, PLT, and MPV.

**Technological Comparison to Predicate**

The subject device has the same technological characteristics as the legally marketed predicate device. The BC-5D Hematology Control is intended to be used to monitor values obtained on Mindray BC-5000 series hematology analyzers. The predicate device is used to monitor values obtained on Sysmex Hematology Instruments.

The predicate device is assayed for : WBC, RBC, HGB, HCT, MCV, MCH, MCHC, RDW-SD, RDW-CV, PLT, PDW, MPV, P-LCR, IRF, NEUT#, LYMPH#, MONO#, EO#, BASO#, NRBC#, RET#, Neut%, Lymph%, Mono%, Eo%, Baso%, Nrbc%, and Ret%.

**Summary of Performance Data**

Following the manufacture and analysis of three verification lots from which the “assay range” for each parameter was determined, 3 validation lots of BC-5D Hematology Controls were manufactured and tested. Results verified the product to have substantial equivalence in performance, precision and stability to the predicate device. The BC-5D Hematology Control met the acceptance criteria, as determined during verification, over the life of the product. Expiration dating is established at 75 days (closed vial) and 14 days (open vial) when stored at 2 - 8° C, and handled according to instructions for use.

**Substantial Equivalence Conclusion**

The data demonstrates that the BC-5D Hematology Control is substantially equivalent to the legally marketed predicate device.