



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

August 25, 2016

CareFusion 2200, Inc.  
Erika Fernandez  
Regulatory Affairs Manager  
75 N. Fairway Dr.  
Vernon Hills, Illinois 60061

Re: K160764

Trade/Device Name: AirLife Autofill Humidification Chamber  
Regulation Number: 21 CFR 868.5450  
Regulation Name: Respiratory Gas Humidifier  
Regulatory Class: Class II  
Product Code: BTT  
Dated: July 25, 2016  
Received: July 26, 2016

Dear Erika Fernandez:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

*Tejashri Purohit-Sheth, M.D.*  **Tejashri Purohit-Sheth, M.D.**  
Clinical Deputy Director  
DAGRID/ODE/CDRH FOR

Tina Kiang, Ph.D.  
Acting Division Director  
Division of Anesthesiology,  
General Hospital, Respiratory,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K160764

Device Name

AirLife Autofill Humidification Chamber

Indications for Use (Describe)

The AirLife Autofill Humidification Chamber is intended to hold water required to humidify breathing gases delivered to any patient using a heated humidifier. The product is single use device, non-sterile and used in professional healthcare environments under a doctor's supervision and by skilled clinician. The AirLife Autofill Humidification Chamber is compatible with the Fisher and Paykel MR850 system.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number"*

## Summary of Safety and Effectiveness

510k summary complying with 21 CFR 807.92.

### 1. SUBMITTER

CareFusion 2200, Inc.  
75 N Fairway Drive, Vernon Hill, IL 60061  
Phone: 847-362-8097  
Fax: 312 949-0731  
Contact Person: Erika Fernandez  
Date Prepared: August 24<sup>th</sup>, 2016

### 2. Device

Product Name: AirLife Autofill Humidification Chamber  
Device Name: Humidifier Chamber  
Common Name: Humidifier Chamber  
Classification Name: Humidifier, Respiratory Gas (21 CFR 868.5450)  
Regulatory Class: II  
Product Code: BTT

### 3. Predicate Device

This submission demonstrates substantial equivalence to the MR290 Humidification Chamber, K934140 cleared on January 24, 1994. This predicate device has not been subject to a design-related recall.

### 4. Device Description

The AirLife Autofill Humidification Chamber is intended to hold water required to humidify breathing gases delivered to patients.

### 5. Indication for use

The AirLife Auto Fill Humidification Chamber is intended to hold water required to humidify breathing gases delivered to any patient using a heated humidifier. The product is single use device, non-sterile and used in professional healthcare environments under a doctor's supervision and by skilled clinician. The AirLife Auto Fill Humidification Chamber is compatible with Fisher and Paykel MR850 system.

## 6. Comparison of technological characteristics with the predicate device

The fundamental scientific technology is the same for both the proposed and predicate device. The AirLife Humidification Chamber is substantially equivalent to the predicate device, MR290 Humidification Chamber, regarding safety, effectiveness, design (technology), materials and intended use.

Element of comparison	Proposed Device	Predicate Device
<b>Indications for Use</b>	The AirLife Autofill Humidification Chamber is intended to hold water required to humidify breathing gases delivered to any patient using a heated humidifier. The product is a single use device, non-sterile and used in professional healthcare environments under a doctor's supervision and by skilled clinicians.	The MR290 Humidification Chamber is intended to hold water required to humidify the air being delivered to patients. The MR290 is an auto-fill humidification chamber suitable for all patients and compatible with all F&P MR-series humidifiers
<b>Principal of Operation</b>	The humidification chamber works with a respiratory humidifier (also called a "heater base") that provides the heat source, temperature control system, and alarm system to heat and humidity respiratory gases.	The humidification chamber works with a respiratory humidifier (also called a "heater base") that provides the heat source, temperature control systems, and alarm systems to heat and humidifies respiratory gases.
<b>Characteristics</b>		
<b>Connection</b>	22 mm ISO conical connections for breathing circuit attachment Flexible tubing bonded to close piercing device (spike) and chamber inlet for supplying water	22 mm ISO conical connections for breathing circuit attachment Flexible tubing bonded to close piercing device (spike) and chamber inlet for supplying water
<b>Design</b>	Maximum and minimum waterline marks for visual confirmation of proper water levels	Maximum water line mark for visual confirmation of proper water levels
	Independent primary and secondary floats to regulate water level and to prevent overflow condition if the primary float fails.	Independent primary and secondary floats to regulate water level and prevent overflow condition if the primary float fails open.
	Clear Housing to allow visibility of water level for monitoring	Clear Housing to allow visibility of water level for monitoring
<b>Humidity Output</b>	Invasive mode: 39.0 mg/L Non Invasive mode: 24.3 mg/L	Invasive mode: 39.2 mg/L Non Invasive mode: 24.6 mg/L
<b>Maximum Operating pressure</b>	13.2kPa	13.2kPa

Element of comparison	Proposed Device	Predicate Device
<b>Maximum Continuous Gas Flow Rate</b>	Maximum continuous gas flow rate of 60 LPM in invasive mode and 120 LPM in non-Invasive mode	Maximum continuous gas flow rate of 60 LPM in invasive mode and 120 LPM in non-Invasive mode
<b>Compliance</b>	6.23 mL/kPa	5.54 mL/kPa
<b>Resistance to Flow @60L/min</b>	0.40 cmH <sub>2</sub> O	0.50 cmH <sub>2</sub> O
<b>Enthalpy Maximum Value Enthalpy Averaged Value</b>	156 kJ/kg Max. 117 kJ/kg Avg.	151 kJ/kg Max. 120 kJ/kg Avg.
<b>Compressible volume</b>	245 mL	255 mL
<b>Material</b>	Uses Noryl for float frame	Uses Polycarbonate for float frame
<b>Gas leakage</b>	<30mL/min	<100 mL/min
<b>Duration for Use</b>	14 days	14 days
<b>Shelf life</b>	24 months	Not Published
<b>Compatibility (1)</b>	Compatible with flexible and rigid inhalation water containers (water feed set contains air vent provision for rigid containers)	Compatible with flexible and rigid inhalation water containers (water feed set contains air vent provision for rigid containers)
<b>Compatibility (2)</b>	Compatible with Fisher and Paykel MR850 Humidifiers	Compatible with Fisher and Paykel MR series Humidifiers

## 7. Performance Data

The proposed device was tested to ensure compliance to the following standards:

### Biocompatibility

Tests for an externally communicating, tissue by way of gas path and direct mucosal contact with prolonged contact (greater than 24 hours but less than 30 days): Cytotoxicity, Sensitization, Irritation, Muscle Implantation, Genotoxicity and Extractables/Leachables

## Standards

Performance Characteristic	Standard
Biological Evaluation of Medical Devices--Part 1: Evaluation and Testing	AAMI/ANSI/ISO 10993-1:2009
Biological Evaluation of Medical Devices-Part 5: Tests for In Vitro Cytotoxicity	AAMI/ANSI/ISO 10993-5:2009 (R2014)
Biological Evaluation of Medical Devices-Part 10: Tests for Irritation and Skin Sensitization.	AAMI/ANSI/ISO 10993-10:2010 (R2014)
Biological Evaluation of Medical Devices Part 12: Sample Preparation and Reference Materials	AAMI/ANSI/ISO 10993-12:2012
Biological Evaluation of Medical Devices Part 17: Establishment of allowable limits for leachable substances	AAMI/ANSI/ISO 10993-17:2002
Biological Evaluation of Medical Devices Part 18: Chemical characterization of materials	AAMI/ANSI/ISO 10993-18:2005

## Performance

The following tests were performed for the proposed device to support the substantial equivalence decision.

Performance Characteristic	Relevant Standard
Humidity Output Invasive	BS EN ISO 8185:2009
Humidity Output Non-Invasive	
Specific Enthalpy	BS EN ISO 8185:2009
Hazardous Output	BS EN ISO 8185:2009

Performance Characteristic	Relevant Standard
Resistance to Flow	BS EN ISO 5367:2014
Maximum Peak Flow	BS EN ISO 8185:2014
Leakage & Maximum Operating Pressure	BS EN ISO 8185:2009 BS EN ISO 5367:2014
Liquid Overflow	BS EN ISO 8185:2009
Compliance	BS EN ISO 8185:2009 BS EN ISO 5367:2014
Compressible Volume	N/A
Accessible Surface Temperature	BS EN ISO 8185:2009 IEC 60601-1:2005
Steady State Noise	BS EN ISO 8185:2009 IEC 60601-1:2005
Shelf Life/Accelerated Aging	N/A
End-use Simulation/Duration of Use	N/A

## 8. Conclusion

Test results demonstrate that the proposed device is as safe and effective as the predicate device, and therefore, is substantially equivalent to the predicate device.