



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

December 22, 2016

Life Spine, Inc.
Mr. Randy Lewis
General Manager
13951 S. Quality Drive
Huntley, Illinois 60142

Re: K161127

Trade/Device Name: The Small PLATEAU[®] (PLATEAU-C) Spacer System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral body fusion device
Regulatory Class: Class II
Product Code: ODP
Dated: November 23, 2016
Received: November 28, 2016

Dear Mr. Lewis:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Vincent J. Devlin -S

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K161127

Device Name

The Small PLATEAU® (PLATEAU-C) Spacer System

Indications for Use (Describe)

The Small PLATEAU (PLATEAU-C) Spacer System is intended for spinal fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one disc level (C2-T1). DDD is defined as discogenic pain with degeneration of the disc confirmed by history and radiographic studies. This device is intended to be used with autogenous bone graft and a supplemental internal spinal fixation system.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary
The Small PLATEAU® (PLATEAU-C) Spacer System

Submitted By: Life Spine Inc.
13951 S. Quality Drive
Huntley, IL 60142
Telephone: 847-884-6117
Fax: 847-884-6118

510(k) Contact: Randy Lewis
Life Spine, Inc.
13951 S. Quality Drive
Huntley, IL 60142
Telephone: 847-884-6117
Fax: 847-884-6118

Date Prepared: December 20th, 2016

Trade Name: The Small PLATEAU® (PLATEAU-C) Spacer System

Common Name: Intervertebral Body Fusion Device

Classification: ODP, 21 CFR 888.3080, Class II

Primary Predicate : Small PLATEAU (PLATEAU-C) Spacer System (K111835)

Additional Predicate: Titan Spine Endoskeleton TC IBF (K100889)

Device Description:

The Small PLATEAU (PLATEAU-C) Spacer System is intended to serve as an intervertebral body fusion device. The implant is available in a range of sizes and footprints to suit the individual pathology and anatomical conditions of the patient. It is fabricated and manufactured from either Titanium (Ti-6Al-4V ELI) or Polyetheretherketone (PEEK-OPTIMA LT1) with radiographic markers. The implant is hollow to permit packing with autogenous bone graft to help promote intervertebral body fusion. The superior and inferior surfaces have teeth to assist in the interface with the vertebral endplates to prevent rotation and/or migration.

All implants are intended for single use only and should not be reused under any circumstances. **Do not use any of the Small PLATEAU (PLATEAU-C) Spacer System components with components from any other system or manufacturer. The Small PLATEAU (PLATEAU-C) Spacer System components should never be reused under any circumstances.**

Intended Use of the Device:

The Small PLATEAU (PLATEAU-C) Spacer System is intended for spinal fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one disc level (C2-T1). DDD is defined as discogenic pain with degeneration of the disc confirmed by history and radiographic studies. This device is intended to be used with autogenous bone graft and a supplemental internal spinal fixation system.

Technological Characteristics:

The Small PLATEAU (PLATEAU-C) Spacer System is substantially equivalent to the predicate systems in terms of design, materials, indications for use and sizing.

Performance Data:

Finite element analysis and engineering rationale was presented to demonstrate the substantial equivalency of the Small PLATEAU (PLATEAU-C) Spacer System.

Substantial Equivalence:

The Small PLATEAU (PLATEAU-C) Spacer System was shown to be substantially equivalent to the predicate devices in indications for use, design, function, materials used and mechanical performance.