



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

December 22, 2016

Feg Textiltechnik Forschungs-und Entwicklungsgesellschaft
% Mr. Neil Armstrong
RA Advisor To Feg Textiltechnik
Meddiquest Limited
Herlington House
Orton Malborne
Peterborough, PE2 5XS GB

Re: K161550

Trade/Device Name: DynaMesh® -VENTRAL
Regulation Number: 21 CFR 878.3300
Regulation Name: Surgical Mesh
Regulatory Class: Class II
Product Code: FTL
Dated: November 18, 2016
Received: November 23, 2016

Dear Mr. Armstrong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may; therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

David Krause -S

for Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K161550

Device Name

DynaMesh®-VENTRAL

Indications for Use (Describe)

DynaMesh®-VENTRAL serves to support the tissue and stabilize the fascial structures of the abdominal wall.

DynaMesh®-VENTRAL is indicated for repairing hernias in all current extraperitoneal surgical techniques.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



FEG Textiltechnik
Forschungs- und
Entwicklungsgesellschaft mbH
Prager Ring 70
D-52070 Aachen
Germany
Phone: 011 49 241 1892374-0
Fax: 011 49 241 1892374-59

510(k) Summary

(as required by 21 CFR 807.92(c))

Owner's Name:

FEG Textiltechnik Forschungs- und Entwicklungsgesellschaft mbH
Prager Ring 70
D - 52070 Aachen
Germany
Tel: +49 241 189 23 74 - 0
Fax: +49 241 189 23 74 - 59
Contact Person: Stefan Schneemelcher
e-mail: schneeme@feg-textiltechnik.de

Date this Summary was Prepared: December 20, 2016

Classification:

Name: Mesh, Surgical, Polymeric
Regulation: 21 CFR 878.3300
Product Code: 79 FTL
Regulatory Class: II

Common/Usual Name:

Mesh Implant used in Hernia Surgery

Proprietary Name:

DynaMesh®-VENTRAL

Predicate Devices used to Demonstrate Substantial Equivalence:

The FEG Textiltechnik DynaMesh®-VENTRAL is substantially equivalent to:

Product Name: DynaMesh®-CICAT, K131530
Manufacturer: FEG Textiltechnik, Aachen, Germany

Description, including Intended Use

The FEG Textiltechnik Forschungs- und Entwicklungsgesellschaft mbH DynaMesh®-CICAT is intended for hernia surgery for reinforcing connective tissue structures. DynaMesh®-VENTRAL serves to support the tissue and stabilize the fascial structures of the abdominal wall. DynaMesh®-VENTRAL is indicated for repairing hernias in all current extraperitoneal surgical techniques.

The FEG Textiltechnik Forschungs- und Entwicklungsgesellschaft mbH DynaMesh®-VENTRAL is a textile structure knitted from non absorbable, bio-stable Polyvinylidene fluoride (PVDF) monofilament fiber.

The grade of Polyvinylidene Fluoride (PVDF) used is SOLEF®1008 of Solvay Solexis.

The knitted mesh structure consists of uncoloured PVDF Monofilaments with a zig-zag pattern of green-coloured PVDF Monofilaments as visual aid to the clinician that facilitates correct mesh positioning; the mesh is correctly positioned when the zig-zag-stripes are in the craniocaudal direction.

The used green dye is Phtalocyanine green, CAS1328-53-6,PG7 of BASF.

DynaMesh®-VENTRAL is supplied as a sterile, flexible, flat sheet of material without any coatings. The FEG Textiltechnik's DynaMesh®-VENTRAL will be packed in a cardboard box. Within the cardboard box the device will be double-packed in two paper-poly pouches suitable for Ethylene Oxide sterilization; the labeled outer pouch contains the inner pouch and four adhesive labels for use on patient records.

Technological Characteristics Compared to Predicate Devices

The FEG Textiltechnik DynaMesh®-VENTRAL is:

- for the same intended use,
- knitted with the same equipment by the same personnel in the same facilities,
- packaged in same materials by same processes in the same facilities
- labeled with the same label format by the same processes in the same facilities
- contract sterilized in the same chamber with the same cycle at the same subcontract facility

as the FEG Textiltechnik DynaMesh®-CICAT K131530.

	FEG Textiltechnik DynaMesh®-VENTRAL	FEG Textiltechnik DynaMesh®-CICAT
Feature		
Clearance/Approval	Subject Device	K131530
Structure	knitted mesh	same
Material	100% PVDF Monofilament	same
Manufacturing Process	Knitted by FEG	same
Packaging Materials	poly, paper, cardboard	same
Packaging Process	Packed by FEG	same
Sterilization Process	Sterilized by Rose GmbH	same

Summary of Non-Clinical and Clinical Data

Various Laboratory Bench Tests have been conducted on DynaMesh®-VENTRAL to demonstrate performance to an accepted norm and/or substantial equivalence to the predicate products:

- Burst Strength Test
- Tensile Strength
- Maximum elongation
- Suture Pullout Strength Test
- Tear Resistance Testing
- Porosity Test
- Mesh stiffness

Conclusion

From review of the non-clinical and clinical data, FEG Textiltechnik conclude that the DynaMesh®-VENTRAL is substantially equivalent in terms of safety, effectiveness and performance to the predicate device.