



December 6, 2016

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Innovative Health, LLC
Rafal Chudzik
VP, R&D and Operations
1435 North Hayden Road
Suite 100
Scottsdale, Arizona 85257

Re: K161769

Trade/Device Name: Reprocessed Supreme Diagnostic Electrophysiology Catheters
(see attached list of models)

Regulation Number: 21 CFR 870.1220

Regulation Name: Electrode Recording Catheter or Electrode Recording Probe

Regulatory Class: Class II

Product Code: NLH

Dated: November 4, 2016

Received: November 7, 2016

Dear Rafal Chudzik:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address


<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,



Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

The model numbers included in the scope of this submission (K161769) are as follows:

Description	Item Number	French Size	Electrode Spacing (mm)	Curve Type	Usable Length (cm)	Number of Electrodes
Supreme Diagnostic EP Catheter	401957	5	10	CRD	120	2
	401978	6	2-2-2	CSL	65	20
	402033	6	2-2-2-2-2-2-2-2-45-2-2-2-2-2-2-2-2	CSL	120	20
	402020	6	2-5-2-5-2-5-2-5-2-60-2-10-2-10-2-10-2-10-2	CRD	120	20
	401956	6	10	CRD	120	2
	401893	4	2-8-2	CSL	65	10
	401979	4	2-5-2	CSL	65	10
	401863	5	2-8-2	CSL	120	10
	401864	5	2-8-2	CSL	65	10
	401865	5	5-5-5	CSL	120	10
	401966	4	5-5-5	CRD-2	120	6
	401967	4	2-2-2	CRD-2	120	6
	401968	4	2-5-2	CRD-2	120	6
	401872	5	2-2-2	CRD-1	120	6
	402008	5	2-5-2	CRD-2	120	6
	402009	5	5-5-5	CRD-2	120	6
	402010	6	2-5-2	CRD-2	120	6
	402011	6	5-5-5	CRD-2	120	6
	401876	6	5-5-5-175-175	CRD-1	120	6
	401877	6	5-5-5-175-175	JSN	120	6
	401891	4	5-5-5	CRD	120	4
	401890	4	5-5-5	JSN	120	4
	401892	4	5-5-5	DAO	120	4
	401993	4	2-2-2	JSN	120	4
	401994	4	2-5-2	JSN	120	4
	401996	4	2-5-2	CRD	120	4
	402012	4	5-5-5	CRD-2	120	4
	401433	5	10	CRD	120	4
	401859	5	10	CRD-2	120	4
	401435	5	10	JSN	120	4
	401441	5	5	CRD	120	4
	401860	5	5	CRD-2	120	4
	401443	5	5	JSN	120	4
	401444	5	5	DAO	120	4
	401466	5	5	DAO-1	120	4
	401448	5	2-5-2	CRD	120	4
	401450	5	2-5-2	JSN	120	4
	401468	5	2-5-2	DAO-1	120	4
	401871	5	2	CRD-1	120	4
	401434	6	10	CRD	120	4
	401436	6	10	JSN	120	4
	401438	6	10	DAO	120	4
	401442	6	5	CRD	120	4
	401474	6	5	CRD-1	120	4
	402004	6	5	CRD-2	120	4
	401430	6	5	JSN	120	4
	401475	6	5	JSN-1	120	4
	401445	6	5	DAO	120	4
	401449	6	2-5-2	CRD	120	4
	401451	6	2-5-2	JSN	120	4
	401453	6	2-5-2	DAO	120	4
	401950	6	5	CRD	120	4
	401951	6	5	JSN	120	4
	401952	6	10	CRD	120	4

Table 1: Device Scope

Indications for Use

510(k) Number (if known)
K161769

Device Name
Reprocessed Supreme Diagnostic Electrophysiology Catheter

Indications for Use (Describe)

The Reprocessed Supreme Diagnostic Electrophysiology Catheters can be used in the evaluation of a variety of cardiac arrhythmias from endocardial and intravascular sites.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

SECTION 5: 510(k) SUMMARY

As required by 21 CFR 807.92(c)

Submitter's Name and Address:

Innovative Health, LLC.
1435 N. Hayden Road, Suite 100
Scottsdale, AZ 85257

Contact Name and Information:

Rafal Chudzik
VP, R&D and Operations
Innovative Health, LLC.
(480) 525-6006 (office)
(844) 965-9359 (fax)
rchudzik@innovative-health.com

Date prepared:

June 24, 2016

Device Information:

<i>Trade/Proprietary Name:</i>	Reprocessed Supreme Diagnostic Electrophysiology Catheters
<i>Common Name:</i>	Diagnostic Electrophysiology Catheter
<i>Classification Name:</i>	Catheter, Recording, Electrode, Reprocessed
<i>Classification Number:</i>	Class II, 21 CFR 870.1220
<i>Product Code:</i>	NLH

Predicate Device:

510(k) Number	510(k) Device	Manufacturer
K002976	Response Electrophysiology Catheter, Supreme Electrophysiology Catheter	St. Jude Medical, Daig Division
K894500	Daig Electrophysiology Catheter	Daig Corp.

Device Description:

The Reprocessed Supreme Diagnostic Electrophysiology Catheters are manufactured in various fixed curves and electrode spacing for electrophysiological mapping for the evaluation of a variety of cardiac arrhythmias from endocardial and intravascular sites.

The item numbers included in the scope of this submission are as follows:

Description	Item Number	French Size	Electrode Spacing (mm)	Curve Type	Usable Length (cm)	Number of Electrodes
Supreme Diagnostic EP Catheter	401957	5	10	CRD	120	2
	401978	6	2-2-2	CSL	65	20
	402033	6	2-2-2-2-2-2-2-2-2-45-2-2-2-2-2-2-2-2-2	CSL	120	20
	402020	6	2-5-2-5-2-5-2-5-2-60-2-10-2-10-2-10-2-10-2	CRD	120	20
	401956	6	10	CRD	120	2
	401893	4	2-8-2	CSL	65	10
	401979	4	2-5-2	CSL	65	10
	401863	5	2-8-2	CSL	120	10
	401864	5	2-8-2	CSL	65	10
	401865	5	5-5-5	CSL	120	10
	401966	4	5-5-5	CRD-2	120	6
	401967	4	2-2-2	CRD-2	120	6
	401968	4	2-5-2	CRD-2	120	6
	401872	5	2-2-2	CRD-1	120	6
	402008	5	2-5-2	CRD-2	120	6
	402009	5	5-5-5	CRD-2	120	6
	402010	6	2-5-2	CRD-2	120	6
	402011	6	5-5-5	CRD-2	120	6
	401876	6	5-5-5-175-175	CRD-1	120	6
	401877	6	5-5-5-175-175	JSN	120	6
	401891	4	5-5-5	CRD	120	4
	401890	4	5-5-5	JSN	120	4
	401892	4	5-5-5	DAO	120	4
	401993	4	2-2-2	JSN	120	4
	401994	4	2-5-2	JSN	120	4
	401996	4	2-5-2	CRD	120	4
	402012	4	5-5-5	CRD-2	120	4
	401433	5	10	CRD	120	4
	401859	5	10	CRD-2	120	4
	401435	5	10	JSN	120	4
	401441	5	5	CRD	120	4
	401860	5	5	CRD-2	120	4
	401443	5	5	JSN	120	4
	401444	5	5	DAO	120	4
	401466	5	5	DAO-1	120	4
	401448	5	2-5-2	CRD	120	4
	401450	5	2-5-2	JSN	120	4
	401468	5	2-5-2	DAO-1	120	4
	401871	5	2	CRD-1	120	4
	401434	6	10	CRD	120	4
	401436	6	10	JSN	120	4
	401438	6	10	DAO	120	4
	401442	6	5	CRD	120	4
	401474	6	5	CRD-1	120	4
	402004	6	5	CRD-2	120	4
	401430	6	5	JSN	120	4
	401475	6	5	JSN-1	120	4
	401445	6	5	DAO	120	4
	401449	6	2-5-2	CRD	120	4
	401451	6	2-5-2	JSN	120	4
	401453	6	2-5-2	DAO	120	4
	401950	6	5	CRD	120	4
	401951	6	5	JSN	120	4
	401952	6	10	CRD	120	4

Table 5.1: Device Scope

Indications for Use:

The Reprocessed Supreme Diagnostic Electrophysiology Catheters can be used in the evaluation of a variety of cardiac arrhythmias from endocardial and intravascular sites.

Technological Characteristics:

The purpose, design, materials, function, and intended use of the Reprocessed Supreme Diagnostic Electrophysiology (EP) Catheters are identical to the predicate devices. There are no changes to the claims, clinical applications, patient population, performance specifications, or method of operation. In addition, Innovative Health's reprocessing of these devices includes removal of visible soil and decontamination. Each device is inspected and function tested prior to packaging and labeling.

Functional and Safety Testing:

Bench and laboratory testing was conducted to demonstrate performance (safety and effectiveness) of the Reprocessed Supreme Diagnostic EP Catheter. This included the following:

- Biocompatibility
- Cleaning Validation
- Sterilization Validation
- Functional Testing
 - Visual Inspection
 - Dimensional Verification
 - Electrical Continuity and Resistance
 - Simulated Use
 - Mechanical Characteristics
- Electrical Safety Testing
 - Dielectric and Current Leakage
- Packaging Validation

The Reprocessed Supreme Diagnostic EP Catheters are reprocessed no more than one (1) time. Each device is marked and tracked. After the device has reached the maximum number of reprocessing cycles, the device is rejected from further reprocessing. Reprocessing is performed only by Innovative Health. Innovative Health restricts its reprocessing to exclude devices previously reprocessed by other reprocessors.

Conclusion:

Innovative Health concludes that the Reprocessed Supreme Diagnostic EP Catheters are as safe and effective as the predicate devices described herein.