



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

September 15, 2016

Fiagon GmbH
% Ms. Yarmela Pavlovic
Regulatory Consultant
Hogan Lovells
3 Embarcadero Center, Suite 1500
San Francisco, CA 94111

Re: K161940
Trade/Device Name: Guidewire 0.6 Single Use
Regulation Number: 21 CFR 882.4560
Regulation Name: Stereotaxic Instrument
Regulatory Class: Class II
Product Code: PGW
Dated: August 22, 2016
Received: August 22, 2016

Dear Ms. Pavlovic:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Eric A. Mann -S

for Malvina B. Eydelman, M.D.
Director
Division of Ophthalmic and Ear,
Nose and Throat Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K161940

Device Name

GuideWire 0.6 Single Use

Indications for Use (Describe)

The GuideWire 0.6 Single Use is intended as an aid for precisely locating anatomical structures in either open or percutaneous procedures. It is indicated for use with the Fiagon Navigation system using electromagnetic navigation.

It is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure in the field of ENT surgery, such as the paranasal sinuses, mastoid anatomy, can be identified relative to a CT or MR based model of the anatomy.

Example procedures include, but are not limited to:

ENT Procedures;

Transphenoidal access procedures.

Intranasal procedures.

Sinus procedures, such as Maxillary antrostomies, Ethmoidectomies, Sphenoidotomies/Sphenoid explorations, Turbinate resections, and Frontal sinusotomies.

ENT related anterior skull base procedures.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

SECTION 5 – 510(K) SUMMARY

510(k) Summary

July 14, 2016

1. Submitter Information/ 510(k) Holder

Submitter: Fiagon GmbH
Address: Neuendorfstrasse 23b
16761 Hennigsdorf, Germany

Telephone: +49 3302 201 21 10
Telefax: +49 3302 201 21 15

Contact: Mr. Dirk Mucha, CTO

2. Device Information

Trade Name: Fiagon Navigation – GuideWire 0.6 Single Use
Common Name: Image guided surgery system
Classification: Class II per 21 CFR 882.4560
Device: Ear, Nose, and Throat Stereotaxic Instrument
Product Code: PGW

3. Purpose of Submission

The purpose of this submission is to gain clearance for a modified version of the previously cleared instrument, Fiagon Navigation - Guidewire 0.6 (K160369).

4. Predicate Device Information

Fiagon Navigation - Guidewire 0.6 (K160369)

5. Device Description

The Fiagon Navigation – GuideWire 0.6 Single Use is a disposable instrument intended to be used with the Fiagon Navigation system. The instrument is an electromagnetically navigated pointing device (malleable, sensor within the tip).

Each device incorporates a sensor device, which is tracked by the navigation system within the low-energy magnetic field of a field generator (part of the navigation system). The navigation software (part of the navigation system) displays the position of the instruments in preoperative scans (e.g., CT, MRI, fluoroscopy).

6. Intended Use

The GuideWire 0.6 Single Use is intended as an aid for precisely locating anatomical structures in either open or percutaneous procedures. It is indicated for use with the Fiagon Navigation system using electromagnetic navigation.

It is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure in the field of ENT surgery, such as the paranasal sinuses, mastoid anatomy, can be identified relative to a CT or MR based model of the anatomy.

Example procedures include, but are not limited to:

ENT Procedures;

Transphenoidal access procedures.

Intranasal procedures.

Sinus procedures, such as Maxillary antrostomies, Ethmoidectomies,

Sphenoidotomies/Sphenoid explorations, Turbinate resections, and Frontal sinusotomies.

ENT related anterior skull base procedures.

7. Comparison of Technological Characteristics

The GuideWire 0.6 Single Use is a modified version of the previously cleared GuideWire 0.6 (K160369). The reason for this Special 510(k) is to describe a change in material as well as a change from a reprocessed device to one that is provided sterile and is Single Use and disposable. Sterilization validation tests reports are provided to demonstrate that these differences do not raise new issues of safety and effectiveness.

8. Performance Data

Testing was performed in order to determine device precision and accuracy of the modified device.

Testing was also completed to ensure functionality and compatibility with the Fiagon Navigation system. Also testing was performed to ensure sterilization of the modified disposable instrument does not alter the performance characteristics of the device.

9. Conclusion

Based on the indications for use, technological characteristics, performance testing, and comparison to the predicates, the GuideWire 0.6 Single Use has been shown to be substantially equivalent to the comparable device GuideWire 0.6 and the modified device does not present any new issues of safety or effectiveness.