



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

December 13, 2016

BTL Industries, Inc.  
% Jan Zarsky  
Director  
47 Loring Drive  
Framingham, MA 01702

Re: K162010  
Trade/Device Name: HPM-6000U  
Regulation Number: 21 CFR 876.5320  
Regulation Name: Nonimplanted Electrical Continence Device  
Regulatory Class: Class II  
Product Code: KPI  
Dated: September 15, 2016  
Received: September 16, 2016

Dear Jan Zarsky,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Joyce M. Whang -S**

for

Benjamin R. Fisher, Ph.D.  
Director  
Division of Reproductive, Gastro-Renal,  
and Urological Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

Not known  
**K162010**

Device Name  
HPM-6000U

Indications for Use (Describe)

HPM-6000U is intended to provide entirely non-invasive electromagnetic stimulation of pelvic floor musculature for the purpose of rehabilitation of weak pelvic muscles and restoration of neuromuscular control for the treatment of urinary incontinence in women.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### General Information

Sponsor: BTL Industries, Inc.  
47 Loring Drive  
Framingham, MA 01702  
Tel: [+1-866-285-1656](tel:+1-866-285-1656)  
Fax: +1-888-499-2502

Applicant: BTL Industries, Inc.  
47 Loring Drive  
Framingham, MA 01702  
Tel: [+1-866-285-1656](tel:+1-866-285-1656)  
Fax: +1-888-499-2502

Contact Person: Jan Žársky  
Director  
BTL Industries, Inc.  
[zarsky@btl.net](mailto:zarsky@btl.net)

Summary Preparation  
Date: July 19, 2016

### Device Name

Trade/Proprietary Name: HPM-6000U  
Primary Classification Name: Stimulator, Electrical, Non-Implantable, For Incontinence  
Classification Regulation: 21 CFR 876.5320, Class II  
Classification Product Code: KPI

### Legally Marketed Predicate Devices

The HPM-6000U is a non-invasive therapeutic device, and is substantially equivalent to the current product that is already cleared for distribution in the USA under the following 510(k) Premarket Notification number:

Neotonus Model 1000 Muscle Stimulator System (K973096)

## Product Description

The HPM-6000U is a non-invasive therapeutic device. The device produces electromagnetic field that interacts with the tissues of the human body. The electromagnetic field delivered in the muscular or neuronal tissue area is triggering the stimulation and tonisation.

The HPM-6000U is equipped with a color touch screen with wide view angle that significantly facilitates the use of the device. The on-screen information guides the user step-by-step through the entire therapy procedure. The therapeutic parameters are easily set using the touch screen, buttons and knob on the device. During the therapy the device keeps information about the applied therapy type, remaining therapy time and main therapy parameters on the screen.

## Intended Use

HPM-6000U is intended to provide entirely non-invasive electromagnetic stimulation of pelvic floor musculature for the purpose of rehabilitation of weak pelvic muscles and restoration of neuromuscular control for the treatment of urinary incontinence in women.

## Non-clinical Testing

The HPM-6000U device has been thoroughly evaluated for electrical safety. The HPM-6000U has been found to comply with the following applicable medical device safety standards:

ISO 14971	Medical devices – Application of risk management to medical devices
IEC 62304	Medical device software – Software life cycle processes
IEC 60601-1	Medical electrical equipment – Part 1: General requirements for basic safety and essential performance
IEC 60601-1-2	Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility – Requirements and tests
IEC 60601-1-6	Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability
ISO 10993-1	Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process
ISO 10993-5	Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity
ISO 10993-10	Biological evaluation of medical devices – Part 10: Tests for irritation and skin sensitization

## Clinical testing

The substantial equivalence determination for the HPM-6000U is not based on clinical testing. The device safety and efficacy was demonstrated by comparison of technical characteristics between the HPM-6000U and the predicate device.

## Comparison with the Predicate Device

<b>510(k) number</b>	<b>Not Assigned</b>	<b>K973096</b>
<b>Device name</b>	HPM-6000U	Neotonus Model 1000 Muscle Stimulator System
<b>Company name</b>	BTL Industries, Inc.	Neotonus, Inc.
<b>Product Code and Regulation</b>	<u>Gastroenterology-Urology Devices</u> 21 CFR 876.5320 KPI – Stimulator, Electrical, Non-Implantable, For Incontinence	<u>Gastroenterology-Urology Devices</u> 21 CFR 876.5320 KPI – Stimulator, Electrical, Non-Implantable, For Incontinence
<b>Intended Use</b>	HPM-6000U is intended to provide entirely non-invasive electromagnetic stimulation of pelvic floor musculature for the purpose of rehabilitation of weak pelvic muscles and restoration of neuromuscular control for the treatment of urinary incontinence in women.	Neotonus Model 1000 Muscle Stimulator System is intended to provide entirely non-invasive electromagnetic stimulation of pelvic floor musculature for the purpose of rehabilitation of weak pelvic muscles and restoration of neuromuscular control for the treatment of urinary incontinence in women.
<b>Primary Function</b>	Stimulation of pelvic floor musculature	Stimulation of pelvic floor musculature
<b>Principle of Action</b>	Initiating action potential of nerves results in muscle contraction	Initiating action potential of nerves results in muscle contraction
<b>Type of Energy</b>	Magnetic field	Magnetic field
<b>Energy Source</b>	100 – 240 V AC, 50 – 60 Hz, max 14 A	110 V AC, 50 – 60 Hz, max 12 A
<b>Type of Coil</b>	Single magnetic coil	Single magnetic coil
<b>Number of</b>	1	1

<b>510(k) number</b>	<b>Not Assigned</b>	<b>K973096</b>
<b>Device name</b>	HPM-6000U	Neotonus Model 1000 Muscle Stimulator System
<b>Company name</b>	BTL Industries, Inc.	Neotonus, Inc.
<b>Magnetic Coils in the Applicator</b>		
<b>Type of Applicator</b>	Chair	Chair
<b>Number of Applicators</b>	1	1
<b>Patient Position during Therapy</b>	Sitting position	Sitting position
<b>Position of Coil</b>	Center of applicator seat	Center of applicator seat
<b>Type of Operation</b>	Continuous	Continuous
<b>Magnetic Field Intensity</b>	0.7 – 2.5 T	Up to 2.2 T
<b>Pulse Repetition Rate</b>	1 – 150 Hz	1 – 55 Hz
<b>Step of Frequency Setting</b>	1 Hz	1 Hz
<b>Pulse Duration</b>	280 $\mu$ s ( $\pm$ 20%)	275 $\mu$ s
<b>Pulse Amplitude</b>	0 – 100 %	0 – 100 %
<b>Shape of Stimulation Pulse</b>	Sine, biphasic	Sine, biphasic
<b>Therapy Time</b>	30 min	30 min
<b>Operating Temperature</b>	+10 to +30 °C (50 – 86 °F)	+10 to +30 °C (50 – 86 °F)
<b>Interface</b>	Touch-screen	Graphical Display
<b>Firmware controlled</b>	Yes	Yes
<b>Environmental Specifications</b>	For indoor use only	For indoor use only
<b>Clinical Use</b>	Prescription Use	Prescription Use

<b>510(k) number</b>	<b>Not Assigned</b>	<b>K973096</b>
<b>Device name</b>	HPM-6000U	Neotonus Model 1000 Muscle Stimulator System
<b>Company name</b>	BTL Industries, Inc.	Neotonus, Inc.
<b>External Exchangeable Fuse</b>	Yes	Yes
<b>Main Unit Dimensions (W×H×D)</b>	500×970×580 mm (20×38×23 in)	580×170×320 mm (23×7×12.5 in)
<b>Applicator Dimensions (W×H×D)</b>	730×730×730 mm (29×29×29 in)	700×1250×785 mm (27.5×49×31 in)
<b>System Weight</b>	46 kg (101 lb)	42.5 kg
<b>Position</b>	Vertical – On castors	Horizontal
<b>Electrical Protection</b>	Class II, BF	Class II

### Substantial Equivalence

Based upon the intended use and known technical information provided in this pre-market notification, the HPM-6000U device has been shown to be substantially equivalent to currently marketed predicate device. Thorough Substantial Equivalence Discussion is provided in Section 12.

### Conclusion

Based on the aforementioned information, the HPM-6000U is safe and effective and substantially equivalent to the identified predicate device.