



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

September 29, 2016

GENICON, Inc.  
Ms. Katlyn Kachman  
Regulatory Compliance  
6869 Stapoint Court, Suite 114  
Winter Park, Florida 32792

Re: K162059  
Trade/Device Name: Ezee Retrieval  
Regulation Number: 21 CFR 876.1500  
Regulation Name: Endoscope and Accessories  
Regulatory Class: Class II  
Product Code: GCJ  
Dated: September 13, 2016  
Received: September 16, 2016

Dear Ms. Kachman:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR Part 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Christopher J. Ronk -S**

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.  
Director  
Division of Surgical Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure



# Section 1 Indications for Use

510(k) Number (if known) K162059

Device Name  
GENICON EZEE Retrieval

Indications for Use  
The GENICON EZEE Retrieval is indicated for use in laparoscopic procedures to capture organs or tissue to be removed from the body cavity.

Prescription Use X and/or Over-The-Counter Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

TEL 407.657.4851 — WWW.GENICONENDO.COM — FAX 407.677.9773  
US OFFICE • 6869 Stapoint Court #114, Winter Park, Florida USA 32792 — EU OFFICE • 44 Airedale Avenue, Chiswick, London U.K W4 2NW





## Section 2

### 510(k) Summary per 21 CFR 807.92

1. Contact Information

GENICON  
6869 Stapoint Court, Suite 114, Winter Park, FL 32792  
Phone (407) 657-4851 Fax (407) 677-9773  
Katlyn Kachman, Regulatory Compliance  
September 20, 2016

2. Device Name

GENICON EZEE Retrieval

3. Substantially Equivalent Device

Covidien Endo Catch Gold Specimen Pouch, 10mm, 173050G [K922123]  
Legally Marketed (unmodified) Device: GeniStrong Single Use Specimen Retrieval Bag, 10mm, 550-000-003 [K132375]

4. Description

The GENICON EZEE Retrieval is comprised of a flexible plastic bag with a large, easily accessible opening, an actuation rod with thumb ring handle, finger rings, string and closure suture, and an introducer shaft. In the fully deployed condition, the bag opening is maintained in a fully-open position by a metallic rim, and the size of the specimen bag is 4" x 5" with a volume of 230ml. A string with a closure suture facilitates closure of the specimen bag after the specimen had been collected. This device is disposable device packaged and sterilized for single use only. Do not re-use, reprocess, or re-sterilize. Discard after use.

5. Indications for Use

The GENICON EZEE Retrieval is indicated for use in laparoscopic procedures to capture organs or tissue to be removed from the body cavity.

6. Technical Specifications

The GENICON EZEE Retrieval is intended for Laparoscopic Surgery (GCJ) and contains a Bag, Biasing Arms, Introducer, Closure Suture/String, Handle and Actuation/Deployment Shaft. The shaft diameter is 10mm and is composed of PC while the Bag is Plastic and the Introducer is Stainless Steel. There are no FDA performance standards for these products. The sterilization is performed by Ethylene Oxide per ISO 11135:2014. This device is available by Prescription Only for use in a Hospital Operating Room. This device is compliant with FDA Class II requirements for ISO 10993.

Area	Genicon 510K Number: K132375	Covidien 510K Number: K922123	EZEE Retrieval 510K Number: K162059
Intended Use	Laparoscopic Surgery (GCJ)	Laparoscopic Surgery (GCI)	Laparoscopic Surgery (GCJ)
Design	Bag, Biasing Arms, Introducer, Closure Suture/String, Handle, Actuation/Deployment Shaft	SAME	SAME
Introducer	5 to 15 mm	10 to 15 mm	10 mm

TEL 407.657.4851 — WWW.GENICONENDO.COM — FAX 407.677.9773

US OFFICE • 6869 Stapoint Court #114, Winter Park, Florida USA 32792 — EU OFFICE • 44 Airedale Avenue, Chiswick, London U.K W4 2NW



Diameter			
Materials	Shaft – PC Bag – PU Coated Nylon(PA)	Shaft – Plastic and Stainless Steel Bag - TPU	Shaft – ABS Bag - TPU
Performance	There are no FDA performance standards for these products. Bench Testing and Clinical Evaluation performed. (ref. TR-12022-A and CLEV-032-A)	SAME	SAME
Sterilization	Ethylene Oxide per ISO 11135-1:2014	Ethylene Oxide	Ethylene Oxide per ISO 11135-1:2014
Where Used	Hospital Operating Room	SAME	SAME
Prescription Only	Yes	Yes	Yes
Biocompatibility	Compliant with FDA Class II requirements for ISO 10993	SAME	SAME

#### 7. Nonclinical Tests

The GENICON EZEE Retrieval has been evaluated by our Chief Technical Officer, Design Engineers, and Chief Medical Officer through performance studies and bench testing which included Deployment Force, Seam Strength, and Puncture Force.

#### 8. Clinical Tests

There were no clinical trials performed on the GENICON EZEE Retrieval.

#### 9. Conclusions

Based on the indications for use and technological characteristics, the GENICON EZEE Retrieval has shown to be substantially equivalent to the predicate device.

TEL 407.657.4851 — WWW.GENICONENDO.COM — FAX 407.677.9773

US OFFICE • 6869 Stapoint Court #114, Winter Park, Florida USA 32792 — EU OFFICE • 44 Airedale Avenue, Chiswick, London U.K W4 2NW