



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Toshiba Medical Systems Corporation
% Orlando Tadeo, Jr.
Manager, Regulatory Affairs
Toshiba America Medical Systems, Inc.
2441 Michelle Drive
TUSTIN CA 92780

November 1, 2016

Re: K162155
Trade/Device Name: Xario 200 Diagnostic Ultrasound System V5.0
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, ITX
Dated: October 13, 2016
Received: October 14, 2016

Dear Mr. Tadeo:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



For

Robert Ochs, Ph.D.
Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K162155

Device Name

Xario 200 Diagnostic Ultrasound System V5.0

Indications for Use (Describe)

The Diagnostic Ultrasound System Xario 200 Model TUS-X200 and Xario 200 Model TUS-X200S are indicated for the visualization of structures, and dynamic processes with the human body using ultrasound and to provide image information for diagnosis in the following clinical applications: fetal, abdominal, intra-operative(abdominal), laparoscopic, pediatric, small organs, neonatal cephalic, adult cephalic, trans-rectal, trans-vaginal, musculo-skeletal (conventional), musculo-skeletal (superficial), cardiac adult, cardiac pediatric, trans-esoph(cardiac) and peripheral vessel.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

System: Xario 200 TUS-X200,TUS-X200S V5.0

Transducer: _____

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal	P	P	P		P	2	P	P	P		P	4,5,6,7,8,9,11(E),12(N)
Abdominal	P	P	P	P	P	2,3	P	P	P		P	4,5,6,7,8,9,11(E),12(N)
Intra-operative (Abdominal)	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Neuro)												
Laparoscopic	P	P	P		P	2	P	P	P			5,7
Pediatric	P	P	P	P	P	2,3	P	P	P		P	4,5,6,7,8,9,11(E),12(N)
Small Organ (Note 1)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Neonatal Cephalic	P	P	P	P	P	2,3	P	P	P			5,6,7,8,11(E)
Adult Cephalic	P	P	P	P	P	3	P	P	P			7
Trans-rectal	P	P	P		P	2	P	P	P		P	4,5,6,7,8,9,12(N)
Trans-vaginal	P	P	P		P	2	P	P	P		P	4,5,6,7,8,9,12(N)
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Musculo-skeletal (Superficial)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Intravascular												
Other (Specify)												
Cardiac Adult	P	P	P	P	P	3	P	P	P	P		4, 7
Cardiac Pediatric	P	P	P	P	P	3	P	P	P	P		4, 7
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)	P	P	P	P	P	3	P					4, 7
Intra-cardiac												
Other (Specify)												
Peripheral vessel	P	P	P	P	P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PSU-25BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	P	P	P	P	P	3	P	P	P			7
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P	P	P	3	P	P	P			7
Small Organ (Note 1)												
Neonatal Cephalic	P	P	P	P	P	3	P	P	P			7
Adult Cephalic	P	P	P	P	P	3	P	P	P			7
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult	P	P	P	P	P	3	P	P	P	P		4, 7
Cardiac Pediatric	P	P	P	P	P	3	P	P	P	P		4, 7
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PSU-30BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	P	P	P	P	P	3	P	P	P			7
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P	P	P	3	P	P	P			7
Small Organ (Note 1)												
Neonatal Cephalic	P	P	P	P	P	3	P	P	P			7
Adult Cephalic	P	P	P	P	P	3	P	P	P			7
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult	P	P	P	P	P	3	P	P	P	P		4, 7
Cardiac Pediatric	P	P	P	P	P	3	P	P	P	P		4, 7
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PSU-50BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	P	P	P	P	P	3	P	P	P			7
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P	P	P	3	P	P	P			7
Small Organ (Note 1)												
Neonatal Cephalic	P	P	P	P	P	3	P	P	P			7
Adult Cephalic	P	P	P	P	P	3	P	P	P			7
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult	P	P	P	P	P	3	P	P	P	P		4, 7
Cardiac Pediatric	P	P	P	P	P	3	P	P	P	P		4, 7
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PSU-70BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	E	E	E	E	E	3	E	E	E			7
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	E	E	E	E	E	3	E	E	E			7
Small Organ (Note 1)												
Neonatal Cephalic	E	E	E	E	E	3	E	E	E			7
Adult Cephalic	E	E	E	E	E	3	E	E	E			7
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult	E	E	E	E	E	3	E	E	E	E		4, 7
Cardiac Pediatric	E	E	E	E	E	3	E	E	E	E		4, 7
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-375BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal	P	P	P		P	2	P	P	P			4,5,6,7,8,9,11(E),12(N)
Abdominal	P	P	P		P	2	P	P	P			4,5,6,7,8,9,11(E),12(N)
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P		P	2	P	P	P			4,5,6,7,8,9,11(E),12(N)
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-382BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Abdominal	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-674BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal	P	P	P		P	2	P	P	P			5,6,7,8,11(E),12(N)
Abdominal	P	P	P		P	2	P	P	P			5,6,7,8,11(E),12(N)
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P		P	2	P	P	P			5,6,7,8,11(E),12(N)
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix
 Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-674MV

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal	P	P	P		P	2	P	P	P		P	5,6,7,8
Abdominal	P	P	P		P	2	P	P	P		P	5,6,7,8
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P		P	2	P	P	P		P	5,6,7,8
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-712BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI 2D	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Small Organ (Note 1)												
Neonatal Cephalic	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-745BTF

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Abdominal)	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-745BTH

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Abdominal)	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-745BTV

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Abdominal)	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)	P	P	P		P	2	P	P	P			5,6,7,8,11(E)
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-681MVL

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal	P	P	P		P	2	P	P	P		P	5,6,7,8,12(N)
Trans-vaginal	P	P	P		P	2	P	P	P		P	5,6,7,8,12(N)
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-770ST

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Specific (Tracks 3)	Mode of Operation											
	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal	P	P	P		P	2	P	P	P			4,5,6,7,8,9
Trans-vaginal	P	P	P		P	2	P	P	P			4,5,6,7,8,9
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-781VT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal	P	P	P		P	2	P	P	P			4,5,6,7,8,9,12(N)
Trans-vaginal	P	P	P		P	2	P	P	P			4,5,6,7,8,9,12(N)
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix
 Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVU-781VTE

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal	E	E	E		E	2	E	E	E			4,5,6,7,8,9,12(N)
Trans-vaginal	E	E	E		E	2	E	E	E			4,5,6,7,8,9,12(N)
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix
 Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PVL-715RS

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal	P	P	P		P	2	P	P	P			5,6,7,8,12(N)
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PLU-704BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)	P	P	P		P	2	P	P	P			5,6,7,8,10,11(E),12(N)
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)	P	P	P		P	2	P	P	P			5,6,7,8,10,11(E),12(N)
Musculo-skeletal (Superficial)	P	P	P		P	2	P	P	P			5,6,7,8,10,11(E),12(N)
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel	P	P	P		P	2	P	P	P			5,6,7,8,10,11(E),12(N)
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix
 Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PLU-1005BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Musculo-skeletal (Superficial)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix
 Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PLU-1204BT

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Musculo-skeletal (Superficial)	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel	P	P	P		P	2	P	P	P			4,5,6,7,8,9,10,11(E),12(N)
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix
 Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PET-512MC

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)	P	P	P	P	P	3	P					4, 7
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PET-512MD

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult												
Cardiac Pediatric												
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)	N	N	N	N	N	3	N					4, 7
Intra-cardiac												
Other (Specify)												
Peripheral vessel												
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: N/A

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PET-805LA

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation												
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Dynamic Flow	Power	CHI 2D	4D	Other	[Note]
Ophthalmic													
Fetal													
Abdominal													
Intra-operative (Abdominal)													
Intra-operative (Neuro)													
Laparoscopic	P	P	P		P	2	P	P	P			5,7	
Pediatric													
Small Organ (Note 1)													
Neonatal Cephalic													
Adult Cephalic													
Trans-rectal													
Trans-vaginal													
Trans-urethral													
Trans-esoph. (non-Card.)													
Musculo-skeletal (Conventional)													
Musculo-skeletal (Superficial)													
Intravascular													
Other (Specify)													
Cardiac Adult													
Cardiac Pediatric													
Intravascular (Cardiac)													
Trans-esoph. (Cardiac)													
Intra-cardiac													
Other (Specify)													
Peripheral vessel													
Other (Specify)													

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PC-20M

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult				P								
Cardiac Pediatric				P								
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel				P								
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

System: Xario 200 TUS-X200, TUS-X200S V5.0Transducer: PC-50M

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation											
Specific (Tracks 3)	B	M	PWD	CWD	Color Doppler	Combined (Specify) *	THI	Advanced Dynamic Flow	Power	CHI	4D	Other [Note]
Ophthalmic												
Fetal												
Abdominal												
Intra-operative (Abdominal)												
Intra-operative (Neuro)												
Laparoscopic												
Pediatric												
Small Organ (Note 1)												
Neonatal Cephalic												
Adult Cephalic												
Trans-rectal												
Trans-vaginal												
Trans-urethral												
Trans-esoph. (non-Card.)												
Musculo-skeletal (Conventional)												
Musculo-skeletal (Superficial)												
Intravascular												
Other (Specify)												
Cardiac Adult				P								
Cardiac Pediatric				P								
Intravascular (Cardiac)												
Trans-esoph. (Cardiac)												
Intra-cardiac												
Other (Specify)												
Peripheral vessel				P								
Other (Specify)												

N = new indication; P = previously cleared by FDA; E = added under this appendix

Previous 510(k) of the transducers: K143027

Note 1 Small organ includes thyroid, breast and testicle.

Note 2 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD

Note 3 Combined mode includes B/M; B/PWD; BDF/PWD; BDF/MDF; BDF/MDF/PWD; 2D/CWD; BDF/CWD

Note 4 TDI

Note 5 ApliPure

Note 6 ApliPure Plus

Note 7 Precision Imaging

Note 8 Differential THI

Note 9 Elastography

Note 10 BEAM

Note 11 Smart 3D

Note 12 SMI

Prescription Use Only (Per 21 CFR 801.109)

510(k) SUMMARY**1. SUBMITTER'S NAME:**

Toshiba Medical Systems Corporation
1385 Shimoishigami
Otawara-shi, Tochigi-ken, Japan 324-8550

2. OFFICIAL CORRESPONDENT

Akinori Hatanaka

3. ESTABLISHMENT REGISTRATION:

9614698

4. CONTACT PERSON:

Orlando Tadeo, Jr.
Manager, Regulatory Affairs
Toshiba America Medical Systems, Inc
2441 Michelle Drive
Tustin, CA 92780
(714) 669-7459

5. Date Prepared:

October 13, 2016

6. TRADE NAME(S):

Xario 200 Diagnostic Ultrasound System V5.0

7. COMMON NAME:

System, Diagnostic Ultrasound

8. DEVICE CLASSIFICATION:

Class II

Ultrasonic Pulsed Doppler Imaging System – Product Code: 90-IYN [per 21 CFR 892.1550]

Ultrasonic Pulsed Echo Imaging System – Product Code: 90-IYO [per 21 CFR 892.1560]

Diagnostic Ultrasonic Transducer – Product Code: 90-ITX [per 21 CFR 892.1570]

9. PREDICATE DEVICE:

Product	Marketed by	510(k) Number	Clearance Date
Xario200, V3.0 (Models: TUS-X200 and TUS-X200S)	Toshiba America Medical Systems	K143027	November 20, 2014

Reference Device

Product	Marketed by	510(k) Number	Clearance Date
Aplio 500/400/300 Diagnostic Ultrasound System, V6.0	Toshiba America Medical Systems	K151451	July 9, 2015

10. REASON FOR SUBMISSION:

Modification of a cleared device

11. DEVICE DESCRIPTION:

The Xario200 Model TUS-X200 and Model TUS-X200S are mobile diagnostic ultrasound systems. These systems are Track 3 devices that employ a wide array of probes including flat linear array, convex linear array, and sector array with frequency ranges between approximately 2 MHz to 12 MHz.

12. INDICATIONS FOR USE:

The Diagnostic Ultrasound System Xario 200 Model TUS-X200 and Xario 200 Model TUS-X200S are indicated for the visualization of structures, and dynamic processes with the human body using ultrasound and to provide image information for diagnosis in the following clinical applications: fetal, abdominal, intra-operative(abdominal), laparoscopic, pediatric, small organs, neonatal cephalic, adult cephalic, trans-rectal, trans-vaginal, musculo-skeletal (conventional), musculo-skeletal(superficial), cardiac adult, cardiac pediatric, trans-esoph(cardiac) and peripheral vessel.

13. SUBSTANTIAL EQUIVALENCE:

This device is substantially equivalent to the Xario200 V3.0 Diagnostic Ultrasound System, 510(k) cleared under K143027, marketed by Toshiba America Medical Systems. The **Xario 200 Model TUS-X200 Version 5.0 and Xario200 Model TUS-X200S Version 5.0** functions in a manner similar to and is intended for the same use as the predicate devices referenced within this submission. The subject device includes Precision Plus Imaging, a modification to an existing feature which improves upon image quality and does not raise new questions of safety and effectiveness. Also, new features determined to be substantially equivalent to features cleared under the predicate devices referenced within this submission including SMI, Luminance and a new transducer (which is indicated for transesophageal use) are being implemented to the subject device.

	Xario 200 Model TUS-X200, TUS-X200S v3.0	Xario 200 Model TUS-X200 v5.0	Xario 200 Model TUS-X200S v5.0	Comment
510(K) Control Number	K143027 (Predicate Device)	N/A (Subject Device)	N/A (Subject Device)	
Precision Plus Imaging	No	Yes	No	Image quality improvement(s) to existing feature
Monitor	19 inch LCD (CCFL backlight)	19 inch LCD (LED backlight)	19 inch LCD (LED backlight)	New hardware
PET-512MD Transducer	Not Available	Available	Available	New TEE transducer, similar to PET-512MC (K103629)

Previously cleared software options being implemented to the subject device:

Application	510(k) Clearance	Comments
Superb Micro Vascular Imaging	Previously cleared under K151451	Available on Model TUS- X200 only
Luminance	Previously cleared under K151451	N/A

14. SAFETY:

The device is designed and manufactured under the Quality System Regulations as outlined in 21 CFR § 820 and ISO 13485 Standards. This device is in conformance with the applicable parts of the IEC60601-1 (2005), IEC 60601-1-2:2007, IEC 60601-2-37 (2007), IEC 62304 (2006), AIUM RTD2-2004 Output Display and ISO 10993-1 standards.

15. TESTING

Design Control Activities including risk management following ISO14971, verification/validation testing and Acoustic Output testing (UD3, 2004) were conducted and included in this submission.

Performance Testing – Clinical Images

Representative clinical images of volunteers were obtained to demonstrate that the implementation of SMI onto the subject device performed as expected. As concluded in the study, SMI was capable of imaging low velocity flow with a significant reduction in clutter noise and was capable of imaging with a high frame rate.

Software Documentation for a Moderate Level of Concern, per the FDA guidance document, “Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices Document” issued on May 11, 2005, is also included as part of this submission.

Additionally, testing of the modified system was conducted in accordance with the

applicable standards published by the International Electrotechnical Commission (IEC) for Medical Devices.

16. CONCLUSION

The modifications incorporated into the **Xario 200 Model TUS-X200 Version 5.0** and **Xario200 Model TUS-X200S Version 5.0** do not change the intended use of the device.

Based upon the acquisition of representative clinical images, successful completion of software validation, application of risk management and design controls, it is concluded that this device is safe and effective for its intended use.