



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

July 21, 2017

Suzhou Kangjie Medical Inc.  
% Di Wu, Ph.D.  
Regulatory Affairs Consultant  
DIWU Regulatory  
3440 Indian Queen Lane, Rear  
Philadelphia, PA 19129

Re: K162778  
Trade/Device Name: Portable Slit Lamp Microscope  
Regulation Number: 21 CFR 886.1850  
Regulation Name: AC-Powered Slitlamp Biomicroscope  
Regulatory Class: Class II  
Product Code: HJO  
Dated: May 25, 2017  
Received: May 30, 2017

Dear Dr. Di Wu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification"

(21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

  
*Kesia Alexander*

for Malvina B. Eydelman, M.D.  
Director  
Division of Ophthalmic and Ear,  
Nose and Throat Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)

K162778

Device Name

Portable Slit Lamp Microscope Model KJ5S

Indications for Use (Describe)

Portable Slit Lamp Microscope Model KJ5S is intended for use in the examination of the anterior and posterior segment of the eye.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**Date prepared** July 13, 2017

**Submission Sponsor (Manufacturer):**

Suzhou Kangjie Medical Inc.  
No.129, Weixin Rd., Weiting Town  
Suzhou, Jiangsu, P.R. China 215000

**Submission Correspondent (Agent):**

Mrs. Di Wu, PhD  
3440 Indian Queen Lane, Rear  
Philadelphia, PA 19129 USA  
Phone: (386) 487-8384  
Email: diwuregulatory@gmail.com

**Trade/Device Name:**

Portable Slit Lamp Microscope  
Model KJ5S

**Common or Usual Name:** Slit Lamp Microscope

**Device Class:** II

**Classification Name:** AC-Powered Slitlamp Biomicroscope

**Regulation Number:** 21 CFR 886.1850

**Product Code:** HJO

**Review Panel:** Ophthalmic

**Predicate Device:**

- K131711, SUZHOU 66 VISION TECH CO., LTD  
YZ3 PORTABLE SLIT LAMP

**Device Description:**

Portable Slit Lamp Microscope  
Model KJ5S

The Portable Slit Lamp Microscope is a hand held converging stereomicroscope system powered by rechargeable batteries.

---

The illumination system and fixation targets are activated using a single/double click trigger located on the rear of the grip/handle. A rheostat located below the eyepieces on the front of the grip /handle is used to increase or reduce the light intensity.

Environment of Use: healthcare facility/hospital

Duration and type of contact: less than 24 hours, surface device

**Intended Use:**

Portable Slit Lamp Microscope Model KJ5S is intended for use in the examination of the anterior and posterior segment of the eye.

**Comparison to Predicate Device:**

Refer to the Comparison Table on Pages 4 and 5.

**Discussion of Non-Clinical Tests Performed:**

The performance tests were conducted:

IEC 60601-1, Medical electrical equipment - Part 1: General requirements for basic safety and essential performance

IEC 60601-1-2 Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests

ISO 15004-1, Ophthalmic instruments - Fundamental requirements and test methods - Part 1: General requirements applicable to all ophthalmic instruments

ISO 15004-2, Ophthalmic instruments - Fundamental requirements and test methods - Part 2: Light hazard protection

ISO 10939, Ophthalmic instruments — Slit-lamp microscopes

The biocompatibility evaluation for the subject device was conducted in accordance with the FDA Blue Book Memorandum #G95-1 "Use of International

---

Standard ISO-10993, 'Biological Evaluation of Medical Devices Part 1: Evaluation and Testing,'" May 1, 1995, and International Standard ISO 10993-1 "Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing Within a Risk Management Process," as recognized by FDA.

ISO 10993-1, Biological Evaluation of Medical Devices -- Part 1: Evaluation And Testing Within A Risk Management Process. (Biocompatibility)

ISO 10993-5, Biological Evaluation of Medical Devices -- Part 5: Tests For In Vitro Cytotoxicity. (Biocompatibility)

ISO 10993-10, Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization. (Biocompatibility)

IEC 62133, Secondary cells and batteries containing alkaline or other non-acid electrolytes - Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications

**Discussion of Clinical Tests Performed:**

None

**Conclusion:**

The subject device has the same intended uses and indications, similar technological characteristics, and principles of operation as its predicate device. The minor differences between the subject device and its predicate devices raise no new issues of safety or effectiveness. Thus, the subject devices are substantially equivalent to its predicate devices.

## Comparison Table

### KJ5S Slit Lamp Microscope vs. K131711

Descriptive Information	Proposed Device	Predicate Device
<b>510(k) Number</b>	<b>K162778</b>	<b>K131711</b>
<b>Manufacturer</b>	Suzhou Kangjie Medical Inc.	66 Vision Tech Co., Ltd.
<b>Proprietary or Model Name</b>	KJ5S portable slit lamp microscope	YZ3 Portable Slit Lamp microscope
<b>Indications for Use</b>	Same	Same
<b>Flammability of materials near the light source</b>	NONE	NONE
<b>Maximum temperature of parts of the device held by the operator or accessible to the patient</b>	Maximum temperature of parts of the device held by the operator: - Eyepiece: 35° - Grip: 35 ° - Slit width control ring: 35° Maximum temperature of parts of the device accessible to the patient: Forehead rest 35°	Same
<b>Brightness controls</b>	Maximal Illumination ≥12000Lx	Maximal Illumination ≥30000Lx
<b>Operating temperature</b>	10° to 35°	Same
<b>Slit Width</b>	0 to 10mm continuously adjustable	0 to 12mm continuously adjustable
<b>Slit Length</b>	1mm, 3mm, 5mm, 10mm	0.2mm, 1mm, 2mm, 12mm
<b>Illumination field diameter</b>	1mm, 3mm, 5mm, 10mm	0.2mm, 1mm, 2mm, 12mm
<b>Radial movement of the slit light illumination relative to the microscope axis</b>	Horizontal ± 30°	Same
<b>Stereo angle</b>	13°	Same
<b>Light sources</b>	White LED Single Light Source	Same
<b>Pupil-distance</b>	48 to 72mm	50 to 75mm
<b>Eye piece</b>	10X, 16X	Same
<b>Objective</b>	1X	Same
<b>Total magnifications</b>	10X, 16X(Optional)	Same
<b>Filter</b>	Heat-absorption, Cobalt blue, Red-free, Gray	Cobalt blue, Red-free, Color Temperature Compensation

## Comparison Table

---

<b>Descriptive Information</b>	<b>Proposed Device</b>	<b>Predicate Device</b>
<b>510(k) Number</b>	<b>K162778</b>	<b>K131711</b>
<b>Illumination rotation angle</b>	Horizontal $\pm$ 30°	Same
<b>Working distance</b>	80mm	60mm
<b>Power</b>	7.4V 680mAh Li Battery, Rechargeable	7.4V 2200mA Li Battery, Rechargeable
<b>Working time</b>	5-6 hours	2.5 hours
<b>Net Weight</b>	750g	900g