



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

February 8, 2017

Lucenxia Prescience Ag  
Robert Hill  
Regulatory Director  
Rathausstrasse 7  
Baar, 6341 CH

Re: K162852

Trade/Device Name: Finessis Zero Chemo Flexylon Powder Free Sterile White Surgical Glove

Regulation Number: 21 CFR 878.4460

Regulation Name: Surgeon's Glove

Regulatory Class: Class I

Product Code: KGO, LZC

Dated: January 11, 2017

Received: January 11, 2017

Dear Robert Hill:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

  
**Michael J. Ryan -S**

for Tina Kiang, Ph.D.

Acting Division Director

Division of Anesthesiology,

General Hospital, Respiratory,

Infection Control, and Dental Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

### Indications for Use

510(k) Number (if known)  
K162852

Device Name  
FINESSIS ZERO CHEMO Flexylon Powder Free Sterile White Surgical Gloves

Indications for Use (Describe)

This surgeon's glove is a device made of Flexylon intended to be worn by operating room personnel to protect a surgical wound from contamination.

Flexylon; a styrene-ethylene/butylene-styrene (SEBS) material.

These Gloves are tested for use with Chemotherapy Drugs  
DO NOT USE with Carmustine or Thiotepa  
Chemical Resistance Data per ASTM D 6978-05

Drug	Minimum Breakthrough Detection Time (minutes)
Carmustine (BCNU) 3.3 mg/ml (3,300 ppm)	15.3
Cisplatin 1.0 mg/ml (1,000 ppm)	240
Cyclophosphamide (cytoxan) 20 mg/ml (20,000 ppm)	240
Dacarbazine (DTIC) 10.0 mg/ml (10,000)	240
Doxorubicin Hydrochloride 2.0 mg/ml (2,000 ppm)	240
Etoposide (Toposar) 20.0 mg/ml (20,000 ppm)	240
Fluorouracil 50.0 mg/ml (50,000 ppm)	240
Ifosfamide 50.0 mg/ml (50,000 ppm)	240
Methotrexate 25 mg/ml (25,000 ppm)	240
Mitomycin C 0.5 mg/ml (500 ppm)	240
Paclitaxel (Taxol) 6.0 mg/ml (6,000 ppm)	240
Thiotepa 10.0 mg/ml (10,000 ppm)	15.8
Vincristine Sulfate 1.0 mg/ml (1,000 ppm)	240

Type of Use (Select one or both, as applicable)

- Prescription Use (Part 21 CFR 801 Subpart D)       Over-The-Counter Use (21 CFR 801 Subpart C)

#### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*