



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

December 06, 2016

Medtronic Cardiac and Vascular Group
Renee Cveykus
Principal Regulatory Affairs Specialist
8200 Coral Sea St. NE
Mounds View, MN 55112

Re: K162896

Trade/Device Name: Affinity NT Oxygenator (Model 511), Affinity NT Oxygenator with
Trillium Biosurface (Model 511T)

Regulation Number: 21 CFR 870.4350

Regulation Name: Cardiopulmonary Bypass Oxygenator

Regulatory Class: Class II

Product Code: DTZ, DTR

Dated: October 14, 2016

Received: October 17, 2016

Dear Ms. Cveykus:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

A handwritten signature in black ink, appearing to read "Bram D. Zuckerman", is written over a large, light blue "FDA" watermark.

for

Bram D. Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K162896

Device Name

Affinity NT Oxygenator

Indications for Use (Describe)

Model 511:

The Affinity NT Oxygenator is intended to be used in an extracorporeal perfusion circuit to oxygenate and remove carbon dioxide from the blood and to cool or warm the blood during routine cardiopulmonary bypass (CPB) procedures up to 6 hours in duration.

Model 511T:

The Affinity NT Oxygenator with Trillium Biosurface is intended to be used in an extracorporeal perfusion circuit to oxygenate and remove carbon dioxide from the blood and to cool or warm the blood during routine cardiopulmonary bypass (CPB) procedures up to 6 hours in duration.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

5.1 510(k) Summary

Date Prepared: October 14, 2016

Submitter: Medtronic, Inc.
Medtronic Perfusion Systems
7611 Northland Drive
Minneapolis, MN 55428
Establishment Registration Number: 2184009

Contact Person: Renee L. Cveykus
Principal Regulatory Affairs Specialist
Medtronic Perfusion Systems
Phone: (763) 505-3059
Fax: (763) 367-0401
Email: renee.l.cveykus@medtronic.com

Alternate Contact: Susan C. Fidler
Sr. Regulatory Affairs Manager
Medtronic Perfusion Systems
Phone: (763) 514-9839
Fax: (763) 367-8360
Email: susan.c.fidler@medtronic.com

Proprietary Name:

Models	Description
511	Affinity NT™ Oxygenator
511T	Affinity NT™ Oxygenator with Trillium™ Biosurface

Device Name and Classification:

Trade Name:	Affinity NT™ Oxygenator with Trillium™ Biosurface
Common Name:	Oxygenator
Classification Name:	Cardiopulmonary bypass Oxygenator
Classification Panel:	Cardiovascular
Regulation Number:	21 CFR 870.4350
Product Code:	DTZ
Classification:	Class II

Predicate Device:

Medtronic Affinity NT Oxygenators (K143073). The scope of this predicate submission is below and included both uncoated and Trillium coated models.

Product Family	Model	Description
Affinity NT Oxygenator	511	Affinity NT Oxygenator
	511T	Affinity NT Oxygenator with Trillium Biosurface

Device Description

The device listed in this 510(k) Notification is single use, non-toxic, non-pyrogenic, and is supplied sterile in packaging.

The Medtronic Affinity NT Oxygenator is a single use device designed to oxygenate and remove carbon dioxide from the blood and with the heat exchanger and arterial filter cools or warms the blood during extracorporeal circulation. The oxygenator has either a blood contacting Trillium™ Biosurface or is uncoated.

The Affinity NT Oxygenator is intended to be used in an extracorporeal perfusion circuit to oxygenate and remove carbon dioxide from the blood and to cool or warm the blood during routine cardiopulmonary bypass procedures up to 6 hours in duration.

Affinity NT Oxygenator is designed to be an integral part of the cardiopulmonary heart lung bypass circuit for use during cardiac surgery. Blood that comes from the patient is delivered through a blood pump to the oxygenator and other auxiliary devices, and then back to the patient.

The purpose of this 510(k) Notification is to notify the FDA of a labeling change to the disinfectant warning on the Instructions for Use to allow for disinfectant use in the water path of the oxygenator as well as report previous changes for the Affinity NT Oxygenator models included in this submission.

Indications for Use

There is no change to the intended use of the devices within the scope of the proposed change in this Traditional 510(k) Notification. The current Indications for Use statements for these devices are listed below:

The Affinity NT Oxygenator (Model 511) is intended to be used in an extracorporeal perfusion circuit to oxygenate and remove carbon dioxide from the blood and to cool or warm the blood during routine cardiopulmonary bypass (CPB) procedures up to 6 hours in duration.

The Affinity NT Oxygenator with Trillium Biosurface (model 511T) is intended to be used in an extracorporeal perfusion circuit to oxygenate and remove carbon dioxide from the blood and to cool or warm the blood during routine cardiopulmonary bypass (CPB) procedures up to 6 hours in duration.

Comparison to Predicate Devices

The Affinity NT Oxygenator has the same intended use, design and materials, and principles of operation and technology when compared to the predicate Affinity NT Oxygenator.

- Intended Use: The intended use is the same as the predicate device.
- Design: The design is the same as the predicate device.
- Materials: The materials of the Affinity NT Oxygenator are the same as the predicate device.
- Principles of Operation and Technology: The principles of operation are the same as the predicate device.
- Performance: The performance of the device is the same as the predicate device.

Summary of Performance Data

Testing was used to verify the performance characteristics of this device. Clinical testing was not required to establish substantial equivalence. The following performance tests were conducted:

Testing	Description	Result
Pressure Integrity	Water path must withstand 45 PSI pressure for 6 hours without leaking	Pass
Burst	Water path burst testing should be comparable to that of the control devices	Pass
Port Break	Water path break force shall be comparable to that of the control device	Pass

An analysis was also completed to characterize the physical properties of the materials used to construct the water side of the heat exchanger.

Conclusion

In summary, the information provided within the submission demonstrates that the Affinity NT Oxygenator Uncoated or with Trillium Biosurface is substantially equivalent to the marketed predicate devices.