



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

July 22, 2017

BIOMERIEUX, INC.
NATHAN HARDESTY
MANAGER, REGULATORY AFFAIRS MICROBIOLOGY
595 ANGLUM RD.
HAZELWOOD MO 63042

Re: K162950

Trade/Device Name: VITEK MS

Regulation Number: 21 CFR 866.3361

Regulation Name: Mass spectrometer system for clinical use for the identification of
microorganisms

Regulatory Class: II

Product Code: PEX

Dated: June 16, 2017

Received: June 19, 2017

Dear Mr. Hardesty:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulations (21 CFR Parts 801 and 809), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Kristian M. Roth -S

For: Uwe Scherf, M.Sc., Ph.D.
Director
Division of Microbiology Devices
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K162950

Device Name
VITEK® MS

Indications for Use (Describe)

VITEK® MS is a mass spectrometry system using matrix-assisted laser desorption/ionization – time of flight mass spectrometry (MALDI-TOF MS) for the identification of microorganisms cultured from human specimens.

The VITEK® MS is a qualitative in vitro diagnostic device indicated for use in conjunction with other clinical and laboratory findings to aid in the diagnosis of bacterial, yeast and mould infections.

(See Attached for 'List of Claimed Organisms')

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Indications for Use Attachment

<i>Abiotrophia defectiva</i>	<i>Candida parapsilosis</i>
<i>Achromobacter denitrificans</i> ¹	<i>Candida pelliculosa</i>
<i>Achromobacter xylosoxidans</i> ¹	<i>Candida rugosa</i>
<i>Acinetobacter baumannii</i> complex	<i>Candida tropicalis</i>
<i>Acinetobacter haemolyticus</i>	<i>Candida utilis</i>
<i>Acinetobacter junii</i>	<i>Candida zeylanoides</i>
<i>Acinetobacter lwoffii</i>	<i>Chryseobacterium indologenes</i>
<i>Actinomyces meyeri</i>	<i>Citrobacter amalonaticus</i>
<i>Actinomyces neuii</i>	<i>Citrobacter braaki</i> ⁵
<i>Actinomyces odontolyticus</i>	<i>Citrobacter freundii</i> ⁵
<i>Aerococcus viridans</i>	<i>Citrobacter koseri</i>
<i>Aeromonas hydrophila/caviae</i> ²	<i>Citrobacter youngae</i> ⁵
<i>Aeromonas sobria</i> ^{2 3}	<i>Clostridium clostridioforme</i>
<i>Aggregatibacter actinomycetemcomitans</i>	<i>Clostridium difficile</i>
<i>Aggregatibacter aphrophilus</i>	<i>Clostridium perfringens</i>
<i>Aggregatibacter segnis</i>	<i>Clostridium ramosum</i>
<i>Alcaligenes faecalis ssp faecalis</i>	<i>Corynebacterium jeikeium</i>
<i>Bacteroides caccae</i>	<i>Cronobacter sakazakii</i>
<i>Bacteroides fragilis</i>	<i>Cryptococcus neoformans</i>
<i>Bacteroides ovatus</i> ⁴	<i>Edwardsiella hoshinae</i>
<i>Bacteroides thetaiotaomicron</i>	<i>Edwardsiella tarda</i>
<i>Bacteroides uniformis</i>	<i>Eikenella corrodens</i>
<i>Bacteroides vulgatus</i>	<i>Elizabethkingia meningoseptica</i>
<i>Bordetella parapertussis</i>	<i>Enterobacter aerogenes</i>
<i>Bordetella pertussis</i>	<i>Enterobacter asburiae</i> ⁶
<i>Brevundimonas diminuta</i>	<i>Enterobacter cancerogenus</i>
<i>Burkholderia multivorans</i>	<i>Enterobacter cloacae</i> ⁶
<i>Campylobacter coli</i>	<i>Enterococcus avium</i>
<i>Campylobacter jejuni</i>	<i>Enterococcus casseliflavus</i>
<i>Candida albicans</i>	<i>Enterococcus durans</i>
<i>Candida dubliniensis</i>	<i>Enterococcus faecalis</i>
<i>Candida famata</i>	<i>Enterococcus faecium</i>
<i>Candida glabrata</i>	<i>Enterococcus gallinarum</i>
<i>Candida guilliermondii</i>	<i>Escherichia coli</i> ⁷
<i>Candida haemulonii</i>	<i>Escherichia fergusonii</i>
<i>Candida inconspicua</i>	<i>Escherichia hermannii</i>
<i>Candida intermedia</i>	<i>Ewingella americana</i>
<i>Candida kefyr</i>	<i>Fingoldia magna</i>
<i>Candida krusei</i>	<i>Fusobacterium necrophorum</i>
<i>Candida lambica</i>	<i>Fusobacterium nucleatum</i>
<i>Candida lipolytica</i>	<i>Gardnerella vaginalis</i>
<i>Candida lusitanae</i>	<i>Gemella haemolysans</i>
<i>Candida norvegensis</i>	<i>Gemella morbillorum</i>

Geotrichum capitatum
Granulicatella adiacens
Haemophilus influenzae
Haemophilus parahaemolyticus
Haemophilus parainfluenzae
Hafnia alvei
Kingella denitrificans
Kingella kingae
Klebsiella oxytoca
Klebsiella pneumoniae
Kodamaea ohmeri
Lactococcus garvieae
*Lactococcus lactis ssp lactis*⁸
Leclercia adecarboxylata
Legionella pneumophila
Leuconostoc mesenteroides
Leuconostoc pseudomesenteroides
Listeria monocytogenes
Malassezia furfur
Malassezia pachydermatis
Micrococcus luteus
Mobiluncus curtisii
Moraxella (Neisseria) ovis
Morganella morganii
Neisseria cinerea
*Neisseria gonorrhoeae*⁹
Neisseria meningitidis
*Neisseria mucosa*¹⁰
Ochrobactrum anthropi
Oligella ureolytica
Oligella urethralis
Pantoea agglomerans
Parvimonas micra
Pasteurella multocida
Pediococcus acidilactici
Peptoniphilus asaccharolyticus
Peptostreptococcus anaerobius
Pluralibacter gergoviae
Prevotella bivia
Prevotella buccae
Prevotella denticola
Prevotella intermedia
Prevotella melaninogenica
Propionibacterium acnes

Proteus mirabilis
*Proteus penneri*¹¹
*Proteus vulgaris*¹¹
Providencia rettgeri
Providencia stuartii
Pseudomonas aeruginosa
Pseudomonas fluorescens
Pseudomonas putida
Pseudomonas stutzeri
Ralstonia pickettii
Raoultella ornithinolytica
Raoultella planticola
Rhizobium radiobacter
Rhodotorula mucilaginosa
Rothia mucilaginosa
Saccharomyces cerevisiae
*Salmonella group*⁹
Saprochaete capitata
Serratia fonticola
Serratia liquefaciens
Serratia marcescens
Serratia odorifera
Sphingobacterium multivorum
Sphingobacterium spiritivorum
Sphingomonas paucimobilis
Staphylococcus aureus
Staphylococcus capitis
Staphylococcus cohnii ssp cohnii
Staphylococcus cohnii ssp urealyticus
Staphylococcus epidermidis
Staphylococcus haemolyticus
*Staphylococcus hominis ssp hominis*¹²
Staphylococcus lugdunensis
Staphylococcus saprophyticus
Staphylococcus schleiferi
Staphylococcus sciuri
Staphylococcus simulans
Staphylococcus warneri
Stenotrophomonas maltophilia
Streptococcus agalactiae
Streptococcus anginosus
Streptococcus constellatus
Streptococcus dysgalactiae
Streptococcus gallolyticus ssp gallolyticus

Streptococcus infantarius ssp coli
Streptococcus infantarius ssp infantarius
Streptococcus intermedius
Streptococcus mitis/Streptococcus oralis
Streptococcus mutans
Streptococcus pneumoniae
Streptococcus pyogenes
*Streptococcus salivarius ssp salivarius*¹³
Streptococcus sanguinis
Trichosporon asahii
Trichosporon inkin
Trichosporon mucoides
Vibrio cholerae
Vibrio parahaemolyticus
Vibrio vulnificus
Yersinia enterocolitica
Yersinia frederiksenii
Yersinia intermedia
Yersinia kristensenii
*Yersinia pseudotuberculosis*¹⁴

MYCOBACTERIUM

Mycobacterium abscessus
Mycobacterium avium
Mycobacterium chelonae
*Mycobacterium fortuitum group:*¹⁵
 Mycobacterium alvei
 Mycobacterium farcinogenes
 Mycobacterium fortuitum
 Mycobacterium houstonense
 Mycobacterium peregrinum
 Mycobacterium porcinum
 Mycobacterium senegalense
Mycobacterium gordonae
Mycobacterium haemophilum
Mycobacterium immunogenum
Mycobacterium intracellulare
Mycobacterium kansasii
Mycobacterium lentiflavum
Mycobacterium malmoense
Mycobacterium marinum
Mycobacterium mucogenicum
Mycobacterium scrofulaceum
Mycobacterium simiae
Mycobacterium smegmatis
Mycobacterium szulgai

*Mycobacterium tuberculosis complex:*¹⁶
 Mycobacterium africanum
 Mycobacterium bovis
 Mycobacterium canettii
 Mycobacterium microti
 Mycobacterium pinnipedii
 Mycobacterium tuberculosis
Mycobacterium xenopi

NOCARDIA

Nocardia abscessus
Nocardia asteroides
Nocardia brasiliensis
Nocardia cyriacigeorgica
Nocardia farcinica
*Nocardia nova*¹⁷
Nocardia otitidiscaviarum
Nocardia paucivorans
Nocardia pseudobrasiliensis
Nocardia transvalensis
Nocardia veterana
Nocardia wallacei

MOULDS

Acremonium sclerotigenum
Alternaria alternata
Aspergillus brasiliensis
Aspergillus flavus/oryzae
Aspergillus fumigatus
Aspergillus lentulus
Aspergillus nidulans
Aspergillus niger complex
Aspergillus sydowii
*Aspergillus terreus complex*¹⁸
Aspergillus calidoustus
Aspergillus versicolor
Blastomyces dermatitidis
Cladophialophora bantiana
Coccidioides immitis/posadasii
Curvularia hawaiiensis
Curvularia spicifera
Epidermophyton floccosum
Exophiala dermatitidis

14. There is a possibility of cross-identification between *Yersinia similis* and *Yersinia pseudotuberculosis*
15. The *Mycobacterium fortuitum* complex or group displayed in VITEK[®] MS includes the seven most prominent and closely related species. The VITEK[®] MS group differs from the ones reported in the literature that may include up to 13 species.
16. The *Mycobacterium tuberculosis* complex displayed in VITEK[®] MS includes the four most prevalent species based on different worldwide geographic regions. It does not include two additional species *Mycobacterium microti* and *Mycobacterium pinnipedii* as reported in the literature.
17. In KB V3.0.0, *Nocardia nova* is displayed as a low discrimination result with *Nocardia africana* but only *Nocardia nova* has been clinically validated.
18. Both *Aspergillus alabamensis* and *Aspergillus niveus* are not identified as *Aspergillus terreus* complex. No identification is expected with either species.
19. All of the no identifications for this organism were from multiple replicates of the same isolate.
20. *Mucor racemosus* f. *sphaerosporus* is not identified as *Mucor racemosus* complex. No identification is expected. The *Mucor racemosus* complex is comprised of *M. racemosus* f. *brunneus*, *M. racemosus* f. *chibinensis*, *M. racemosus* f. *racemosus*, and *M. racemosus* f. *sphaerosporus*.
21. 3 out of the 5 no identifications for this organism were from multiple replicates of the same isolate.
22. All of the discordant identifications for this organism were from multiple replicates of the same isolate.



510(k) SUMMARY

VITEK[®] MS

510(k) Submission Information:

Submitter's Name:	bioMérieux, SA
Address:	3 Route de Port Michaud La Balme les Grottes, 38390 (France)
Contact Person:	Nathan Hardesty Manager, Regulatory Affairs Microbiology
Phone Number:	314 -731-8666
Fax Number:	314-731-8689
Date of Preparation:	October 14, 2016

B. Device Name:

Formal/Trade Name:	VITEK [®] MS
Classification Name:	21 CFR 866.3361, Mass spectrometer system for clinical use for the identification of microorganisms
Common Name:	VITEK [®] MS, VITEK MS

C. Predicate Device: VITEK[®] MS (DEN130013 / K124067)

D. 510(k) Summary:

VITEK[®] MS is mass spectrometry system using matrix-assisted laser desorption/ionization–time of flight mass spectrometry (MALDI-TOF MS) for the identification of microorganisms cultured from human specimens. The VITEK[®] MS is a qualitative *in vitro* diagnostic device indicated for use in conjunction with other clinical and laboratory findings to aid in the diagnosis of bacterial, yeast and mould infections.

The VITEK[®] MS is intended for laboratory use by professional users who are trained in microbiology and good laboratory practices.

This 510(k) is an update to the VITEK[®] MS (Mass Spectrometry) clinical knowledge base (KB v3.0.0) for the purposes of identifying *Mycobacterium*, *Nocardia*, and mould isolates. As the VITEK[®] MS KB v3.0.0 update includes new indications for use on the VITEK[®] MS system, new clinical data was required to establish safety and effectiveness. To account for the detection of higher mass peaks, relevant for some

bioMérieux, Inc.

595 Anglum Road, Hazelwood, Missouri 63042-2320, USA Phone: 314/731-8500 800/638-4835 Fax: 314/731-8700

<http://www.biomerieux-usa.com>

Page 001 - 010-1

moulds and *Mycobacterium*, the VITEK® MS acquisition station software was optimized (in v1.5.0) .

Microorganism identifications are made via matrix-assisted laser desorption/ionization – time of flight mass spectrometry (MALDI-TOF MS) technology, which includes the three basic principles of ionization, separation, and detection. Depending on the isolate culture, the analyte sample may be directly spotted to a target slide, or for *Mycobacterium*, *Nocardia* and mould it must be processed/inactivated before adding to the target slide. Once spotted to the target slide, a matrix is added for the purpose of easy sublimation and strong absorbance in the laser wavelength employed by the instrument.

The slide is then loaded onto the instrument, where a laser targets the sample spot and pulses the isolate spot, resulting in vibrational excitation of matrix and analyte molecules. The matrix transfer protons to the analyte resulting in a positive charge. The ionized molecules are then accelerated in an electromagnetic field and a grid electrode in the ionization chamber. The velocity of the molecules depends on the mass-to-charge (m/z) ratio of the analyte, with heavier molecules having a higher moment of inertia resulting in a lower velocity.

The time of flight is measured precisely by the ions arrival at a particle detector. Based on the time of flight, the m/z ratio of each particle can be determined, and a mass spectrum of the analyte sample mixture is generated. The mass spectrum displays results as a series of peaks (spectrum) which correspond to the ionized proteins derived from the analyte sample. The mass spectra are sufficiently distinctive to allow taxonomic characterization at the genus and species.

Clinical performance testing for the VITEK® MS v3 / KB v3.0.0 was conducted at five sites, for the purpose of an expanded intended use for *Mycobacterium*, *Nocardia*, and moulds. Performance testing included a total of 2,695 *Mycobacteria*, *Nocardia* and mould isolates, and evaluation of performance was a comparison of the VITEK MS identification results to molecular sequencing methods (i.e. the reference method). Testing *Mycobacteria* isolates included samples from both solid and liquid culture media.

Clinical strains for all isolates tested from solid and liquid media at all sites combined show an acceptable agreement rate of 94.6% (correct single choice identification plus a low discrimination correct genus result), as well as with each organism group [moulds (92.7%), mycobacteria (96.5%) and *Nocardia* (97.9%)]. A combined very low error rate of 0.7% (19/2695) was obtained with all isolates tested, and the combined no identification rate was acceptable at 4.7% (127/2695).

When excluding No ID results, clinical strains for all isolates tested from solid and liquid media at all sites combined show a high agreement rate of 99.3% (correct single choice identification plus a low discrimination correct genus result), as well as high agreement

with each organism group [moulds (99.1%), mycobacteria (99.6%) and *Nocardia* (99.2%)]. A very low error rate of 0.7% (19/2568) was obtained with all isolates combined.

One hundred well-characterized challenge strains included 50 moulds, 35 mycobacteria and 15 *Nocardia* strains. Three trial sites tested the mould challenge strains, and three trial sites tested the mycobacteria and *Nocardia* challenge strains. 95.7% (287/300) of Challenge results were correct one choice, and 1.0% (3/300) of results were low discrimination correct genus, for a combined agreement of 96.7% (290/300), for the mould, mycobacteria and *Nocardia* challenge isolates tested at all sites combined. No misidentifications were obtained, and no identification was obtained with 10 (3.3%) of all challenge isolates tested at all sites combined.

When excluding the No ID Challenge results, 99.0% (287/290) of results were correct one choice, and 1.0% (3/290) of results were low discrimination correct genus, for a combined agreement of 100% (290/290) for the mould, mycobacteria and *Nocardia* challenge isolates tested at all clinical trial sites combined.

Quality control strains *Mycobacterium smegmatis*, ATCC 19420, and *Nocardia farcinica*, ATCC 3308, showed 100% agreement, the quality control strain *Aspergillus brasiliensis*, ATCC 16404, showed 97.5% agreement, and the negative control showed 99.7% agreement.