



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

MedCom GmbH  
% Mr. Johannes Messow  
Quality Manager  
Rundeturmstr. 50  
Darmstadt, Hessen 64283  
GERMANY

May 1, 2017

Re: K163119  
Trade/Device Name: NaviSuite SSI Edition  
Regulation Number: 21 CFR 892.2050  
Regulation Name: Picture archiving and communications system  
Regulatory Class: II  
Product Code: LLZ  
Dated: February 3, 2017  
Received: April 3, 2017

Dear Mr. Messow:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

 For

Robert Ochs, Ph.D.  
Director  
Division of Radiological Health  
Office of In Vitro Diagnostics  
and Radiological Health  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K163119

Device Name

NaviSuite SSI Edition

### Indications for Use (Describe)

The NaviSuite SSI Edition is used to support a clinical ultra-sound session and follow percutaneous procedure by means of providing additional real-time image information from a 2nd modality like CT or MR or US or its combination with the possibility to plan the ideal treatment procedure. The second modality or their combinations are acquired and 3D is generated inside the software permitting the anatomical correlation with the patient position.

"The second modality image is not intended to be used as a standalone diagnostic image since it represents information of a patient that could not be congruent with the current (actual) patient position and shall therefore always be seen as an additional source of information."

"The 2nd modality image only should not be used for diagnostic purposes."

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

#### **\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## V. 510(k) Summary of Safety and Effectiveness

### A. Submitter

Name: MedCom GmbH

Address: Dolivostr. 11  
Darmstadt, HE 64293  
Germany

Establishment Reg.: 3006579682

Telephone: +49 (6151) 95147-285

Fax: +49 (6151) 95147-20

Contact: Mr. Johannes Messow  
[jmessow@medcom-online.de](mailto:jmessow@medcom-online.de)

Date: April 26, 2017

### B. Device

Trade Name: NaviSuite SSI Edition

Common name: SuperSonic Navigator

Classification: Regulatory Class: II

Product Code: LLZ / System, Image  
Processing, Radiological

CFR Section: 892.2050

Panel: Radiology

**C. Predicate Devices**

Device trade name: MIM 3.5 (CIRCA)  
510(k) number: K052379  
Company name: MIMVISTA CORP.  
Classification Number: 892.2050  
Classification: Class II  
Product code: LLZ

Device trade name: MyLabClass C 6250 MyLabTwice 6200  
510(k) number: K153277  
Company name: Esaote S.P.A.  
Classification Number: 892.1550  
Classification: Class II  
Product code: IYN  
Subsequent codes: ITX, IYO

**D. Reason for Submission**

New device application

**E. Standards**

1. ISO 14971:2007, Medical devices - Application of risk management to medical devices. (General I (QS/RM))
2. IEC 62304:2006, Medical Device Software - Software Life Cycle Processes. (Software/Informatics)

**F. Description**

NaviSuite SSI Edition 0.9.10 is a medical product to support a clinical ultrasound session and follow percutaneous procedure by means of providing additional real-time image information from a 2nd modality like CT or MR or US or its combination with the possibility to plan the ideal treatment procedure. The second modality or their combinations are acquired and a 3D is generated inside the software permitting the anatomical correlation with the patient position.


By using the 2nd modality image the user gains security in assessing the morphology of the ultrasound image and obtains a virtual feed-back on the procedure provided by the virtual percutaneous tool position enabled by a dedicated receiver, its treatment region delineated by custom tables and a dedicated computation in 3D of the target region.

The 2nd modality image is not intended to be used as a standalone diagnostic image since it represents information of a patient probably not congruent with the current (real) patient situation and shall therefore always been seen as an additional source of information.

NaviSuite SSI Edition is a software designed for use together with an ultrasound device:

| Device                                                     | Type      | 510(k) / registration number |
|------------------------------------------------------------|-----------|------------------------------|
| <b>System, Imaging,<br/>Pulsed Doppler,<br/>Ultrasonic</b> | AIXPLORER | K142100                      |

This hardware device is legally marketed in the US as listed in the table above.



System Overview

## **G. Indications for Use**

The NaviSuite SSI Edition is used to support a clinical ultra-sound session and follow percutaneous procedure by means of providing additional real-time image information from a 2nd modality like CT or MR or US or its combination with the possibility to plan the ideal treatment procedure. The second modality or their combinations are acquired and 3D is generated inside the software permitting the anatomical correlation with the patient position.

The second modality image is not intended to be used as a standalone diagnostic image since it represents information of a patient that could not be congruent with the current (actual) patient position and shall therefore always be seen as an additional source of information.

The 2nd modality image only should not be used for diagnostic purposes.

NaviSuite SSI Edition is intended for prescription use.

## **H. Intended Use**

The main purpose of the medical product NaviSuite SSI Edition is to support a clinical ultra-sound session and follow percutaneous procedure by means of providing additional real-time image information from a 2nd modality like CT or MR or US or its combination with the possibility to plan the ideal treatment procedure. The second modality or their combinations are acquired and 3D is generated inside the software permitting the anatomical correlation with the patient position.

By using the 2nd modality image the user gains security in assessing the morphology of the ultrasound image and obtains a virtual feed-back on the procedure provided by the virtual percutaneous tool position enabled by a dedicated receiver, its treatment region delineated by custom tables and a dedicated computation in 3D of the target region.

The 2nd modality image is not intended to be used as a standalone diagnostic image since it represents information of a patient probably not congruent with the current (real) patient situation and shall therefore always be seen as an additional source of information.

NaviSuite SSI Edition works in Virtual Biopsy mode also without 2nd modality data. In this case needle guidance is shown as an overlay over the real-time US image. The user can adjust the entry angle and trajectory of the needle before doing percutaneous procedure.

NaviSuite SSI Edition system should only be operated by trained personnel who are familiar with its user manual and the operation

The 2nd modality image only should not be used for diagnostic purposes.

The product NaviSuite SSI Edition was developed by MedCom in cooperation with SSI (France).

## I. Technological Comparison to Predicate Devices

The device NaviSuite SSI Edition is substantially equivalent to the predicate device MyLabTwice 6200 (K153277).

It is the same software used in the MyLabTwice 6200 device which was ported to the LINUX operating system of the AIXPLORER ultrasound device. Differences exist in the hardware and the user interfaces. In comparison to MyLabTwice, the NaviSuite software receives the live US image through a shared memory channel where in the MyLabTwice it is grabbed by a frame grabber card. In contrast to the MyLabTwice solution, the NaviSuite SSI Edition does not contain a planning environment and the contouring possibilities have been reduced to defining a ball target at places of interest instead of freehand contouring used in the MyLabTwice device.

### MIM 3.5 (CIRCA)

The MIM 3.5 and NaviSuite SSI Edition software can load and display DICOM datasets. The DICOM files are written following a standard which every DICOM compliant software has to follow in order to parse the data. Both software order the slices based on their spatial position and orientation to generate a 3D object of the scanned body region and can compute and visualize virtual cuts and slices through the 3D object.

## J. Non-clinical Performance Data

Non-clinical verification and validation software tests were conducted to confirm that NaviSuite SSI Edition meets its intended use and is safe and effective.

### System Accuracy Test:

The software cannot determine the accuracy of the treatment procedure. The Therapist or Physician is responsible for verifying the correctness of the registration. For this means the software offers side by side and overlay views (with a real-time US image).

Needle navigation was tested using a CIRS Abdominal Phantom to confirm the software projected virtual needle position over the real time US image.

### Tracking Quality Test:

The system can detect low sensor tracking quality and is able to show if the tracking is distorted (for example by metallic objects etc.).

Tests for verifying the tracking accuracy and sensor in range status have been carried out.

Statement from the User Manual about the systems accuracy:

## 2. Security

*It is not admitted to ground a diagnosis just on the results of NaviSuite. In particular the displayed image of the 2nd modality shall never be used as a standalone source for assessments since it represents computed and thus artificial information. MedCom is not engaging in any responsibility for indirect harms due to the use and evaluation of diagnosis data furnished by NaviSuite.*

## 6. System Accuracy

### **2<sup>nd</sup> Modality Image Display**

*The system accuracy concerning the matching between the ultrasound image and the 2<sup>nd</sup> modality image depends highly on the registration process and the condition of the patient compared to the situation when the 2<sup>nd</sup> modality (i.e. the CT scan) has been acquired. This also includes the breathing of the patient. Thus, please be aware that the current patient breath state should be comparable to the one during the 2<sup>nd</sup> modality acquisition.*

*However, from the technical point of view an “overlapping” accuracy of at least 5mm can be achieved under the presumption that the registration has been done with an overall error better than 5mm.*

## K. Conclusion

Based on the information provided in this Premarket Notification MedCom concludes that NaviSuite SSI Edition is as safe and effective and substantially equivalent to the predicate devices described herein.