



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

OsteoMed LLC  
Kathryn Jayne  
Senior Specialist, Regulatory Affairs  
3885 Arapaho Rd  
Addison, Texas 75001

April 4, 2017

Re: K163303

Trade/Device Name: OsteoMed ExtremiFix Mid And Large Screw System  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener  
Regulatory Class: Class II  
Product Code: HWC, HTN  
Dated: March 8, 2017  
Received: March 10, 2017

Dear Ms. Kathryn Jayne:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Mark N. Melkerson -S**

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: January 31, 2017  
See PRA Statement below.

## Indications for Use

510(k) Number (if known)

K163303

Device Name

OsteoMed ExtremiFix Mid & Large Screw System

Indications for Use (Describe)

The OsteoMed ExtremiFix Mid & Large Screw System is indicated for use in bone reconstruction, osteotomy, arthrodesis, and fracture fixation of foot, ankle, and long bones (upper and lower extremity). The screws are intended for single use only. The system drills and guide wires are single use instruments.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

### FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



---

## 510(k) SUMMARY

### I. SUBMITTER

OsteoMed  
3885 Arapaho Rd.  
Addison, TX 75001

Phone: 972-677-4766  
Fax: 800-390-2620  
Email: [kjayne@osteomed.com](mailto:kjayne@osteomed.com)

Contact: Kathryn Jayne  
Phone: 972-677-4766  
Fax: 800-390-2620  
Email: [kjayne@osteomed.com](mailto:kjayne@osteomed.com)

Alternate Contact: Alicia Hemphill  
Phone: 972-677-4786  
Email: [ahemphill@osteomed.com](mailto:ahemphill@osteomed.com)

Date Prepared: March 28, 2017

### II. DEVICE

Name of the Device: *OsteoMed ExtremiFix Mid and Large Screw System*  
Common or Usual Name: Bone Fixation Screw System  
Classification Name: Smooth or threaded metallic bone fixation fastener  
Regulation: 888.3040  
Regulatory Class: II  
Product Code: HWC, HTN

### III. PREDICATE DEVICE

Primary Predicate: OsteoMed Headless Cannulated Screw (K063298)  
Additional Predicate: OsteoMed Cannulated Screw System (K151021)  
Additional Predicate: Synthes Sterile 4.5 mm Cannulated Screws (K963172)  
Additional Predicate: Smith & Nephew Richards, Inc. Screw (K951389)  
Additional Predicate: OsteoMed Super Screw System (K954330)

---

#### **IV. DEVICE DESCRIPTION**

The OsteoMed ExtremiFix Mid & Large Screw System is a rigid fixation system consisting of screws in both cannulated and solid versions, all of which are available in various overall and distal thread lengths to accommodate specific patient anatomies. The headed and headless compression screws are made of biocompatible Ti 6-Al 4-V Titanium Alloy. Longer sizes of the single-use only implants are provided individually sterile packed, while the majority of screws and all instruments are provided in modules to allow for customization specific to the surgical indication. The system includes washers for use with 4.5mm, 5.5mm, 6.5mm and 7.0mm headed screws. The washers are made of biocompatible Titanium Alloy. The OsteoMed ExtremiFix Mid & Large Screw System is sterilized in a sterilization tray which is available from OsteoMed.

The subject device is intended for use in a healthcare facility/hospital for use by a clinician. It is a prescription device.

#### **V. INDICATIONS FOR USE**

The OSTEOMED ExtremiFix Mid & Large Screw System is indicated for use in bone reconstruction, osteotomy, arthrodesis, and fracture fixation of foot, ankle, and long bones (upper and lower extremity). The screws are intended for single use only. The system drills and guide wires are single use instruments.

#### **VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE**

##### **Performance Characteristics & Testing & Clinical**

Verification testing was conducted to ensure the subject screws performed equal or better compared to the predicate devices. The subject screws were also tested against the Synthes predicate reference device to ensure the design features met the required mechanical strength criteria for their intended use. The screws with the dual lead technology underwent verification evaluation to ensure the new design features met the mechanical strength criteria for the intended use. The screws were compared to their respective predicate.





Performance equivalence was shown through the verification comparison to the predicate devices.





Clinical Testing is not required to support substantial equivalence.

##### **Substantial Equivalence**

---

The basis of substantial equivalence for this device, the OsteoMed ExtremiFix Mid and Large Screw System, is based on similarities in indications for use, intended use, material, function, performance, design, technology, sterilization, and operational principles to the OsteoMed predicates and Synthes predicate. Performance comparisons were performed which verified that the new screw system met required mechanical strength criteria for their intended use compared to the legally marketed predicate devices listed in this summary. OsteoMed has shown that the non-clinical tests demonstrate that the devices are as safe and as effective as the predicate devices.

System/Device Name	Subject Device	Primary Predicate	Secondary Predicate	Reference Predicate
	OsteoMed ExtremiFix Mid and Large Screw System 	OsteoMed Headless Cannulated Screw System 	OsteoMed Cannulated Screw System 	Synthes Sterile 4.5mm Cannulated Screws 
510(k) Number	Pending	K063298	K151021	K963172
Regulation No.	888.3040	888.3040	888.3040	888.3040
Product Code	HWC, HTN	HWC	HWC	HWC
Device Class	II	II	II	II
Regulatory Panel	Orthopedic	Orthopedic	Orthopedic	Orthopedic
Indication for Use / Intended Use	Fracture fixation, reconstruction, osteotomy, and arthrodesis procedures of foot, ankle, and long bones (upper and lower extremity).	Bone reconstruction, osteotomy, arthrodesis, joint fusion, fracture repair, and fracture fixation of bones appropriate for the size of the device. Screws are intended for single use only.	Bone Fixation of hand and foot following trauma or osteotomy. Screws and washers are intended for single use only.	Fracture fixation of long bones and long bone fragments.
Material	Titanium Alloy - Ti 6Al-4V (ASTM F136)	Titanium Alloy - Ti 6Al-4V (ASTM F136)	Titanium Alloy - Ti 6Al-4V (ASTM F136)	Titanium Alloy - Ti 6Al-4V and Stainless Steel
Technology	Screws to fixate Bone	Screws to fixate Bone	Screws to Fixate Bone	Screws to Fixate Bone
Operating Principles	Rigid fixation	Rigid fixation	Rigid fixation	Rigid fixation
Insertion Material	Bone	Bone	Bone	Bone
Screw Length (mm)	20mm - 180mm	10mm -120mm	10mm -140mm	20mm - 80mm
Screw Diameter	4.5mm - 7.0mm	2.0mm - 6.5mm	3.0mm -8.0mm	4.0 mm
Head Type	Hexalobe	Trilobe	Cruciform Hex	Hex
Headless (HL)	Yes	Yes	No	No
Headed (HD)	Yes	No	Yes	Yes
Cannulated	Yes	Yes	Yes	Yes
Solid	Yes	No	No	No
Short Thread (ST)	Yes	Yes	Yes	Yes

System/Device Name	Subject Device	Primary Predicate	Secondary Predicate	Reference Predicate	
		<b>OsteoMed ExtremiFix Mid and Large Screw System</b> 	<b>OsteoMed Headless Cannulated Screw System</b> 	<b>OsteoMed Cannulated Screw System</b> 	<b>Synthes Sterile 4.5mm Cannulated Screws</b> 
<b>510(k) Number</b>	<b>Pending</b>	<b>K063298</b>	<b>K151021</b>	<b>K963172</b>	
Long Thread (LT)	Yes	Yes	No	No	
Full Thread (FT)	Yes	No	No	Yes	
Screw Type	HL Cannulated ST & LT HD Cannulated ST, LT,& FT Headed Solid ST	Headless Cannulated	Headed Cannulated	Headed Cannulated	
Double-Lead	Yes	No	No	No	
Low Profile Heads	Yes	Yes	Yes	TBD	
Self-Drilling	Yes	Yes	Yes	Yes	
Self-Tapping	Yes	Yes	Yes	Yes	
Lag	Yes	Yes	Yes	Yes	
Reverse Cutting Flutes	Yes	No	No	Yes	
Washers	Availability	Yes	No	Yes	
	Part Numbers	316-0325 316-0326	N/A	316-0225	419.91
	Screw Size	4.5 - 5.5 mm	N/A	3.0 mm	4.5mm
	Thickness	1.0 mm	N/A	1.0 mm	N/A
	Outer Diameter	10.0 -11.0 mm	N/A	12.4 mm	10 mm
	Material	Titanium Alloy	N/A	Titanium Alloy	CPTi
Instrumentation	Single Use and Re-usable Preparation Instruments	Single Use and Re-usable Preparation Instruments	Single Use and Re-usable Preparation Instruments	Single Use and Re-usable Preparation Instruments	
Instrumentation Materials	Stainless Steel, Anodized Aluminum, Medical grade polymer	Stainless Steel, Anodized Aluminum, Medical grade plastic	Stainless Steel, Anodized Aluminum, Medical grade polymer	Stainless Steel	

## VII. PERFORMANCE DATA

The following performance data were provided in support of the substantial equivalence determination.

### Biocompatibility

The biocompatibility evaluation for the device was conducted in accordance with the FDA Guidance Document, issued on June 16, 2016, *Use of International Standard ISO 10993-1, "Biological evaluation of medical devices - Part 1:*

*Evaluation and testing within a risk management process” – Guidance for Industry and Food and Drug Administration Staff.*

- Cytotoxicity
- Sensitization
- Irritation
- Systemic toxicity
- Sub-chronic toxicity
- Genotoxicity
- Implantation
- Pyrogen testing

The instruments are considered tissue contacting for a duration of less than 24 hours while the implants are considered permanent implants. The Titanium alloy implant material conforms to ASTM F-136 for chemical composition.

#### **Electrical Safety and Electromagnetic Compatibility (EMC)**

The OsteoMed ExtremiFix Mid & Large Screw System is not an active device; therefore, Electrical Safety and Electromagnetic Compatibility (EMC) testing is not applicable.

#### **Software Verification and Validation Testing**

The OsteoMed ExtremiFix Mid & Large Screw System does not contain software; therefore, Software Verification and Validation Testing is not applicable.

#### **Bench Testing**

The following bench evaluations were conducted:

- Self-Tapping Force
- Safety Factor (Failure Torque / Insertion Torque)
- Pullout Force
- Torsional Strength (Solid Screws Only)
- Bending Strength
- Sterilization Validation
- Endotoxin Testing

#### **Animal Study**

No animal studies were performed to demonstrate safety and efficacy.

#### **Clinical Studies**

No clinical studies were performed to demonstrate safety and efficacy.

## **VIII. CONCLUSIONS**

---





The basis of substantial equivalence for the proposed device is based on similarities in indications for use, material, function, performance, design, technology, sterilizations, and operating principles. The non-clinical data support the substantial equivalence and demonstrates that the device should perform as intended in the specified use conditions.

(End of Summary)

---