



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

GE Healthcare
% Mr. Bryan Behn
Regulatory Affairs Director
9900 Innovation Drive
WAUWATOSA WI 53226

February 10, 2017

Re: K163596
Trade/Device Name: LOGIQ P9 and LOGIQ P7
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, ITX
Dated: December 20, 2016
Received: December 21, 2016

Dear Mr. Behn:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

 For

Robert Ochs, Ph.D.
Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K163596

Device Name

LOGIQ P9 and LOGIQ P7

Indications for Use (Describe)

The device is intended for use by a qualified physician for ultrasound evaluation of Fetal; Abdominal; Pediatric; Small Organ (breast, testes, thyroid); Neonatal Cephalic; Adult Cephalic; Cardiac (adult and pediatric); Peripheral Vascular; Musculo-skeletal Conventional and Superficial; Urology (including prostate); Transrectal; Transvaginal; and Intraoperative (abdominal, thoracic, vascular).

When Pinpoint™ GT Technology, from C.R. Bard, Inc., is included with the system, the Indications for Use include: Pinpoint GT Technology is intended to provide clinicians with visual tools for passive magnetic tracking of a needle with respect to ultrasound image data

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



Indications for Use Forms

The following forms represent indications with clinical applications and exam types along with the modes of operation for the LOGIQ P9 and LOGIQ P7 system. Combinations identified “P” for the transducers represents those previously cleared with LOGIQ P9 and LOGIQ P7 and *P for those cleared on another GE Ultrasound system and those identified and “N” are new. Please see section 11 Table 11.2.1 for information on previous clearance information on these transducers.



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 Ultrasound System

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]	P	P	P	P	P	P	P	P	P	P	5,6
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P	3,5,6
Pediatric	P	P	P	P	P	P	P	P	P	P	3,5,6
Small Organ ^[2]	P	P	P	P	P	P	P	P	P	P	3,5,6
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P	5,6
Adult Cephalic	P	P	P	P	P	P	P	P	P	P	5,6
Cardiac Adult & Pediatric	P	P	P	P	P	P	P	P	P	P	
Peripheral Vascular	P	P	P	P	P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional	P	P	P	P	P	P	P	P	P	P	3,5,6
Musculo-skeletal Superficial	P	P	P	P	P	P	P	P	P	P	3,5,6,9
Other ^[4]	P	P	P	P	P	P	P	P	P	P	3,5,6
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]	P	P	P	P	P	P	P	P	P	P	3,5,6
Transvaginal	P	P	P	P	P	P	P	P	P	P	3,5,6
Transurethral											
Intraoperative ^[8]	P	P	P	P	P	P	P	P	P	P	3,5,6
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[9] Needle guidance with Pinpoint™ GT needle technology;

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with L12n-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic	N	N			N		N	N	N		
Fetal / Obstetrics ^[7]											
Abdominal ^[1]	N	N			N		N	N	N		
Pediatric	N	N			N		N	N	N		
Small Organ ^[2]	N	N			N		N	N	N		
Neonatal Cephalic	N	N			N		N	N	N		
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	N	N			N		N	N	N		
Musculo-skeletal Conventional	N	N			N		N	N	N		
Musculo-skeletal Superficial	N	N			N		N	N	N		[9]
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]	N	N			N		N	N	N		
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [9] Needle guidance with Pinpoint™ GT needle technology;
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare

510(k) Premarket Notification Submission

GE LOGIQ P9 and LOGIQ P7 with P6D Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric											
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic			*P	*P							
Cardiac Adult & Pediatric			*P	*P							
Peripheral Vascular			*P	*P							
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; *P = previously cleared by FDA in (K160182)

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

GE LOGIQ P9 and LOGIQ P7 with L4-12t-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric	P	P	P		P	P	P	P	P	P	[3,6]
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	[3,6]
Neonatal Cephalic	P	P	P		P	P	P	P	P	P	[3,6]
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	[3,6]
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	[3,6]
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	[3,6]
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with L10-22-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric											
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic	P	P	P		P	P	P	P	P	P	
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with L3-9i-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric											
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]	P	P	P		P	P	P	P	P	P	
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with E8CS--RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]	P	P	P		P	P	P	P	P	P	[3,6]
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	[3,6]
Pediatric											
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	[3,6]
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]	P	P	P		P	P	P	P	P	P	[3,6]
Transvaginal	P	P	P		P	P	P	P	P	P	[3,6]
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with BE9CS-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric											
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	[3,6]
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]	P	P	P		P	P	P	P	P	P	[3,6]
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with RIC5-9A-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]	P	P	P		P	P	P	P	P	P	[5,6]
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	[5,6]
Pediatric											
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	[5,6]
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal	P	P	P		P	P	P	P	P	P	[5,6]
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with 12S-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric	P	P	P	P	P	P	P	P	P	P	
Small Organ ^[2]											
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P	
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with C1-5-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]	P	P	P	P	P	P	P	P	P	P	5,6
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P	3,5,6
Pediatric	P	P	P	P	P	P	P	P	P	P	3,5,6
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P	P	P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P	P	P	P	P	P	P	P	3,5,6
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with 4C-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]	P	P	P	P	P	P	P	P	P	P	5,6
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P	3,5,6
Pediatric	P	P	P	P	P	P	P	P	P	P	3,5,6
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P	P	P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P	P	P	P	P	P	P	P	3,5,6
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with 9L-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	3,5,6
Pediatric	P	P	P		P	P	P	P	P	P	3,5,6
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	3,5,6
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with ML6-15-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric	P	P	P		P	P	P	P	P	P	3,5,6
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	3,5,6
Neonatal Cephalic	P	P	P		P	P	P	P	P	P	
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	3,5,6
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with E8C-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]	P	P	P		P	P	P	P	P	P	5,6
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	
Pediatric											
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	3,5,6
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]	P	P	P		P	P	P	P	P	P	3,5,6
Transvaginal	P	P	P		P	P	P	P	P	P	3,5,6
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with L8-18i-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric	P	P	P		P	P	P	P	P	P	5,6
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic	P	P	P		P	P	P	P	P	P	
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	3,5,6
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]	P	P	P		P	P	P	P	P	P	3,5,6
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with P8D Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric											
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic			P	P							
Cardiac Adult & Pediatric			P	P							
Peripheral Vascular			P	P							
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

- Notes:
- [1] Abdominal includes Renal, GYN/Pelvic.
 - [2] Small organ includes breast, testes and thyroid
 - [3] Elastography Imaging - Elasticity.
 - [4] Other use includes Urology/Prostate
 - [5] 3D/4D Imaging mode
 - [6] Needle guidance imaging
 - [7] Includes infertility monitoring of follicle development
 - [8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)
 - [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with RAB2-6-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]	P	P	P		P	P	P	P	P	P	5,6
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	5,6
Pediatric											
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	5,6
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with L6-12-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	3,5,6
Pediatric	P	P	P		P	P	P	P	P	P	3,5,6
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	3,5,6
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	3,5,6
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with 12L-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	3,5,6
Pediatric	P	P	P		P	P	P	P	P	P	3,5,6
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	3,5,6
Neonatal Cephalic											
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	3,5,6
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	3,5,6
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging - Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D/4D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with 8C-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	5
Pediatric	P	P	P		P	P	P	P	P	P	5
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	5
Neonatal Cephalic	P	P	P		P	P	P	P	P	P	5
Adult Cephalic											
Cardiac Adult & Pediatric											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging – Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with 3Sc-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P	5,6
Pediatric	P	P	P	P	P	P	P	P	P	P	5,6
Small Organ ^[2]											
Neonatal Cephalic											
Adult Cephalic	P	P	P	P	P	P	P	P	P	P	
Cardiac Adult & Pediatric	P	P	P	P	P	P	P	P	P	P	
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging – Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

Diagnostic Ultrasound Indications for Use Form

GE LOGIQ P9 and LOGIQ P7 with 6S-RS Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other [Notes]
Ophthalmic											
Fetal / Obstetrics ^[7]											
Abdominal ^[1]											
Pediatric	P	P	P	P	P	P	P	P	P	P	
Small Organ ^[2]											
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P	
Adult Cephalic											
Cardiac Adult & Pediatric	P	P	P	P	P	P	P	P	P	P	
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal ^[8]											
Transvaginal											
Transurethral											
Intraoperative ^[8]											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA

Notes: [1] Abdominal includes Renal, GYN/Pelvic.

[2] Small organ includes breast, testes and thyroid

[3] Elastography Imaging – Elasticity.

[4] Other use includes Urology/Prostate

[5] 3D Imaging mode

[6] Needle guidance imaging

[7] Includes infertility monitoring of follicle development

[8] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV)

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostics and Radiological Health (OIR)

Prescription User (Per 21 CFR 801.109)



GE Healthcare
510(k) Premarket Notification Submission

510(k) Summary

In accordance with 21 CFR 807.92 the following summary of information is provided:

Date: Dec 20, 2016

Submitter: GE Healthcare
9900 Innovation Dr
Wauwatosa, WI 53226

Primary Contact Person: Bryan Behn
Regulatory Affairs Director
GE Healthcare
T:(414)721-4214
F:(414)918-8275

Secondary Contact Person: Chan Sook Kim
Regulatory Affairs Leader
GE Healthcare
GE Ultrasound Korea, Ltd.
T: +82 31 740 6307

Device: Trade Name: LOGIQ P9 and LOGIQ P7

Common/Usual Name: LOGIQ P9 and LOGIQ P7

Classification Names: Class II

Product Code: Ultrasonic Pulsed Doppler Imaging System, 21CFR 892.1550 90-IYN
Ultrasonic Pulsed Echo Imaging System, 21CFR 892.1560, 90-IYO
Diagnostic Ultrasound Transducer, 21 CFR 892.1570, 90-ITX

Primary Predicate Device: K161047 LOGIQ P9/P7 Diagnostic Ultrasound System

Reference Predicate K152758 Venue 50 Diagnostic Ultrasound System

Device(s): K160182 LOGIQ S7 Expert Diagnostic Ultrasound System

Device Description: The LOGIQ P9 and LOGIQ P7 is a full featured, general purpose diagnostic ultrasound system which consists of a mobile console approximately 53 cm wide, 69 cm deep and 157 cm high that provides digital acquisition, processing and display capability. The user interface includes a computer keyboard, specialized controls, 10.4-inch LCD touch screen and color 21.5-inch LCD image display.

Intended Use: The device is intended for use by a qualified physician for ultrasound evaluation of Fetal; Abdominal; Pediatric; Small



GE Healthcare

510(k) Premarket Notification Submission

Organ (breast, testes, thyroid); Neonatal Cephalic; Adult Cephalic; Cardiac (adult and pediatric); Peripheral Vascular; Musculo-skeletal Conventional and Superficial; Urology (including prostate); Transrectal; Transvaginal; and Intraoperative (abdominal, thoracic, vascular).

When Pinpoint™ GT Technology, from C.R. Bard, Inc., is included with the system, the Indications for Use include: Pinpoint GT Technology is intended to provide clinicians with visual tools for passive magnetic tracking of a needle with respect to ultrasound image data.

Technology: The LOGIQ P9 and LOGIQ P7 employs the same fundamental scientific technology as its predicate devices

Determination of Substantial Equivalence: Comparison to Predicate Devices
The LOGIQ P9 and P7 systems are substantially equivalent to the predicate devices with regard to intended use, imaging capabilities, technological characteristics and safety and effectiveness.

- The systems are all intended for diagnostic ultrasound imaging and fluid flow analysis.
- The LOGIQ P9/P7 and predicate LOGIQ P9/P7 systems have the same clinical intended uses.
- The LOGIQ P9/P7 and predicate LOGIQ P9/P7 systems have the same imaging modes.
- The LOGIQ P9/P7 and predicate LOGIQ P9/P7 systems transducers are identical except for the addition of 2 new transducers L12n-RS (similar to L12n-SC on predicate Venue50 K152758) and P6D (identical to LOGIQ S7 Expert K160182)
- The LOGIQ P9/P7 and predicate LOGIQ P9/P7 systems have the same indications for use
- The systems are manufactured with materials which have been evaluated and found to be safe for the intended use of the device.
- The systems have acoustic power levels which are below the applicable FDA limits.
- The LOGIQ P9/P7 and predicate LOGIQ P9/P7 systems have similar capability in terms of performing measurements, capturing digital images, reviewing and reporting studies.
- The LOGIQ P9/P7 and predicate LOGIQ P9/P7 systems



GE Healthcare

510(k) Premarket Notification Submission

have been designed in compliance with approved electrical and physical safety standards.

- Migration of the Pinpoint GT Technology from C.R. Bard, which allows the user to track the needle location during interventional guidance. This technology is equivalent to that cleared by C.R. Bard, Inc. (previously cleared with Venue 50 K152758).
- The LOGIQ P9/P7 adds an scanning parameter of image rotation for 90° and 270° as a minor change to the 180° rotation already cleared.

Summary of Non-Clinical Tests:

The device has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical, electromagnetic, and mechanical safety, and has been found to conform to applicable medical device safety standards. LOGIQ P9 and LOGIQ P7 and its applications comply with voluntary standards:

- AAMI/ANSI ES60601-1, Medical Electrical Equipment – Part 1: General Requirements for Safety, 2005/C1:2012
- IEC60601-1-2, Medical Electrical Equipment – Part 1-2:General Requirements for Safety – Collateral Standard: Electromagnetic Compatibility Requirements and Tests 2007
- IEC60601-2-37, Medical Electrical Equipment – Part 2-37:Particular Requirements for the Safety of Ultrasonic Medical Diagnostic and Monitoring Equipment, 2007
- NEMA UD 3, Standard for Real Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment:2004
- ISO10993-1, Biological Evaluation of Medical Devices- Part 1: Evaluation and Testing- Third Edition, 2009
- NEMA UD 2, Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment:2004
- ISO14971, Application of risk management to medical devices: Second edition 2007
- NEMA PS 3.1 - 3.20 (2011), Digital Imaging and



GE Healthcare

510(k) Premarket Notification Submission

Communications in Medicine (DICOM) Set. (Radiology)

The following quality assurance measures are applied to the development of the system:

- Risk Analysis
- Requirements Reviews
- Design Reviews
- Testing on unit level (Module verification)
- Integration testing (System verification)
- Performance testing (Verification)
- Safety testing (Verification)

Transducer materials and other patient contact materials are biocompatible.

Summary of Clinical Tests:

The subject of this premarket submission, LOGIQ P9 and LOGIQ P7, did not require clinical studies to support substantial equivalence.

Conclusion: GE Healthcare considers the LOGIQ P9 and LOGIQ P7 to be as safe, as effective, and performance is substantially equivalent to the predicate device(s).