



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

March 24, 2017

Beijing Adss Development Co., Ltd  
% Ray Wang  
General Manager  
Beijing Believe Technology Service Co., Ltd  
5-1206, Build 332, Dafangju, No. 25 Banbidian Rd.  
Liyuan Town, Tongzhou District, Beijing, 101121 CN

Re: K161692

Trade/Device Name: Diode Laser Therapy Machine  
Regulation Number: 21 CFR 878.4810  
Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And  
In Dermatology  
Regulatory Class: Class II  
Product Code: GEX,  
Dated: October 13, 2016  
Received: October 17, 2016

Dear Ray Wang:

This letter corrects our substantially equivalent letter of November 16, 2016.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR

Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

  
Jennifer R. Stevenson -S

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.  
Director  
Division of Surgical Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)

K161692

Device Name

Diode Laser Therapy Machine

Indications for Use (Describe)

The Diode Laser Therapy Machine is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.

Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**Tab #7 510(k) Summary**

This 510(k) Summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of SMDA 1990 and Title 21, CFR Section 807.92.

The assigned 510(k) Number: \_\_\_\_\_ K161692

1. Date of Preparation

06/07/2016

2. Sponsor

**Beijing ADSS Development Co., Ltd**

F6, Xiandao Bldg., Jinyuan Rd. 36, Daxing Economic Development zone, Beijing, China, 102628

Establishment Registration Number: Not yet registered or the Number

Contact Person: Gao Yurong

Position: sales manager

Tel: 86-10-83625120

Fax:86-10-83625121

Email: gyr@adss.com.cn

3. Submission Correspondent

Mr. Ray Wang

**Beijing Believe Technology Service Co., Ltd.**

5-1206, Build 332, DaFangJu, No.25 BanBiDian Rd.,

LiYuan Town, TongZhou District, Beijing, 101121, China

Tel: +86-10-18910677558

Fax: +86-10-52214696

Email: ray.wang@believe-med.com

#### 4. Identification of Proposed Device

Trade Name: Diode Laser Therapy Machine

Common Name: Powered Laser Surgical Instrument

Model(s): FG 2000/FG 2000-B/FG 2000-C/FG 2000-D

Regulatory Information:

Classification Name: Powered Laser Surgical Instrument

Classification: II;

Product Code: GEX;

Regulation Number: 21 CFR 878.4810;

Review Panel: General& Plastic Surgery;

Intended Use:

The Diode Laser Hair Therapy Machine is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.

Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.

#### 5. Device Description

The proposed device, Diode Laser Therapy System, is a surgical device, which is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI);

There are 4 models included, FG 2000, FG 2000-B, FG 2000-C and FG 2000-D, the four models have same intended use, mechanism of action, principle and specification, only difference is the configuration. The detailed difference shown as following:

Table 1 The Difference of Models

Model	FG 2000-B	FG 2000-C	FG 2000-D	FG 2000
Size	42*63*54cm	56*46*110cm	59*59*146 cm	56*46*112 cm
Weight	30kg	45kg	45kg	49 kg
Handpiece	Refer to Fig 16			Refer to Fig 4
Cooler Number	1	2	1	2

The main components of proposed device shown as following:

Table 2 Main Components of Proposed Device

Components	Function Description	Applied Model(s)
Handpiece	Deliver the laser to area to be treated	FG 2000/FG 2000-B/FG 2000-C/FG 2000-D
Touchscreen	The user interface and for controlling of the system	FG 2000/ FG 2000-B/ FG

		2000-C/FG 2000-D
Emergency Switch	Stop the system in case of emergency situation	FG 2000/ FG 2000-B/ FG 2000-C/FG 2000-D
Key Switch	Start the system	FG 2000/ FG 2000-B/ FG 2000-C/FG 2000-D
Connector	Connection of the device with the handpiece	FG 2000/ FG 2000-B/ FG 2000-C/FG 2000-D
Indicator Lamp	Indicate current working state of the appliance	FG 2000/ FG 2000-C
Foot Switch	Activate the laser emission	FG 2000/ FG 2000-B/ FG 2000-C/FG 2000-D

The proposed system provides 36 working modes, which are six modes for men and six modes for women, the men or women mode includes face, armpit, arm, body, bikini, leg mode respectively for different treatment part, and each particular treat mode includes three submode as mode 1, mode 2 and mode 3.

The difference for each mode is only the default parameter, but all parameters for each mode can be adjustable in the parameter range.

The treatment can be applied on different Fitzpatrick skin type, including I (White), II (White with pigment), III (Yellow), IV (Yellow with pigment), V (Brown) and VI (Black);

6. Identification of Predicate Device

Predicate 1#

510(k) Number: K141973

Product Name: Diode Laser Hair Removal System

Manufacturer: Beijing Anchorfree Technology Co., Ltd

Predicate 2#

510(k) Number: K123483

Product Name: Diode Laser

Manufacturer: Beijing Syntech Laser Co., Ltd

7. Non-Clinical Test Conclusion

Non clinical tests were conducted to verify that the proposed device met all design specifications as was Substantially Equivalent (SE) to the predicate device. The test results demonstrated that the proposed device complies with the following standards:

- IEC 60601-1:2012 Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance;
- IEC 60601-2-2:2007, Medical Electrical Equipment - Part 2-22: Particular Requirements For Basic Safety And Essential Performance Of Surgical, Cosmetic, Therapeutic And Diagnostic Laser Equipment;
- IEC 60825-1: 2007, Safety of laser products - Part 1: Equipment classification and requirements.
- IEC 60601-1-2:2007, Medical electrical equipment- Part 1-2: General requirements for basic safety and essential performance- Collateral standard: Electromagnetic compatibility- Requirements and tests.
- ISO 10993-5:2009, Biological Evaluation of Medical Device, Part 5-Tests for Vitro cytotoxicity
- ISO 10993-10:2002/Amd. 1: 2006, Biological Evaluation of Medical Device, Part 10-Test for irritation and delay-type hypersensitivity AMENDMENT 1
- Performance Testing for Spot Size Accuracy and Energy Output Accuracy.

8. Clinical Test Conclusion

No clinical study is included in this submission.

## 9. Substantially Equivalent (SE) Comparison

Table 3 General Comparison

Item	Proposed Device	Predicate Device	Predicate Device	Remark
Product Code	GEX	GEX	GEX	SE
Regulation Number	21 CFR 878.4810	21 CFR 878.4810	21 CFR 878.4810	SE
Intended Use	<p>The Diode Laser Therapy Machine is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.</p> <p>Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.</p>	<p>The Diode Laser Hair Removal System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.</p> <p>Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.</p>	<p>The Diode Laser is intended for use in dermatologic and general surgical procedures.</p> <p>The Standard Mode is intended for hair removal, permanent hair reduction.</p> <p>The FHR Mode is intended for hair removal, permanent hair reduction.</p> <p>The diode laser system is intended for use on all skin types (Fitzpatrick skin types I-VI), including tanned skin.</p>	SE, with difference in wording, while the actual indications are the same.
Configuration	Main Unit	Main Unit	Main Unit	SE
	Handpiece	Handpiece	Handpiece	SE
	Foot Control	Foot Control	Foot Control	SE
Principle of Operation	Diode Laser	Diode Laser	Diode Laser	SE



Table 4 Performance Comparison

Item	Proposed Device	Predicate Device	Predicate Device	Remark
Laser Type	Diode Laser	Diode Laser	Diode Laser	SE
Laser Classification	Class IV	Class IV	Class IV	SE
Laser Wavelength	808 nm	808 nm	808 nm	SE
Spot Size	1.44 cm <sup>2</sup>	1.44 cm <sup>2</sup>	1.2 cm <sup>2</sup>	SE
Fluence	2-120J/ cm <sup>2</sup>	1-120 J/cm <sup>2</sup>	1-120 J/cm <sup>2</sup>	SE
Frequency	1-10Hz	1Hz, 2Hz, 3Hz,10Hz	≤10 Hz	SE
Pulse Duration	9-143ms	2.9-348ms	5-200 ms	Discussion 1
Power Supply	AC 110V/50Hz-60Hz	AC220V, 50Hz/ AC110V, 60Hz	100-240 V 50/60Hz	SE
Dimension	56*46*112 cm	380mm×540mm×1200mm	460X 365 X350 mm	Discussion 2
	42*63*54cm			
	56*46*110cm			
	59*59*146 cm			
Weight	49 kg	55kg	25 kg	Discussion 2
	30kg			
	45kg			
	45kg			

## Discussion 1

The proposed device is different in Pulse Duration of HR mode from the predicate, the proposed devices' pulse duration range is cover the predicates' and both proposed device and predicate has same Fluence, therefore, this difference will not affect the substantially equivalency.

## Discussion 2

The proposed device is different in dimension and weight from the predicate device, because the proposed device is a trolley type, while the predicate device is a desktop type. By complying with IEC 60601-1, the mechanical performance of the proposed device is determined to be accepted, therefore, this difference will not affect the substantially equivalency.

Table 5 Safety Comparison

Item	Proposed Device	Predicate Device	Remark
Patient Contact Materials and Biocompatibility			
Patient Contact Materials	Sapphire in handpiece and handpiece tip (Stainless Steel)	Stainless steel and Sapphire in handpiece	SE
Cytotoxicity	No Cytotoxicity	Comply with ISO 10993-1	SE
Sensitization	No evidence of sensitization		
Irritation	No evidence of irritation		
EMC, Electrical and Laser Safety			
Electrical Safety	Comply with IEC 60601-1, IEC 60601-2-22	Comply with IEC 60601-1, IEC 60601-2-22	SE
EMC	Comply with IEC 60601-1-2	Comply with IEC 60601-1-2	SE
Laser Safety	Comply with IEC 60601-2-22, IEC 60825	Comply with IEC 60601-2-22, IEC 60825	SE

#### 10. Substantially Equivalent (SE) Conclusion

Based on the comparison and analysis above, the proposed device is determined to be Substantially Equivalent (SE) to the predicate device.