



December 15, 2017

International Biophysics Corporation  
Geoff Marcek  
Director, Engineering and Quality  
2101 E. St. Elmo Road  
Austin, Texas 78744

Re: K170029

Trade/Device Name: FloPump 32mL Centrifugal Pump  
Regulation Number: 21 CFR 870.4360  
Regulation Name: Nonroller-type cardiopulmonary bypass blood pump  
Regulatory Class: Class II  
Product Code: KFM  
Dated: November 6, 2017  
Received: November 7, 2017

Dear Mr. Marcek:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820);

and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Nicole G. Ibrahim -S

for Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

**Indications for Use**

Form Approved: OMB No. 0910-0120  
Expiration Date: January 31, 2017  
See PRA Statement below.

510(k) Number (if known)

K170029

Device Name

FloPump 32mL Centrifugal Pump

**Indications for Use (Describe)**

The FloPump 32mL Centrifugal Pump is a device that uses a method other than revolving rollers to pump the blood through an extracorporeal circuit for periods lasting less than 6 hours for the purpose of providing either:

- i) Full or partial cardiopulmonary bypass (i.e., circuit includes an oxygenator) during open surgical procedures on the heart or great vessels; or
- ii) Temporary circulatory bypass for diversion of flow around a planned disruption of the circulatory pathway necessary for open surgical procedures on the aorta or vena cava.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

Date: November 06, 2017

Manufacturer:  
International Biophysics Corporation  
2101 E. St. Elmo Road  
Austin, TX 78744

Contact Person:  
Geoff Marcek  
VP, Engineering and Quality  
Phone: (512) 814-0046  
Email: gmarcek@biophysics.com

Product	Classification	Product Codes
FloPump 32mL Centrifugal Pump	Class II	KFM

Product Code	Regulation and Classification Name
KFM	Pump, Blood, Cardiopulmonary Bypass, Non-Roller Type per 21 CFR 870.4360

### Description:

The FloPump 32mL is a single use disposable centrifugal pump head. The pump has an inlet that draws blood from a patient and an outlet that pushes the blood out, where it then passes through an oxygenator and back to the patient. The pump mates with the Maquet RotaFlow console and rotates the internal impeller using a magnetic driver. The FloPump 32mL is a non-occlusive pump. The pump has a spinning rotor with flow channels which imparts rotary motion to the incoming blood, directing it through a spiral housing to the outflow port. The FloPump 32mL is part of the extracorporeal circuit, and is therefore in contact with the patient's blood while circulating. The FloPump 32mL does not have any other patient contact.

### Specifications:

Model Numbers	6400S (Sterile), 6400N (Non-Sterile)
Priming Volume	Approx. 32 mL
Inlet/Outlet I.D.	9.5mm (3/8")
Max. rated pressure	750 mmHg
Flow rates	0 – 7 L/min
Materials	
Housing, rotor, connector	Polycarbonate
Bearing Cup	HDPE
Spherical Bearing	Alumina
Magnet	Neodymium

### Indications for Use:

The FloPump 32mL Centrifugal Pump is a device that uses a method other than revolving rollers to pump the blood through an extracorporeal circuit for periods lasting less than 6 hours for the purpose of providing either:

- i) Full or partial cardiopulmonary bypass (i.e., circuit includes an oxygenator) during open surgical procedures on the heart or great vessels; or
- ii) Temporary circulatory bypass for diversion of flow around a planned disruption of the circulatory pathway necessary for open surgical procedures on the aorta or vena cava.

**Predicate Devices:** Maquet (Jostra) RotaFlow Centrifugal Pump – K991864  
IBC FloPump Centrifugal Pump – K983272

**Comparable Features to Predicate Device(s):** This device is comparable to the predicate devices in indications, material, design features, specifications, manufacturing methods, raw materials, intended use, packaging, labeling and sterilization.

**Key Differences in Subject Device to Predicate:** There are no key differences between the subject device and the predicate. The subject device has been designed and constructed to be identical to the predicate.

**Non-Clinical Testing:**

The following non-clinical testing was performed to determine substantial equivalence:

<b>Testing</b>	<b>Results Summary</b>
Flow curves	Substantially equivalent to predicates
Heat generation	Substantially equivalent to predicates
Prime volume	Substantially equivalent to predicates
Air handling	Substantially equivalent to predicates
Hemolysis	Substantially equivalent to predicates
Reliability	Substantially equivalent to predicates
Biocompatibility	Substantially equivalent to predicates
Sterilization	The sterilization process results in an SAL of $10^{-6}$
Packaging durability	No signs of damage and functioned as intended following testing
Shelf-life	No signs of damage and functioned as intended following testing

**Clinical Testing:** Clinical testing was not required