



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

October 13, 2017

Shenzhen Pango Electronic Co., Ltd.
% Ms. Diana Hong
General Manager
Mid-Link Consulting Co., Ltd
P.O. Box 120-119
Shanghai, 200120 CN

Re: K170151

Trade/Device Name: Electronic Blood Pressure Monitor
Regulation Number: 21 CFR 870.1130
Regulation Name: Noninvasive Blood Pressure Measurement System
Regulatory Class: Class II
Product Code: DXN
Dated: September 14, 2017
Received: September 15, 2017

Dear Ms. Diana Hong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,


for

Bram D. Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K170151

Device Name

Electronic Blood Pressure Monitor

Indications for Use (Describe)

The Electronic Blood Pressure Monitor is intended to measure the systolic and diastolic blood pressure as well as the pulse rate of adult person via non-invasive oscillometric technique in which an inflatable cuff is wrapped around the upper arm. It can be used at medical facilities or at home. The intended arm circumference includes 22 cm~32 cm and 32 cm~42 cm.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This 510(k) Summary is being submitted in accordance with requirements of Title 21, CFR Section 807.92.

The assigned 510(k) Number: K170151

1. Date of Preparation: 03/9/2017
2. Sponsor Identification

Shenzhen Pango Electronic Co.,Ltd.

No.25, 1st Industrial Park, Fenghuang Road, Xikeng, Henggang, Longgang District Shenzhen, Guangdong, 518115, China.

Establishment Registration Number: 3006792041

Contact Person: Ms. Xiaoyun Yang

Position: Vice General Manager

Tel: +86-755-33825988

Fax: +86-755-33825989

Email: sales@pan-go.com

3. Designated Submission Correspondent

Ms. Diana Hong (Primary Contact Person)

Ms. Jing Cheng (Alternative Contact Person)

Mid-Link Consulting Co., Ltd

P.O. Box 120-119, Shanghai, 200120, China

Tel: +86-21-22815850,

Fax: 240-238-7587

Email: info@mid-link.net

4. Identification of Proposed Device

Trade Name: Electronic Blood Pressure Monitor;

Common Name: Arm Blood Pressure Monitor;

Models: PG-800B22, PG-800B23, PG-800B26, PG-800B27, PG-800B31, PG-800B32, PG-800B33, PG-800B35, PG-800B36, PG-800B37, PG-800B42 and PG-800B43

Regulatory Information

Classification Name: Noninvasive blood pressure measurement system

Classification: II

Product Code: DXN;

Regulation Number: 21 CFR 870.1130;

Review Panel: Cardiovascular;

Indications for Use:

The Electronic Blood Pressure Monitor is intended to measure the systolic and diastolic blood pressure as well as the pulse rate of adult person via non-invasive oscillometric technique in which an inflatable cuff is wrapped around the upper arm. It can be used at medical facilities or at home. The intended arm circumference includes 22 cm~32 cm and 32 cm~42 cm.

Device Description:

The proposed device, Electronic Blood Pressure Monitor, is a battery driven automatic non-invasive blood pressure monitor. It can automatically complete the inflation, deflation and measurement, which can measure systolic and diastolic blood pressure and pulse rate of the adult person at upper arm within its claimed range and accuracy via the oscillometric technique. User can select the unit of the measurement: mmHg or KPa.

The device has the data storage function in order for data reviewing, including the systolic pressure, diastolic pressure, pulse rate and measurement time. It has a bar indicating function, which can indicate the WHO (World Health Organization) Blood Pressure Classification of the measured blood pressure by referencing Diastolic Blood Pressure issued at Journal of Hypertension 1999. Vol 17, No.2.

The proposed electronic blood pressure monitor has 12 models, including PG-800B22, PG-800B23, PG-800B26, PG-800B27, PG-800B31, PG-800B32, PG-800B33, PG-800B35, PG-800B36, PG-800B37, PG-800B42 and PG-800B43. All models follow the same software, measurement principle and NIBP algorithm. The main differences are product appearance and key numbers.

The proposed device is intended to be used in medical facilities or at home.

The product is provided non-sterile, and not to be sterilized by the user prior to use.

5. Identification of Predicate Device

510(k) Number: K131558

Product Name: Electronic Blood Pressure Monitor

Manufacturer: Shenzhen Pango Electronic Co., Ltd

6. Non-Clinical Test Conclusion

Non clinical tests were conducted to verify that the proposed device met all design specifications as was Substantially Equivalent (SE) to the predicate device. The test results demonstrated that the proposed device complies with the following standards:

IEC 60601-1:2005+CORR.1:2006+CORR.2:2007+AM1:2012, Medical electrical equipment – Part 1: General requirements for basic safety, and essential performance.

IEC 60601-1-11:2010, Medical electrical equipment – Part 1-11: General requirements for basic safety, and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.

IEC 60601-1-2:2007, Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic Compatibility

IEC 80601-2-30:2009, Medical electrical equipment - Part 2-30: Particular requirements for the basic safety and essential performance of automated noninvasive sphygmomanometers

7. Substantially Equivalent

Table 1 Substantially Equivalent Comparison

ITEM	Proposed Device, Electronic Blood Pressure Monitor	Predicate Device, PG-800B series Electronic Blood Pressure Monitor, K131558
Product Code	DXN	DXN
Regulation No.	21 CFR 870.1130	21 CFR 870.1130
Class	II	II
Intended Use	The Electronic Blood Pressure Monitor is intended to measure the systolic and diastolic	PG-800B Electronic Blood Pressure Monitor is intended to measure the systolic and

	blood pressure as well as the pulse rate of adult person via non-invasive oscillometric technique in which an inflatable cuff is wrapped around the upper arm. It can be used at medical facilities or at home. The intended arm circumference includes: 22 cm~32 cm and 32 cm~42 cm.	diastolic blood pressure as well as the pulse rate of adult person via non-invasive oscillometric technique in which an inflatable cuff is wrapped around the arm. It can be used at medical facilities or at home. The intended arm circumference is 22-32 cm.
Measurement Type	Upper arm	Upper arm
Patient Population	Adult	Adult
Measurement Item	Systolic Pressure, Diastolic Pressure, Pulse Rate	Systolic Pressure, Diastolic Pressure, Pulse Rate
Principle	Oscillometric	Oscillometric
Component	LCD / Key / Cuff / MCU / Pump / Batteries	LCD / Key / Cuff / MCU / Pump / Batteries
arm circumference	22~32cm and 32~42cm	22~32cm
Blood Pressure Range	30 ~ 280 mmHg	30 ~ 280 mmHg
Pulse Rate Range	40-199 bpm	40-199 bpm
Patient Contact Material	Cuff – Nylon Enclosure – ABS Key - ABS	Cuff – Nylon Enclosure – ABS Key – ABS
Electrical Safety	Comply with IEC 60601-1	Comply with IEC 60601-1
EMC	Comply with IEC 60601-1-2	Comply with IEC 60601-1-2
Particular Performance	Comply with IEC 80601-2-30:2009 and ISO 81060-2:2013	ANSI/AAMI SP10
Software Level Concern	Moderate	Moderate

The proposed device and its predicate devices have the identical intended use, components, principle and performance. The difference between the proposed device and the predicate device do not raise any question regarding its safety and effectiveness.

The proposed device, Electronic Blood Pressure Monitor PG-800B22, PG-800B23, PG-800B26, PG-800B27, PG-800B31, PG-800B32, PG-800B33, PG-800B35, PG-800B36, PG-800B37, PG-800B42 and PG-800B43, is determined to be Substantially Equivalent (SE) to the predicate device, PG-800B series Electronic Blood Pressure Monitor, (K131558), in respect of safety and effectiveness.