



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Canon, Inc.
% Ms. Diane Rutherford
Submissions Manager
Ken Block Consulting
1201 Richardson Drive, Suite 160
RICHARDSON TX 75080

March 1, 2017

Re: K170332
Trade/Device Name: Digital Radiography CXDI-710C Wireless
Digital Radiography CXDI-810C Wireless
Regulation Number: 21 CFR 892.1680
Regulation Name: Stationary x-ray system
Regulatory Class: II
Product Code: MQB
Dated: January 6, 2017
Received: February 2, 2017

Dear Ms. Rutherford:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in blue ink that reads "Michael D. O'Hara". The signature is written in a cursive style and is positioned over a faint, large "FDA" watermark.

For

Robert Ochs, Ph.D.
Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K170332

Device Name

DIGITAL RADIOGRAPHY CXDI-710C Wireless
DIGITAL RADIOGRAPHY CXDI-810C Wireless

Indications for Use (Describe)

The DIGITAL RADIOGRAPHY CXDI-710C Wireless and CXDI-810C Wireless provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”



5. 510(k) SUMMARY

Submitter: Canon, Inc. – Medical Equipment Group
9-1, Imaikami-cho, Nakahara-ku
Kawasaki, Kanagawa 211-8501 Japan

Contact Person: Mr. Shinji Mori
Manager
TEL: 81-3-3758-2111
FAX: 81-44-739-6493
mori.shinji@canon.co.jp

Date Prepared: January 26, 2017

Proposed Device

Manufacturer:	Canon
Trade Name:	DIGITAL RADIOGRAPHY CXDI-710C Wireless DIGITAL RADIOGRAPHY CXDI-810C Wireless
Common Name:	Solid State X-ray Imager
Classification Name:	Stationary X-ray system
Product Code /	MQB
Regulatory Standard:	892.1680 Stationary X-ray System

Predicate Device:

Clearance:	K131106 dated July 03, 2013
Manufacturer:	Canon
Trade Name:	DIGITAL RADIOGRAPHY CXDI-701C Wireless DIGITAL RADIOGRAPHY CXDI-801C Wireless
Common Name:	Solid State X-ray Imager
Classification Name:	Stationary X-ray system
Product Code /	MQB
Regulatory Standard:	892.1680 Stationary X-ray System

Device Description: The two models of detectors included in this submission are solid state x-ray imagers. Model *CXDI-710C Wireless* has an approximate imaging area of 35.0 x 42.6 cm, while model *CXDI-810C Wireless* has an approximate imaging area of 35.0 x 27.4 cm.

For both models, the detector intercepts x-ray photons and the scintillator emits visible spectrum photons that illuminate an array of photo-detectors that create electrical signals. After the electrical signals are generated, the signals are converted to digital values and the images will be displayed on monitors. The digital value can be communicated to the operator console via wiring connection or wireless.

For the proposed models, temporary image storage is now possible and the detector weight has been reduced from that of the predicates. The proposed models have increased protection against ingress, continue to include the Non-Generator Connection Mode (detection of x-ray irradiation without direct electrical connection to the x-ray generator) and are compatible with the Scatter Correction feature.

Indications for Use: The DIGITAL RADIOGRAPHY CXDI-710C Wireless and CXDI-810C Wireless provide digital image capture for conventional film/screen radiographic examinations. These devices are intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. These devices are not intended for mammography applications.



5. 510(k) SUMMARY (continued)

Summary of
Technological
Characteristics:

Comparisons with the predicate devices show the technological characteristics of the proposed DIGITAL RADIOGRAPHY CXDI-710C Wireless and DIGITAL RADIOGRAPHY CXDI-810C Wireless devices to be substantially equivalent to the predicate devices. The proposed devices are functionally identical to the predicate devices.

The major difference between the proposed and predicate devices is that the proposed devices have a reduction in weight and can save captured in an internal nonvolatile memory of the detector. No other technological changes have been made to the proposed devices.

	New Devices: K170332 CXDI-710C / 810C Wireless	Predicate Device: K131106 CXDI-701C / 801C Wireless
Indication for Use	The DIGITAL RADIOGRAPHY CXDI-710C Wireless and CXDI-810C Wireless provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.	The DIGITAL RADIOGRAPHY CXDI-701G Wireless, CXDI-701C Wireless, CXDI-801G Wireless and CXDI-801C Wireless provide digital image capture for conventional film/screen radiographic examinations. These devices are intended to replace radiographic film/screen systems in all general purpose diagnostic procedures. These devices are not intended for mammography applications.
Application	General Radiography	General Radiography
Scintillator	CsI(Tl) [Cesium Iodide doped with Thallium]	CsI(Tl) [Cesium Iodide doped with Thallium]
Pixel Pitch	125 μm	125μm
Pixels	710C: 2,800 x 3,408 (≈ 9.5 mil) 810C: 2,800 x 2,192 (≈ 6.1 mil)	701C: 2,800 x 3,408 (≈ 9.5 mil) 801C: 2,192 x 2,800 (≈ 6.1 mil)
External Dimensions	710C: 384 x 460 x 15.7 mm 810C: 307.5 x 384 x 15.7 mm	701C: 384 x 460 x 15 mm 801C: 307 x 384 x 15 mm
Weight	710C: 2.3 kg 810C: 1.8 kg	701C: 3.3 kg 801C: 2.3 kg
Spatial Resolution	35% [MTF@2lp/mm]	35% [MTF@2lp/mm]
Control SW	CXDI Control Software Added Standalone function	CXDI Control Software
Device FW	PCA-FE-710 Added Standalone function Added wireless channels with DFS and TPC functions	PCA-FE-701
Wireless Functions	Communication between Detector and: Multi Box Control PC	Communication between Detector and: X-ray I/F Unit Control PC

The User's and Installation Manuals provide detailed instructions and information for safe and effective use of the device and users are expected to adhere to the instructions and other information. The User's Manual explains how to use the detector and other equipment. Connected medical equipment, such as X-ray generators, must comply with IEC 60601-1. Before using the product, be sure to read the manual thoroughly in order to utilize it more effectively.



Summary of
Non-Clinical /
Test Data:

Tests were performed on the models which demonstrated that the device is safe and effective, performs comparably to the predicate device(s), and is substantially equivalent to the predicate device(s). Tests included verification/validation testing to internal functional specifications (including software) and non-clinical image comparisons involving flat panel display images taken with the new device and the predicate device(s). Documentation was provided demonstrating compliance of the CXDI-710C Wireless and CXDI-810C Wireless to all FDA requirements stated in Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices for a moderate LOC, including results of verification/validation plus traceability of verification/validation tests to software requirements and software risk hazards. Other FDA guidance documents used in development include Radio Frequency Wireless Technology in Medical Devices and Content of Premarket Submissions for Management of Cybersecurity in Medical Devices.

Documentation was provided demonstrating that the CXDI-710C / 810C Wireless complies with the FDA requirements stated in Guidance for the Submission of 510(k)'s for Solid State X-ray Imaging Devices. The evaluations of the CXDI-710C / 810C Wireless compared to the CXDI-701C / 801C Wireless, show the CXDI-710C / 810C Wireless to be equivalent to the CXDI-701C / 801C Wireless.

Testing confirmed that the CXDI-710C Wireless and CXDI-810C Wireless complies with the U.S. Performance Standard for radiographic equipment and with relevant voluntary safety standards for Electrical safety and Electromagnetic Compatibility testing, specifically IEC standards 60601-1, 60601-1-2, 60601-1-6, and 60601-2-54.

Together, these verification/validation activities successfully demonstrated that the CXDI-710C Wireless and CXDI-810C Wireless correctly performs as designed, has been validated for its intended use, and raises no new questions regarding either safety or effectiveness when compared to the predicate device(s). Therefore, the verification/validation testing conducted supports a determination of substantial equivalence for the CXDI-710C Wireless and CXDI-810C Wireless device.

Conclusion:

Canon, Inc. – Medical Equipment Group considers the DIGITAL RADIOGRAPHY CXDI-710C Wireless and DIGITAL RADIOGRAPHY CXDI-810C Wireless devices to be substantially equivalent to the predicate devices listed above. This conclusion is based on the similarities in primary intended use, principles of operation, functional design, and established medical use.