



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Arthrex Inc.
Mr. David L. Rogers
Project Manager, Regulatory Affairs
1370 Creekside Boulevard
Naples, FL 34108

May 11, 2017

Re: K170382
Trade/Device Name: Arthrex Compression Screws
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: HWC
Dated: April 11, 2017
Received: April 14, 2017

Dear Mr. Rogers:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K170382

Device Name

Arthrex Compression Screws

Indications for Use (Describe)

The Arthrex Compression Screws are intended for fixation of fractures, osteotomies and arthrodesis in:

- carpal, metacarpal and small hand bones
- tarsal and metatarsals
- phalanges
- intra-articular fractures
- ankle
- proximal and distal humerus
- proximal and distal radius
- proximal and distal ulna
- osteochondral fixation and fractures
- osteochondritis dissecans
- fixation of fractures and osteotomies about the knee
- oblique fractures of the fibula
- reconstructive surgeries of the foot
- malleolar fixation

where size of offered implant is patient appropriate.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)☐ Over-The-Counter Use (21 CFR 801 Subpart C)**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASTaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Date Prepared	May 9, 2017
Submitter	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
Contact Person	David L Rogers Project Manager, Regulatory Affairs 1-239-643-5553, ext. 71924 david.rogers@arthrex.com
Name of Device	Arthrex Compression Screws
Common Name	Screw, fixation, bone
Product Code	HWC
Classification Name	21 CFR 888.3040: Smooth or threaded metallic bone fixation fastener
Regulatory Class	II
Primary Predicate Device	K132217: Arthrex Compression FT Screws
Reference Predicate Device	K123241: Arthrex Fracture Plates
Purpose of Submission	This Special 510(k) premarket notification is submitted to obtain clearance for a line extension to the Arthrex Compression Screw family cleared under K132217.
Device Description	The Arthrex Compression Screws are a family of titanium, cannulated, compression screws that are offered in a diameter range of 5.0 to 7.0mm, length range of 20 to 140mm, in a fully threaded design.
Indications for Use	The Arthrex Compression Screws are intended for fixation of fractures, osteotomies and arthrodesis in: • carpal, metacarpal and small hand bones • tarsal and metatarsals • phalanges • intra-articular fractures • ankle • proximal and distal humerus • proximal and distal radius • proximal and distal ulna • osteochondral fixation and fractures • osteochondritis dissecans • fixation of fractures and osteotomies about the knee • oblique fractures of the fibula • reconstructive surgeries of the foot • malleolar fixation where size of offered implant is patient appropriate.
Performance Data	Shear, push-out, torque, and compression testing were conducted to demonstrate that the proposed compression screws perform statistically equivalent to the predicate. Bacterial endotoxin per EP 2.6.14/USP <85> was conducted to demonstrate that

the device meets pyrogen limit specifications.

Conclusion

The Arthrex Compression Screws are substantially equivalent to the predicate device in which the basic design features and intended uses are the same. Any differences between the proposed device and the predicate device are considered minor and do not raise questions concerning safety or effectiveness.

Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the proposed device is substantially equivalent to the currently marketed predicate device.