



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

In2bones USA, LLC
% Christine Scifert
Managing Partner
MRC-X, LLC
6075 Poplar Avenue, Suite 500
Memphis, Tennessee 38119

July 14, 2017

Re: K170518

Trade/Device Name: In2bones Fracture and Correction System

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And
Accessories

Regulatory Class: Class II

Product Code: HRS, HWC, HTN

Dated: June 2, 2017

Received: June 6, 2017

Dear Christine Scifert:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act

or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement on last page.

Indications for Use

510(k) Number (if known)

K170518

Device Name

In2Bones Fracture and Correction System

Indications for Use (Describe)

The In2Bones USA LLC, Fracture and Correction System plates and screws are intended to treat fractures, fusions, osteotomies and non-unions of the 5th metatarsal.

The Fracture and Correction System lag screws are intended to be used as stand-alone bone screws, or in a plate-screw system for internal bone fixation for bone fractures, fusions, osteotomies and non-unions of various bones, including humerus, radius, ulna, tibia, calcaneus, fibula, and small bones (metacarpals, metatarsals, and phalanges).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) Summary of Safety and Effectiveness

Summary of 510(k) safety and effectiveness information upon which the substantial equivalence determination is based:

I. Submitter

Date Prepared: July 11, 2017
Device Submitter: In2Bones USA, LLC
6060 Poplar Avenue, Suite 380
Memphis, TN 38119
Phone: 901-260-7931
Contact Person: Christine Scifert

II. Device

Proprietary Name: In2Bones Fracture and Correction System
Common Name: Plate, Fixation, Bone
Screw, Fixation, Bone
**Regulation Number/
Name:** 21 CFR 888.3030: Single/multiple component metallic
bone fixation appliances and accessories
21 CFR 888.3040: Smooth or threaded metallic bone
fixation fastener
Regulatory Class: Class II
Product Code: HRS, HWC, HTN
Review Panel: Orthopedic

III. Predicate Device

Predicate Device: K123241 Arthrex Fracture Plates
K983495 Syntec Taichung Non-Sterile Bone Plate and Screw
Implants
K131920 In2Bones SAS I.B.S. Osteosynthesis Screw
K160174 In2Bones SAS I.B.S. 2.0 Osteosynthesis Screw

IV. Device Description

The In2Bones Fracture and Correction System is a system of plates, screws and washers used to treat fracture and reconstruction of the bones of the extremities.

V. Intended Use and Indications for Use

The In2Bones USA LLC, Fracture and Correction System plates and screws are intended to treat fractures, fusions, osteotomies and non-unions of the 5th metatarsal.

The Fracture and Correction System lag screws are intended to be used as stand-alone bone screws, or in a plate-screw system for internal bone fixation for bone fractures, fusions, osteotomies and non-unions of various bones, including humerus, radius, ulna, tibia, calcaneus, fibula, and small bones (metacarpals, metatarsals, and phalanges).

VI. Comparison of technological characteristics with the predicate devices

The In2Bones USA Fracture and Correction System and the legally marketed predicate devices have the same intended use and similar indications for use, similar dimensions, geometry and materials. The In2Bones device and the predicates are available in multiple sizes and various shapes. The plates are attached to bone using various diameter and length screws. The In2Bones USA Fracture and Correction System implants are made from titanium alloy.

VII. Performance Data

Testing including pyrogenicity, static driving torque, static pullout, static torsion, static bending and fatigue bending were performed. The results of the testing demonstrate that the device is substantially equivalent to the predicate devices identified.

VIII. Conclusions

The In2Bones USA Fracture and Correction System when compared to the predicates have similar intended use and indications for use, technological characteristics, and principals of operation as the predicate devices. Thus, the In2Bones USA Fracture and Correction System design characteristics do not raise any new types of questions of safety or effectiveness and thus is substantially equivalent to the predicate devices.