



S.B.M. SAS Science for Bio Materials
Ms. Anne Cospin-Latapie
Quality/Regulatory Affairs Manager
ZI du Monge
Lourdes, France 65100

November 30, 2017

Re: K170868

Trade/Device Name: FIXIT Threaded Anchor System

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/multiple component metallic bone fixation appliances and accessories

Regulatory Class: Class II

Product Code: MAI

Dated: October 19, 2017

Received: October 23, 2017

Dear Ms. Cospin-Latapie:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

Indications for Use

510(k) Number (if known)
K170868

Device Name
FIXIT Threaded Anchor System

Indications for Use (Describe)

The FIXIT® threaded anchor system is intended for fixation of suture (soft tissue) to bone in the shoulder, foot/ankle, knee, hand/wrist, and elbow, but not limited to, the following procedures:

Shoulder: Rotator Cuff Repair, Bankart Repair; SLAP Lesion Repair; Biceps Tenodesis; Acromio-clavicular Separation Repair; Deltoid Repair; Capsule Shift or Capsulolabral Reconstruction;

Ankle/Foot: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Hallux Valgus Reconstruction, Mid-foot Reconstruction, Metatarsal Ligament Repair/Tendon Repair, Bunionectomy

Knee: Anterior Cruciate Ligament Repair; Medial Collateral Ligament Repair; Lateral Collateral Ligament Repair; Patellar Tendon Repair ; Posterior Oblique Ligament Repair; Iliotibial Band Tenodesis ;

Wrist/Hand: Scapholunate Ligament Reconstruction; Ulnar or Radial Collateral Ligament Reconstruction, Radial Collateral Ligament Reconstruction.

Elbow: Biceps Tendon Reattachment; Ulnar or Radial Collateral Ligament Reconstruction; Tennis Elbow Repair and Lateral Epicondylitis Repair

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) SUMMARY

1. SUBMITTER

S.B.M. SAS SCIENCE FOR BIOMATERIALS ZI du Monge F 65100 LOURDES – FRANCE Registration Number: 3004549189
Phone: +33 (0)5 62 42 21 01 Fax: +33 (0)5 62 42 21 00
Contact Person: Anne COSPIN-LATAPIE e-mail : anne.cospin@sbm-fr.com
Date prepared: November 29, 2017

2. DEVICE

Name of Device	FIXIT® Threaded Anchor System
Common or Usual Name	Suture Anchor
Classification Name	Fastener, fixation, biodegradable, soft tissue
Regulatory Class	II
Product Code	MAI

3. PREDICATE DEVICE

Arthrex Bio-Composite Corkscrew manufactured by Arthrex, Inc., K082810

Referenced devices:

K070673 Force Fiber® Black Co-braid Polyethylene non-absorbable Suture

K063778 Force Fiber® Polyethylene non-absorbable Suture

4. DEVICE DESCRIPTION

FIXIT® is a threaded anchor system made of Duosorb® 30 (β -TCP 30%/ PLDLA 70%), a composite, bioabsorbable material.

The implant is available in 2 different configurations:

- FIXIT®Knotless: supplied in a holder,

510 (k) FIXIT Threaded Anchor System



- FIXIT®: pre-loaded on a disposable screwdriver with 2 sutures.

The implant is supplied sterile, ready to use.

5. INDICATIONS FOR USE

The FIXIT® threaded anchor system is intended for fixation of suture (soft tissue) to bone in the shoulder, foot/ankle, knee, hand/wrist, and elbow, but not limited to, the following procedures:

Shoulder: Rotator Cuff Repair, Bankart Repair; SLAP Lesion Repair; Biceps Tenodesis; Acromio-clavicular Separation Repair; Deltoid Repair; Capsule Shift or Capsulolabral Reconstruction;

Ankle/Foot: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Hallux Valgus Reconstruction, Mid-foot Reconstruction, Metatarsal Ligament Repair/Tendon Repair, Bunionectomy

Knee: Anterior Cruciate Ligament Repair; Medial Collateral Ligament Repair; Lateral Collateral Ligament Repair; Patellar Tendon Repair ; Posterior Oblique Ligament Repair; Iliotibial Band Tenodesis ;

Wrist/Hand: Scapholunate Ligament Reconstruction; Ulnar or Radial Collateral Ligament Reconstruction, Radial Collateral Ligament Reconstruction.

Elbow: Biceps Tendon Reattachment; Ulnar or Radial Collateral Ligament Reconstruction; Tennis Elbow Repair and Lateral Epicondylitis Repair.

6. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

FIXIT® Threaded Anchor System is compared to Arthrex Bio-Composite Corkscrew (K082810) manufactured by Arthrex, Inc.

The applicant device has the same intended use as the predicate device.

The technological characteristics of this product are believed to be substantially equivalent as those for the predicate device. This device and its predicates use similar performance characteristics, manufacturing materials, and design.

510 (k) FIXIT Threaded Anchor System



	FIXIT® Threaded Anchor System Present submission	Arthrex Bio-Composite Corkscrew K082810
Intended use (same)	Fixation of suture (soft tissue) to bone	
Indications for use (same)	<p><u>Shoulder</u>: Rotator Cuff Repair, Bankart Repair; SLAP Lesion Repair; Biceps Tenodesis; Acromio-clavicular Separation Repair; Deltoid Repair; Capsule Shift or Capsulolabral Reconstruction;</p> <p><u>Ankle/Foot</u>: Lateral Stabilization; Medial Stabilization; Achilles Tendon Repair; Hallux Valgus Reconstruction; Mid-foot Reconstruction, Metatarsal Ligament Repair/ Tendon Repair, Bunnectomy;</p> <p><u>Knee</u>: Anterior Cruciate Ligament Repair; Medial Collateral Ligament Repair; Lateral Collateral Ligament Repair; Patellar Tendon Repair ; Posterior Oblique Ligament Repair; Iliotibial Band Tenodesis ;</p> <p><u>Wrist/Hand</u>: Scapholunate Ligament Reconstruction; Ulnar or Radial Collateral Ligament Reconstruction, Radial Collateral Ligament Reconstruction.</p> <p><u>Elbow</u>: Biceps Tendon Reattachment; Ulnar or Radial Collateral Ligament Reconstruction; Tennis Elbow Repair and Lateral Epicondylitis Repair,</p>	
Materials	TCP/PLDLA	TCP/PLLA or TCP/PLDLA
Configuration/ dimensions	<p>Ø 4,5mm x14,5 mm, 2 sutures</p> <p>Ø 5,5mm x17,6 mm, 2 sutures</p> <p>Ø 6,5mm x17,6 mm, 2 sutures</p>	<p>Ø 4,5mm x15 mm, 2 sutures</p> <p>Ø 5,5mm x15 mm, 2 or 3 sutures</p> <p>Ø 6,5mm x15 mm, 2 or 3 sutures</p>

Packaging	Anchor loaded with suture on driver sealed in foil pouch	Anchor loaded with suture on driver sealed in plastic tray with Tyvek pouch, which is then sealed in a foil pouch
------------------	--	---

4. PERFORMANCE DATA

Non-clinical performance testing

Non-clinical testing including biocompatibility, biological and mechanical performances was performed to determine substantial equivalence. The results indicated that the devices were functional within their intended use and equivalent to the predicate devices.

Bacterial endotoxin testing has been completed and results have demonstrated that the proposed devices meet the endotoxin limits.

Clinical performance testing:

Clinical performance data was not included.

5. CONCLUSIONS

The FIXIT® Threaded Anchor System is substantially equivalent to its predicate device Arthrex Bio-Composite Corkscrew (K082810). Verification and validation tests demonstrate that the FIXIT® Threaded Anchor System is as safe, as effective, and performs as safely and effectively as its predicate device.